



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 12/21/2022 16:00 TEMP1 OUT: 103 °C 12/22/2022 08:00
TEMP2 IN: 104 °C 12/22/2022 08:30 TEMP2 OUT: 104 °C 12/22/2022 10:00
TEMP3 IN: 540 °C 12/22/2022 10:30 TEMP3 OUT: 540 °C 12/22/2022 12:00
TEMP4 IN: 560 °C 12/22/2022 12:30 TEMP4 OUT: 560 °C 12/22/2022 14:00

Run Number: LB123371

SUPERVISOR: Iwona

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB123371BL	98.6716	98.6716	100	98.6716	98.6716	98.6716	98.6716	0.0000	0
2	N6152-01	89.9945	89.9945	100	92.6293	92.6293	91.9895	91.9895	0.6398	6398
3	N6152-01DUP	89.9867	89.9867	100	92.7114	92.7114	92.0859	92.0859	0.6255	6255

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \star 1000000$

WORKLIST(Hardcopy Internal Chain)

VB 123371

WorkList Name : TVS W N6152 **WorkList ID :** 166009 **Department :** Wet-Chemistry **Date :** 12-21-2022 10:28:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method	
12/27/2022	Water	N6152-01	TVS	A	Cool 4 deg C	PSEG03	N12	GROUND-WATER	12/20/2022	SM2540 G

Date/Time 12-21-22 12:40
Raw Sample Received by: JD WCJ
Raw Sample Relinquished by: CP SM

Date/Time 12-21-22 16:10
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JD WCJ