

LB123378


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45269



A= 0.0000e+000

B= 2.2034e-004 slope

C= 1.1095e-002 $y = mx + c$

Rho= 0.9998757

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.038	0.038	122	0.000	122					
0.05	0.050					185	0				
0.20	0.200	0.211	0.011	907	0.0 %	907					
2.50	2.500	2.473	-0.027	11174	0.0 %	11174					
5.00	5.000	4.902	-0.098	22195	0.0 %	22195					
7.50	7.500	7.590	0.090	34397	0.0 %	34397					
10.0	10.000	9.986	-0.014	45269	0.0 %	45269					

90D
116
112
10

LB123378

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	122	-		0 SFAM01.1	PPB	12/21/2022 15:51	Std
	0.2 S01	907	-		0.2 SFAM01.1	PPB	12/21/2022 15:53	Std
	2.5 S02	11174	-		2.5 SFAM01.1	PPB	12/21/2022 15:55	Std
	5 S03	22195	-		5 SFAM01.1	PPB	12/21/2022 15:58	Std
	7.5 S04	34397	-		7.5 SFAM01.1	PPB	12/21/2022 16:00	Std
	10 S05	45269	-		10 SFAM01.1	PPB	12/21/2022 16:02	Std
ICV069	ICV069	18517	4.0912 -		SFAM01.1	PPB	12/21/2022 16:05	SMPL
ICB069	ICB069	73	0.0272 -		SFAM01.1	PPB	12/21/2022 16:07	SMPL
CCV019	CCV019	22930	5.0636 -		SFAM01.1	PPB	12/21/2022 16:09	SMPL
CCB019	CCB019	65	0.0254 -		SFAM01.1	PPB	12/21/2022 16:12	SMPL
PB149829BL	PBS829	20	0.0155 -		SFAM01.1	PPB	12/21/2022 16:14	SMPL
N5893-01	YBK86	1059393	HIGH -		SFAM01.1	PPB	12/21/2022 16:16	
N5893-02	YBK87	607426	HIGH -		SFAM01.1	PPB	12/21/2022 16:18	
N5893-03	YBK87D	600434	HIGH -		SFAM01.1	PPB	12/21/2022 16:20	
N5893-04	YBK87S	635164	HIGH -		SFAM01.1	PPB	12/21/2022 16:22	
N5893-05	YBK90	1119287	HIGH -		SFAM01.1	PPB	12/21/2022 16:23	
N5893-06	YBK91	1080765	HIGH -		SFAM01.1	PPB	12/21/2022 16:25	
N5893-01DLX10	YBK86	205877	45.3751 -		SFAM01.1	PPB	12/21/2022 16:27	SMPL
N5893-02DLX10	YBK87S	86540	19.0797 -		SFAM01.1	PPB	12/21/2022 16:29	SMPL
N5893-03DLX10	YBK87D	86096	18.9819 -		SFAM01.1	PPB	12/21/2022 16:32	SMPL
N5893-04DLX10	YBK87S	89848	19.8086 -		SFAM01.1	PPB	12/21/2022 16:34	SMPL
N5893-05DLX10	YBK90	226322	49.88 -		SFAM01.1	PPB	12/21/2022 16:36	SMPL
N5893-06DLX10	YBK91	204847	45.1481 -		SFAM01.1	PPB	12/21/2022 16:38	SMPL
CCV020	CCV020	22122	4.8856 -		SFAM01.1	PPB	12/21/2022 16:41	SMPL
CCB020	CCB020	-309	-0.057 -		SFAM01.1	PPB	12/21/2022 16:43	SMPL
N5893-01DL2X100	YBK86	19597	4.3292 -		SFAM01.1	PPB	12/21/2022 16:45	SMPL
N5893-02DL2X50	YBK87	15201	3.3606 -		SFAM01.1	PPB	12/21/2022 16:48	SMPL
N5893-03DL2X50	YBK87D	15179	3.3557 -		SFAM01.1	PPB	12/21/2022 16:50	SMPL
N5893-04DL2X50	YBK87S	15500	3.4264 -		SFAM01.1	PPB	12/21/2022 16:52	SMPL
N5893-05DL2X100	YBK90	20067	4.4328 -		SFAM01.1	PPB	12/21/2022 16:54	SMPL
N5893-06DL2X100	YBK91	16847	3.7232 -		SFAM01.1	PPB	12/21/2022 17:00	SMPL
CCV021	CCV021	22634	4.9984 -		SFAM01.1	PPB	12/21/2022 17:02	SMPL
CCB021	CCB021	-174	-0.0272 -		SFAM01.1	PPB	12/21/2022 17:04	SMPL