



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123381

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

True Value of CCV4 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	12/21/2022	16:40
2	CAL2	1	Water	NA	NA	20.7	7.00	12/21/2022	16:41
3	CAL3	1	Water	NA	NA	20.7	10.03	12/21/2022	16:42
4	ICV	1	Water	NA	NA	20.7	6.98	12/21/2022	16:45
5	CCV1	1	Water	NA	NA	20.7	2.01	12/21/2022	16:47
6	N6152-01	1	Water	NA	NA	20.7	6.44	12/21/2022	16:50
7	N6152-01DUP	1	Water	NA	NA	20.8	6.45	12/21/2022	16:51
8	N6161-10	1	Water	NA	NA	20.2	7.90	12/21/2022	17:00
9	N6165-01	1	Water	NA	NA	20.2	7.58	12/21/2022	17:02
10	N6165-02	1	Water	NA	NA	20.4	7.92	12/21/2022	17:05
11	N6165-03	1	Water	NA	NA	20.6	7.83	12/21/2022	17:06
12	N6165-04	1	Water	NA	NA	20.6	7.39	12/21/2022	17:10
13	N6165-05	1	Water	NA	NA	20.2	9.87	12/21/2022	17:13
14	N6165-06	1	Water	NA	NA	20.3	9.88	12/21/2022	17:18
15	N6165-07	1	Water	NA	NA	20.6	9.84	12/21/2022	17:22
16	CCV2	1	Water	NA	NA	20.8	12.03	12/21/2022	17:25
17	N6165-08	1	Water	NA	NA	20.2	8.83	12/21/2022	17:29
18	N6168-01	1	Water	NA	NA	20.2	9.20	12/21/2022	17:33
19	N6168-02	1	Water	NA	NA	20.1	8.94	12/21/2022	17:36
20	N6168-03	1	Water	NA	NA	20.2	6.33	12/21/2022	17:39

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	N6168-04	1	Water	NA	NA	20.2	8.10	12/21/2022	17:43
22	N6168-05	1	Water	NA	NA	20.3	5.67	12/21/2022	17:45
23	N6168-06	1	Water	NA	NA	20.2	5.37	12/21/2022	17:49
24	N6168-07	1	Water	NA	NA	20.4	7.14	12/21/2022	17:52
25	N6168-11	1	Water	NA	NA	20.4	7.50	12/21/2022	18:00
26	N6168-12	1	Water	NA	NA	20.2	7.49	12/21/2022	18:02
27	CCV3	1	Water	NA	NA	20.8	2.01	12/21/2022	18:10
28	N6168-13	1	Water	NA	NA	20.2	7.52	12/21/2022	18:13
29	N6168-14	1	Water	NA	NA	20.2	7.63	12/21/2022	18:15
30	N6168-14DUP	1	Water	NA	NA	20.3	7.64	12/21/2022	18:16
31	CCV4	1	Water	NA	NA	20.8	12.03	12/21/2022	18:20

WORKLIST(Hardcopy Internal Chain)

WB 123381

WorkList Name : ph n6165

WorkList ID : 166030

Department : Wet-Chemistry

Date : 12-21-2022 16:29:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/20/2022	Water	N6152-01	pH	Cool 4 deg C	PSEG03	N12	GROUND-WATER	12/20/2022	9040C
12/20/2022	Water	N6161-10	pH	Cool 4 deg C	PSEG03	N12	CLI-0659	12/20/2022	9040C
12/21/2022	Water	N6165-01	pH	Cool 4 deg C	ATCE02	N12	12x5	12/21/2022	9040C
12/21/2022	Water	N6165-02	pH	Cool 4 deg C	ATCE02	N12	12x6	12/21/2022	9040C
12/21/2022	Water	N6165-03	pH	Cool 4 deg C	ATCE02	N12	12x7	12/21/2022	9040C
12/21/2022	Water	N6165-04	pH	Cool 4 deg C	ATCE02	N12	12x8	12/21/2022	9040C
12/21/2022	Water	N6165-05	pH	Cool 4 deg C	ATCE02	N12	12x9	12/21/2022	9040C
12/21/2022	Water	N6165-06	pH	Cool 4 deg C	ATCE02	N12	12x10	12/21/2022	9040C
12/21/2022	Water	N6165-07	pH	Cool 4 deg C	ATCE02	N12	12x11	12/21/2022	9040C
12/21/2022	Water	N6165-08	pH	Cool 4 deg C	ATCE02	N12	12x12	12/21/2022	9040C
12/21/2022	Water	N6168-01	pH	Cool 4 deg C	PSEG03	N12	AUD-1893	12/21/2022	9040C
12/21/2022	Water	N6168-02	pH	Cool 4 deg C	PSEG03	N12	AUD-1894	12/21/2022	9040C
12/21/2022	Water	N6168-03	pH	Cool 4 deg C	PSEG03	N12	AUD-1895	12/21/2022	9040C
12/21/2022	Water	N6168-04	pH	Cool 4 deg C	PSEG03	N12	AUD-1692	12/21/2022	9040C
12/21/2022	Water	N6168-05	pH	Cool 4 deg C	PSEG03	N12	AUD-1693	12/21/2022	9040C
12/21/2022	Water	N6168-06	pH	Cool 4 deg C	PSEG03	N12	AUD-1691	12/21/2022	9040C
12/21/2022	Water	N6168-07	pH	Cool 4 deg C	PSEG03	N12	AUD-1741(1421)	12/21/2022	9040C
12/21/2022	Water	N6168-11	pH	Cool 4 deg C	PSEG03	N12	AUD-1839	12/21/2022	9040C
12/21/2022	Water	N6168-12	pH	Cool 4 deg C	PSEG03	N12	AUD-1840	12/21/2022	9040C
12/21/2022	Water	N6168-13	pH	Cool 4 deg C	PSEG03	N12	AUD-1841	12/21/2022	9040C
12/21/2022	Water	N6168-14	pH	Cool 4 deg C	PSEG03	N12	AUD-1842	12/21/2022	9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: