



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/23/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 14:20
In Date: 12/22/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 12/23/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123386

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N6153-01	DBXP5	1	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
N6153-02	DBXP6	2	1.00	1.00	2.00	2.00	100.0	P.E.SAMPLE
N6153-03	MDBXR6	3	1.15	8.35	9.5	9.35	98.2	
N6153-04	MDBXR8	4	1.16	8.56	9.72	9.52	97.7	
N6153-05	MDBXS2	5	1.16	8.41	9.57	9.44	98.5	
N6153-06	MDBXS4	6	1.16	8.34	9.5	9.09	95.1	
N6153-07	MDBXS8	7	1.15	8.69	9.84	9.75	99.0	
N6153-08	MDBXS9	8	1.15	8.48	9.63	9.54	98.9	
N6153-09	MDBXT0	9	1.15	8.60	9.75	9.25	94.2	
N6153-10	MDBXT1	10	1.12	8.55	9.67	9.44	97.3	
N6153-11	MDBXT3	11	1.15	8.39	9.54	9.21	96.1	
N6153-12	MDBXT6	12	1.15	8.35	9.5	9.29	97.5	
N6153-13	MDBXQ0	13	1.15	5.31	6.46	6.33	97.6	
N6153-14	MDBXQ1	14	1.12	8.39	9.51	9.04	94.4	
N6153-15	MDBXQ2	15	1.11	8.70	9.81	9.6	97.6	
N6153-16	MDBXQ3	16	1.12	8.66	9.78	9.5	96.8	
N6153-17	MDBXQ5	17	1.11	8.42	9.53	9.37	98.1	
N6153-18	MDBXQ8	18	1.14	8.41	9.55	9.4	98.2	
N6153-19	MDBXQ9	19	1.13	8.45	9.58	9.44	98.3	
N6153-20	MDBXQ9D	20	1.13	8.45	9.58	9.44	98.3	
N6153-21	MDBXQ9S	21	1.13	8.45	9.58	9.44	98.3	
N6153-22	MDBXR0	22	1.15	8.51	9.66	9.5	98.1	
N6153-23	MDBXR1	23	1.12	8.40	9.52	9.21	96.3	
N6153-24	MDBXR5	24	1.14	8.38	9.52	9.33	97.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$