

Clear table

Instrument ID: IC-2										Analyst : IZ		Method: 300.0 / 9056A	
ident	concentra tion F-	concentratio n CL-	concentratio on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst		
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-12-19_	12/19/22 10:55		IZ/MA		
STD2	0.4930	0.6520	0.7170	2.2030	0.5380	0.9890	3.3100	2022-12-19_	12/19/22 12:00		IZ/MA		
STD3	0.7840	1.1990	1.2150	3.9970	1.0000	2.0180	6.0820	2022-12-19_	12/19/22 12:32		IZ/MA		
STD4	0.9750	1.4510	1.4550	4.8530	1.2080	2.3980	7.3260	2022-12-19_	12/19/22 13:03		IZ/MA		
STD5	1.9060	2.9770	2.9280	9.9480	2.5130	5.1560	14.8460	2022-12-19_	12/19/22 14:38		IZ/MA		
STD6	4.0420	6.0380	5.8390	19.8410	4.9570	9.9310	29.4720	2022-12-19_	12/19/22 15:10		IZ/MA		
STD7	5.0000	7.4840	7.6450	25.1570	6.2840	12.5080	37.9630	2022-12-19_	12/19/22 15:41		IZ/MA		
ICV	1.8290	2.9160	2.9020	9.7570	2.4510	4.8320	14.6320	2022-12-19_	12/19/22 16:13		IZ/MA		
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-12-19_	12/19/22 16:45		IZ/MA		
CCV	2.0980	2.8850	2.8220	9.7120	2.4110	4.8450	14.0430	2022-12-21_	12/21/22 10:47		IZ/MA		
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-12-21_	12/21/22 11:22		IZ/MA		
LB123390BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-12-21_	12/21/22 11:54		IZ/MA		
LB123390BSW	1.9510	2.9090	2.8170	9.7800	2.4320	4.9500	14.2100	2022-12-21_	12/21/22 12:25		IZ/MA		
N6152-01	0.3880	4.7640	0.4530	0.0000	1.3020	0.2120	44.9180	2022-12-21_	12/21/22 12:56		IZ/MA		
N6152-01MS	3.3020	7.7670	3.3460	10.2640	3.8110	8.9620	56.5820	2022-12-21_	12/21/22 13:28		IZ/MA		
N6152-01MSD	3.1510	7.6480	3.2260	10.0670	3.7480	8.9810	55.8920	2022-12-21_	12/21/22 13:59		IZ/MA		
N6174-01X10	0.0000	145.4900	0.3270	0.8360	1.0610	0.2810	3.2850	2022-12-21_	12/21/22 16:06		IZ/MA		
N6174-01DLX300	0.0000	3.2810	0.0000	0.0000	0.0000	0.0000	0.0000	2022-12-21_	12/21/22 16:38		IZ/MA		
CCV	1.8660	2.7720	2.8400	9.3870	2.3230	5.4980	13.5660	2022-12-21_	12/21/22 18:13		IZ/MA		
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-12-21_	12/21/22 18:44		IZ/MA		


Clear table

Instrument ID: IC-2 Analyst: IZ Method: 300.0 / 9056A

ident	concentra tion F-	concentratio n CL-	concentratio on NO2	concentratio on BR-	concentratio on NO3	concentratio on HPO4	concentratio on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2022-12-19_	12/19/22 10:55		IZ/MA
STD2	0.4930	0.6520	0.7170	2.2030	0.5380	0.9890	3.3100	2022-12-19_	12/19/22 12:00		IZ/MA
STD3	0.7840	1.1990	1.2150	3.9970	1.0000	2.0180	6.0820	2022-12-19_	12/19/22 12:32		IZ/MA
STD4	0.9750	1.4510	1.4550	4.8530	1.2080	2.3980	7.3260	2022-12-19_	12/19/22 13:03		IZ/MA
STD5	1.9060	2.9770	2.9280	9.9480	2.5130	5.1560	14.8460	2022-12-19_	12/19/22 14:38		IZ/MA
STD6	4.0420	6.0380	5.8390	19.8410	4.9570	9.9310	29.4720	2022-12-19_	12/19/22 15:10		IZ/MA
STD7	5.0000	7.4840	7.6450	25.1570	6.2840	12.5080	37.9630	2022-12-19_	12/19/22 15:41		IZ/MA
ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4				
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000				
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000				
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000				
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000				
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000				
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000				
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.0000				
ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4				
STD1											
STD2	23.2500	8.6667	19.5000	10.1500	7.6000	-1.1000	10.3333				
STD3	-2.0000	-0.0833	1.2500	-0.0750	0.0000	0.9000	1.3667				
STD4	-2.5000	-3.2667	-3.0000	-2.9400	-3.3600	-4.0800	-2.3200				
STD5	-4.7000	-0.7667	-2.4000	-0.5200	0.5200	3.1200	-1.0267				
STD6	1.0500	0.6333	-2.6833	-0.7950	-0.8600	-0.6900	-1.7600				
STD7	0.0000	-0.2133	1.9333	0.6280	0.5440	0.0640	2.6027				

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-121922.mtw
 Equation: $Q = 0.0721308 \cdot A + 3.7121$
 RSD: 3.229 %
 Correlation coefficient: 0.999431




K3 = 0 K2 = 0 K1 = 0.0721308 K0 = 3.7121
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	1.839	85.27	0.4	20	5.877		
3	3.757	165.9	0.8	20	5.877		
4	4.879	218.8	1	20	5.877		
5	11.41	477	2	20	5.877		
6	25.28	1069	4	20	5.877		
7	32.69	1335	5	20	5.877		

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-121922.mtw
 Equation: $Q = 0.127193 \cdot A + 1.45797$
 RSD: 1.297 %
 Correlation coefficient: 0.999908




K3 = 0 K2 = 0 K1 = 0.127193 K0 = 1.45797
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	4.873	90.99	0.6	20	9.035		
3	9.775	177.1	1.2	20	9.035		
4	11.99	216.6	1.5	20	9.035		
5	25.75	456.6	3	20	9.035		
6	52.4	937.9	6	20	9.035		
7	67.5	1165	7.5	20	9.035		
6.&							

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-121922.mtw
 Equation: $Q = 0.0537662 \cdot A + 5.07165$
 RSD: 3.958 %
 Correlation coefficient: 0.999146




K3 = 0 K2 = 0 K1 = 0.0537662 K0 = 5.07165
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.55	172.6	0.6	20	10.93		
3	15.76	357.7	1.2	20	10.93		
4	19.39	446.8	1.5	20	10.93		
5	43.18	994.7	3	20	10.93		
6	88.82	2078	6	20	10.93		
7	113.5	2750	7.5	20	10.93		
1.#							

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-121922.mtw
 Equation: $Q = 0.287163 \cdot A + 10.4243$
 RSD: 1.544 %
 Correlation coefficient: 0.999870




K3 = 0 K2 = 0 K1 = 0.287163 K0 = 10.4243
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	4.81	117.1	2	20	13.41	
3	10.03	242.1	4	20	13.41	
4	12.49	301.7	5	20	13.41	
5	27.54	656.5	10	20	13.41	
6	57.36	1346	20	20	13.41	
7	74.62	1716	25	20	13.41	

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-121922.mtw
 Equation: $Q = 0.0505639 \cdot A + 1.66986$
 RSD: 1.450 %
 Correlation coefficient: 0.999885




K3 = 0 K2 = 0 K1 = 0.0505639 K0 = 1.66986
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	5.832	179.8	0.5	20	16.64		
3	11.82	362.5	1	20	16.64		
4	14.64	444.9	1.25	20	16.64		
5	31.64	961	2.5	20	16.64		
6	64.81	1928	5	20	16.64		
7	83.92	2452	6.25	20	16.64		

□

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-121922.mtw
 Equation: $Q = 0.0585091 \cdot A + 1.84657$
 RSD: 1.822 %
 Correlation coefficient: 0.999819




K3 = 0 K2 = 0 K1 = 0.0585091 K0 = 1.84657
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	4.732	306.6	1	20	21.97		
3	10.6	658.1	2	20	21.97		
4	12.72	788.1	2.5	20	21.97		
5	28.43	1731	5	20	21.97		
6	56.62	3363	10	20	21.97		
7	72.06	4244	12.5	20	21.97		

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-121922.mtw
 Equation: $Q = 0.176123 \cdot A + 8.93076$
 RSD: 2.442 %
 Correlation coefficient: 0.999675



K3 = 0 K2 = 0 K1 = 0.176123 K0 = 8.93076
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	7.072	325.2	3	20	25.5		
3	14.01	640	6	20	25.5		
4	17.26	781.2	7.5	20	25.5		
5	36.04	1635	15	20	25.5		
6	72.89	3296	30	20	25.5		
7	94.26	4260	37.5	20	25.5		

Report date: 12/19/2022 4:49:25 PM
Printed by: wet


Ident: STD1
Analysis from: 12/19/2022 10:55:01 AM
File: _2022-12-19_

Last save: 12/19/2022 4:21:34 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32871

Last save: 12/19/2022 3:39

SAMPLE:
: IZ/MA
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 4:49:33 PM
Printed by: wet


Ident: STD2
Analysis from: 12/19/2022 12:00:33 PM
File: _2022-12-19_

Last save: 12/19/2022 4:21:34 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32873

Last save: 12/19/2022 3:39

SAMPLE:
: IZ/MA
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.87	0.621	1.84	5.01	85.267	6.67
2	8.80	0.243	4.87	13.27	90.987	7.12
3	10.62	0.298	7.55	20.57	172.551	13.51
4	13.72	0.339	4.81	13.10	117.149	9.17
5	15.72	0.405	5.83	15.88	179.802	14.07
6	21.77	0.841	4.73	12.88	306.601	24.00
7	24.89	0.540	7.07	19.26	325.158	25.45
7	29.00	0.470	36.70	99.97	1277.514	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 4:49:41 PM
Printed by: wet


Ident: STD3
Analysis from: 12/19/2022 12:32:11 PM
File: _2022-12-19_

Last save: 12/19/2022 4:21:34 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32874

Last save: 12/19/2022 3:39

SAMPLE:
: IZ/MA
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.573	3.76	4.96	165.942	6.37
2	8.73	0.239	9.78	12.91	177.124	6.80
3	10.54	0.296	15.76	20.81	357.712	13.74
4	13.60	0.329	10.02	13.23	242.112	9.30
5	15.52	0.400	11.82	15.60	362.541	13.93
6	21.65	0.783	10.60	13.99	658.129	25.28
7	24.77	0.533	14.01	18.49	639.953	24.58
7	29.00	0.451	75.74	99.99	2603.512	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 4:49:49 PM
Printed by: wet


Ident: STD4
Analysis from: 12/19/2022 1:03:48 PM
File: _2022-12-19_

Last save: 12/19/2022 4:21:34 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32875

Last save: 12/19/2022 3:39

SAMPLE:
: IZ/MA
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.86	0.577	4.88	5.23	218.798	6.84
2	8.77	0.240	11.99	12.84	216.626	6.77
3	10.57	0.301	19.39	20.77	446.787	13.97
4	13.62	0.329	12.49	13.38	301.727	9.43
5	15.53	0.399	14.64	15.68	444.924	13.91
6	21.72	0.777	12.72	13.62	788.076	24.64
7	24.85	0.528	17.26	18.49	781.233	24.43
7	29.00	0.450	93.36	100.00	3198.172	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 4:49:56 PM
Printed by: wet


Ident: STD5
Analysis from: 12/19/2022 2:38:43 PM
File: _2022-12-19_

Last save: 12/19/2022 4:21:34 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32878

Last save: 12/19/2022 3:39

SAMPLE:
: IZ/MA
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.84	0.551	11.41	5.60	476.997	6.90
2	8.72	0.236	25.75	12.62	456.612	6.61
3	10.49	0.304	43.18	21.17	994.717	14.39
4	13.54	0.320	27.54	13.50	656.545	9.50
5	15.38	0.393	31.64	15.51	961.040	13.90
6	21.66	0.739	28.43	13.94	1731.024	25.04
7	24.93	0.526	36.03	17.66	1635.203	23.66
7	29.00	0.438	203.98	100.00	6912.138	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 4:50:07 PM
Printed by: wet


Ident: STD6
Analysis from: 12/19/2022 3:10:21 PM
File: _2022-12-19_

Last save: 12/19/2022 4:21:34 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32879

Last save: 12/19/2022 3:39

SAMPLE:
: IZ/MA
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.88	0.581	25.28	6.04	1069.300	7.63
2	8.72	0.237	52.39	12.53	937.936	6.69
3	10.52	0.312	88.81	21.24	2077.843	14.82
4	13.56	0.312	57.36	13.72	1345.561	9.60
5	15.36	0.389	64.81	15.50	1927.559	13.75
6	21.69	0.725	56.62	13.54	3363.107	23.99
7	24.89	0.525	72.89	17.43	3296.082	23.51
7	29.00	0.440	418.16	100.00	14017.389	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 4:50:14 PM
Printed by: wet


Ident: STD7
Analysis from: 12/19/2022 3:41:59 PM
File: _2022-12-19_

Last save: 12/19/2022 4:21:34 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32880

Last save: 12/19/2022 3:39

SAMPLE:
: IZ/MA
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.84	0.555	32.68	6.07	1334.932	7.45
2	8.66	0.228	67.50	12.53	1165.311	6.50
3	10.40	0.311	113.51	21.08	2749.645	15.34
4	13.70	0.305	74.61	13.85	1715.799	9.57
5	15.47	0.383	83.91	15.58	2452.382	13.68
6	21.79	0.717	72.06	13.38	4244.003	23.68
7	24.99	0.525	94.26	17.50	4260.266	23.77
7	29.00	0.432	538.53	100.00	17922.338	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/19/2022 4:50:22 PM
Printed by: wet


Ident: ICV
Analysis from: 12/19/2022 4:13:44 PM
File: _2022-12-19_

Last save: 12/19/2022 4:42:45 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32881

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.538	11.17	5.56	455.616	6.80
2	8.67	0.233	25.48	12.68	446.997	6.67
3	10.42	0.299	43.47	21.62	985.021	14.71
4	13.40	0.318	27.32	13.59	643.232	9.60
5	15.23	0.388	31.46	15.65	936.491	13.98
6	21.61	0.745	26.79	13.32	1620.187	24.19
7	24.77	0.528	35.34	17.58	1610.898	24.05
7	29.00	0.435	201.03	99.99	6698.444	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/20/2022 10:49:50 AM
Printed by: wet


Ident: ICV
Analysis from: 12/19/2022 4:13:44 PM
File: _2022-12-19_

Last save: 12/19/2022 4:42:45 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32881

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.538	11.17	5.56	455.616	6.80
2	8.67	0.233	25.48	12.68	446.997	6.67
3	10.42	0.299	43.47	21.62	985.021	14.71
4	13.40	0.318	27.32	13.59	643.232	9.60
5	15.23	0.388	31.46	15.65	936.491	13.98
6	21.61	0.745	26.79	13.32	1620.187	24.19
7	24.77	0.528	35.34	17.58	1610.898	24.05
7	29.00	0.435	201.03	99.99	6698.444	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/20/2022 10:49:58 AM
Printed by: wet


Ident: ICB
Analysis from: 12/19/2022 4:45:54 PM
File: _2022-12-19_

Last save: 12/19/2022 5:14:54 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32882

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/22/2022 11:40:27 AM
Printed by: wet


Ident: CCV
Analysis from: 12/21/2022 10:47:32 AM
File: _2022-12-21_

Last save: 12/21/2022 11:16:32 AM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32911

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.90	0.637	10.86	5.78	530.230	7.96
2	8.88	0.248	23.68	12.59	442.228	6.64
3	10.75	0.319	39.46	20.98	955.399	14.35
4	13.96	0.350	24.61	13.08	640.108	9.62
5	15.95	0.433	27.98	14.88	920.743	13.83
6	21.46	0.746	27.03	14.38	1624.496	24.40
7	24.61	0.524	34.45	18.32	1543.963	23.19
7	29.00	0.465	188.06	100.00	6657.167	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/22/2022 11:40:43 AM
Printed by: wet


Ident: CCB
Analysis from: 12/21/2022 11:22:32 AM
File: _2022-12-21_

Last save: 12/21/2022 11:51:32 AM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32912

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/22/2022 11:40:49 AM
Printed by: wet


Ident: LB123390BLW
Analysis from: 12/21/2022 11:54:11 AM
File: _2022-12-21_

Last save: 12/22/2022 11:40:18 AM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32913

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 12/22/2022 11:40:57 AM
Printed by: wet


Ident: LB123390BSW
Analysis from: 12/21/2022 12:25:49 PM
File: _2022-12-21_

Last save: 12/22/2022 11:40:19 AM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32914

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.87	0.606	10.86	5.62	489.395	7.32
2	8.79	0.244	24.35	12.61	446.016	6.67
3	10.61	0.312	40.34	20.89	953.646	14.26
4	13.73	0.340	25.68	13.30	644.856	9.64
5	15.66	0.417	29.41	15.23	928.991	13.89
6	21.39	0.745	27.62	14.31	1660.645	24.84
7	24.53	0.523	34.81	18.03	1562.991	23.38
7	28.00	0.455	193.07	100.00	6686.540	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/22/2022 11:41:11 AM
Printed by: wet


Ident: N6152-01
Analysis from: 12/21/2022 12:56:42 PM
File: _2022-12-21_

Last save: 12/21/2022 1:25:42 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32915

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 30
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.89	0.594	1.56	0.92	56.131	0.87
2	8.74	0.249	39.68	23.42	737.704	11.45
3	10.58	0.359	2.75	1.62	74.291	1.15
4	0.00	0.000	0.00	0.00	0.000	0.00
5	15.63	0.400	15.56	9.19	481.836	7.48
6	21.55	1.182	0.54	0.32	40.867	0.63
7	24.47	0.523	109.31	64.52	5050.073	78.41
7	29.00	0.473	169.40	100.00	6440.902	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/22/2022 11:41:18 AM
Printed by: wet


Ident: N6152-01MS
Analysis from: 12/21/2022 1:28:19 PM
File: _2022-12-21_

Last save: 12/21/2022 1:57:20 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32916

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 31
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.92	0.641	18.06	4.59	864.145	5.85
2	8.78	0.246	66.39	16.86	1209.899	8.18
3	10.59	0.357	43.70	11.10	1150.246	7.78
4	13.71	0.324	28.19	7.16	678.581	4.59
5	15.57	0.401	48.49	12.31	1474.422	9.97
6	21.43	0.851	47.01	11.93	3031.833	20.51
7	24.52	0.522	142.01	36.06	6374.530	43.12
7	29.00	0.478	393.84	100.00	14783.655	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/22/2022 11:41:31 AM
Printed by: wet


Ident: N6152-01MSD
Analysis from: 12/21/2022 1:59:57 PM
File: _2022-12-21_

Last save: 12/21/2022 2:28:57 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32917

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 32
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.91	0.671	16.68	4.34	822.097	5.64
2	8.75	0.246	65.28	16.99	1191.142	8.18
3	10.56	0.353	42.34	11.02	1105.584	7.59
4	13.68	0.333	27.03	7.03	664.808	4.56
5	15.57	0.401	47.36	12.32	1449.557	9.95
6	21.50	0.872	46.45	12.09	3038.559	20.86
7	24.60	0.527	139.12	36.20	6296.241	43.22
7	29.00	0.486	384.25	100.00	14567.987	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/22/2022 11:41:42 AM
Printed by: wet


Ident: N6174-01X10
Analysis from: 12/21/2022 4:06:25 PM
File: _2022-12-21_

Last save: 12/21/2022 4:35:25 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32921

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 33
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.74	0.261	1202.47	98.18	22865.509	96.53
3	10.44	0.334	1.09	0.09	27.184	0.11
4	13.56	0.274	1.12	0.09	21.895	0.09
5	15.37	0.380	12.95	1.06	386.611	1.63
6	21.48	1.291	0.79	0.06	64.387	0.27
7	24.49	0.553	6.49	0.53	322.380	1.36
7	29.00	0.442	1224.92	100.01	23687.965	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/22/2022 11:41:49 AM
Printed by: wet


Ident: N6174-01DLX300
Analysis from: 12/21/2022 4:38:03 PM
File: _2022-12-21_

Last save: 12/21/2022 5:07:03 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32922

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 34
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.66	0.237	28.53	99.99	504.509	100.00
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	0.00	0.000	0.00	0.00	0.000	0.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	29.00	0.034	28.53	99.99	504.509	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/22/2022 11:41:58 AM
Printed by: wet


Ident: CCV
Analysis from: 12/21/2022 6:13:06 PM
File: _2022-12-21_

Last save: 12/22/2022 9:14:20 AM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32925

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.80	0.549	11.18	5.77	465.830	6.96
2	8.68	0.240	23.86	12.32	424.418	6.34
3	10.46	0.308	41.78	21.57	961.955	14.37
4	13.47	0.333	25.49	13.16	617.496	9.23
5	15.34	0.406	29.24	15.09	885.824	13.24
6	21.28	0.808	29.72	15.34	1847.663	27.61
7	24.41	0.540	32.42	16.74	1489.810	22.26
7	29.00	0.455	193.69	99.99	6692.996	100.00

This report has been created by IC Net
METROHM LTD

Report date: 12/22/2022 11:42:06 AM
Printed by: wet


Ident: CCB
Analysis from: 12/21/2022 6:44:46 PM
File: _2022-12-21_

Last save: 12/21/2022 7:13:46 PM

Method: AnionsIC2-121922.mtw
Run operator: wet
Analysis number: 32926

Last save: 12/19/2022 4:10

SAMPLE:
: IZ/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD