


LB123391

7470A

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.3830e-004

C= 1.2689e-002

Rho= 0.9995164

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.030  | 0.030  | 74    | 0.000      | 74    |       |       |       |       | 1  |
| 0.2    | 0.200  | 0.202  | 0.002  | 793   | 0.0 %      | 793   |       |       |       |       | 0  |
| 2.5    | 2.500  | 2.497  | -0.003 | 10424 | 0.0 %      | 10424 |       |       |       |       | 1  |
| 5.0    | 5.000  | 5.074  | 0.074  | 21240 | 0.0 %      | 21240 |       |       |       |       | -3 |
| 7.5    | 7.500  | 7.265  | -0.235 | 30434 | 0.0 %      | 30434 |       |       |       |       | 1  |
| 10.0   | 10.000 | 10.132 | 0.132  | 42464 | 0.0 %      | 42464 |       |       |       |       |    |

## LB123391 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Dec 2022 11:54:29

| Sample ID       | Extended ID         | μ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|---------------------|--------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50                  | 74     | -       | 0.0000   | 7470A  | PPB   | 22 Dec 2022 12:07:18 | S    | Std  |
| 0.2 - 1         | 50.2                | 793    | -       | 0.2000   | 7470A  | PPB   | 22 Dec 2022 12:09:35 | S    | Std  |
| 2.5 - 1         | 52.5                | 10424  | -       | 2.5000   | 7470A  | PPB   | 22 Dec 2022 12:11:52 | S    | Std  |
| 5.0 - 1         | 55                  | 21240  | -       | 5.0000   | 7470A  | PPB   | 22 Dec 2022 12:14:09 | S    | Std  |
| 7.5 - 1         | 57.5                | 30434  | -       | 7.5000   | 7470A  | PPB   | 22 Dec 2022 12:16:26 | S    | Std  |
| 10.0 - 1        | 51.0                | 42464  | -       | 10.0000  | 7470A  | PPB   | 22 Dec 2022 12:18:42 | S    | Std  |
| ICV70 - 1       | ICV70               | 17247  | 4.1227  | -        | 7470A  | PPB   | 22 Dec 2022 12:21:31 | U    | SMPL |
| ICB70 - 1       | ICB70               | 56     | 0.0260  | -        | 7470A  | PPB   | 22 Dec 2022 12:23:46 | U    | SMPL |
| CCV22 - 1       | CCV22               | 21354  | 5.1014  | -        | 7470A  | PPB   | 22 Dec 2022 12:26:01 | U    | SMPL |
| CCB22 - 1       | CCB22               | -10    | 0.0103  | -        | 7470A  | PPB   | 22 Dec 2022 12:28:16 | U    | SMPL |
| CRA - 1         | CRA                 | 789    | 0.2007  | -        | 7470A  | PPB   | 22 Dec 2022 12:30:31 | U    | SMPL |
| HighStd - 1     | HighStd             | 42630  | 10.1715 | -        | 7470A  | PPB   | 22 Dec 2022 12:32:47 | U    | SMPL |
| ChkStd - 1      | ChkStd              | 29069  | 6.9399  | -        | 7470A  | PPB   | 22 Dec 2022 12:35:02 | U    | SMPL |
| PB149837BL - 1  | PBW                 | -53    | 0.0001  | -        | 7470A  | PPB   | 22 Dec 2022 12:37:17 | U    | SMPL |
| PB149837BS - 1  | LCSW                | 16040  | 3.8350  | -        | 7470A  | PPB   | 22 Dec 2022 12:39:33 | U    | SMPL |
| N6151-01 - 1    | WC-09-(0-2)         | 480    | 0.1271  | -        | 7470A  | PPB   | 22 Dec 2022 12:41:49 | U    | SMPL |
| N6151-06 - 1    | FS-01-(0-2)         | 18     | 0.0170  | -        | 7470A  | PPB   | 22 Dec 2022 12:44:05 | U    | SMPL |
| N6151-08 - 1    | FS-02-(0-2)         | 19     | 0.0172  | -        | 7470A  | PPB   | 22 Dec 2022 12:46:22 | U    | SMPL |
| N6160-02 - 1    | COMP-3051           | -10    | 0.0103  | -        | 7470A  | PPB   | 22 Dec 2022 12:48:40 | U    | SMPL |
| N6161-09 - 1    | N40305              | -2     | 0.0122  | -        | 7470A  | PPB   | 22 Dec 2022 12:50:55 | U    | SMPL |
| CCV23 - 1       | CCV23               | 21602  | 5.1605  | -        | 7470A  | PPB   | 22 Dec 2022 12:53:10 | U    | SMPL |
| CCB23 - 1       | CCB23               | -69    | -0.0038 | -        | 7470A  | PPB   | 22 Dec 2022 12:55:26 | U    | SMPL |
| N6161-09DUP - 1 | N40305DUP           | 16     | 0.0165  | -        | 7470A  | PPB   | 22 Dec 2022 12:57:42 | U    | SMPL |
| N6161-09MS - 1  | N40305MS            | 16543  | 3.9549  | -        | 7470A  | PPB   | 22 Dec 2022 12:59:58 | U    | SMPL |
| N6161-09MSD - 1 | N40305MSD           | 16326  | 3.9032  | -        | 7470A  | PPB   | 22 Dec 2022 13:02:13 | U    | SMPL |
| PB149853BL - 1  | PBW                 | -14    | 0.0094  | -        | 7470A  | PPB   | 22 Dec 2022 13:04:28 | U    | SMPL |
| PB149853BS - 1  | LCSW                | 15677  | 3.7485  | -        | 7470A  | PPB   | 22 Dec 2022 13:06:44 | U    | SMPL |
| N6088-03 - 1    | EB-2022-12-15       | -16    | 0.0089  | -        | 7470A  | PPB   | 22 Dec 2022 13:09:00 | U    | SMPL |
| N6147-01 - 1    | SUMP                | 47     | 0.0239  | -        | 7470A  | PPB   | 22 Dec 2022 13:11:17 | U    | SMPL |
| N6152-01 - 1    | GROUND-WATER        | 2617   | 0.6363  | -        | 7470A  | PPB   | 22 Dec 2022 13:13:34 | U    | SMPL |
| N6161-10 - 1    | CLI-0659            | 151    | 0.0487  | -        | 7470A  | PPB   | 22 Dec 2022 13:15:51 | U    | SMPL |
| N6161-15 - 1    | 31880               | 1507   | 0.3718  | -        | 7470A  | PPB   | 22 Dec 2022 13:18:06 | U    | SMPL |
| CCV24 - 1       | CCV24               | 21080  | 5.0361  | -        | 7470A  | PPB   | 22 Dec 2022 13:20:22 | U    | SMPL |
| CCB24 - 1       | CCB24               | -53    | 0.0001  | -        | 7470A  | PPB   | 22 Dec 2022 13:22:37 | U    | SMPL |
| N6168-01 - 1    | AUD-1893            | 768    | 0.1957  | -        | 7470A  | PPB   | 22 Dec 2022 13:24:53 | U    | SMPL |
| N6168-02 - 1    | AUD-1894            | 1039   | 0.2603  | -        | 7470A  | PPB   | 22 Dec 2022 13:27:09 | U    | SMPL |
| N6168-03 - 1    | AUD-1895            | 9757   | 2.3378  | -        | 7470A  | PPB   | 22 Dec 2022 13:29:24 | U    | SMPL |
| N6168-04 - 1    | AUD-1692            | 1048   | 0.2624  | -        | 7470A  | PPB   | 22 Dec 2022 13:31:41 | U    | SMPL |
| N6168-05 - 1    | AUD-1693            | 291    | 0.0820  | -        | 7470A  | PPB   | 22 Dec 2022 13:33:57 | U    | SMPL |
| N6168-07 - 1    | AUD-1741(1421)      | -8     | 0.0108  | -        | 7470A  | PPB   | 22 Dec 2022 13:36:13 | U    | SMPL |
| N6168-09 - 1    | AUD-1839-42-COMP    | 2530   | 0.6156  | -        | 7470A  | PPB   | 22 Dec 2022 13:38:30 | U    | SMPL |
| N6168-09DUP - 1 | AUD-1839-42-COMPDUP | 2515   | 0.6120  | -        | 7470A  | PPB   | 22 Dec 2022 13:40:47 | U    | SMPL |
| N6168-09MS - 1  | AUD-1839-42-COMPMS  | 11645  | 2.7877  | -        | 7470A  | PPB   | 22 Dec 2022 13:43:04 | U    | SMPL |
| N6168-09MSD - 1 | AUD-1839-42-COMPMSD | 12377  | 2.9621  | -        | 7470A  | PPB   | 22 Dec 2022 13:45:20 | U    | SMPL |
| CCV25 - 1       | CCV25               | 21336  | 5.0971  | -        | 7470A  | PPB   | 22 Dec 2022 13:47:35 | U    | SMPL |
| CCB25 - 1       | CCB25               | -66    | -0.0030 | -        | 7470A  | PPB   | 22 Dec 2022 13:49:51 | U    | SMPL |
| PB149800TB - 1  |                     | 35     | 0.0210  | -        | 7470A  | PPB   | 22 Dec 2022 13:52:07 | U    | SMPL |
| N6161-09LX5 - 1 |                     | 3      | 0.0134  | -        | 7470A  | PPB   | 22 Dec 2022 13:54:22 | U    | SMPL |
| N6161-09A - 1   |                     | 16541  | 3.9544  | -        | 7470A  | PPB   | 22 Dec 2022 13:56:38 | U    | SMPL |
| N6168-09LX5 - 1 |                     | 690    | 0.1771  | -        | 7470A  | PPB   | 22 Dec 2022 14:08:52 | U    | SMPL |
| N6168-09A - 1   |                     | 17204  | 4.1124  | -        | 7470A  | PPB   | 22 Dec 2022 14:11:07 | U    | SMPL |
| CCV26 - 1       | CCV26               | 20616  | 4.9255  | -        | 7470A  | PPB   | 22 Dec 2022 14:13:25 | U    | SMPL |
| CCB26 - 1       | CCB26               | -56    | -0.0007 | -        | 7470A  | PPB   | 22 Dec 2022 14:15:42 | U    | SMPL |