



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB123394

Reagent/Standard	Lot/Log #
calibration std. phosphate 1 ppm	WP100282
calibration std. phosphate 0.5 ppm	WP100281
calibration std. phosphate 0.3 ppm	WP100280
calibration std. phosphate 0.1 ppm	WP100279
calibration std. phosphate 0.05 ppm	WP100278
calibration std. 0 ppm	WP100277
phosphate CCV std.	WP100283
Combined reagent	WP100357
Phenolphthalein indicator	WP100008
Sodium hydroxide, 1N	WP98644
Phosphate ICV-LCS Std	WP100284

Intercept: -0.0017

Slope: 0.6541

Regression: 0.999889

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	%D	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	0.003		12/22/2022	13:35
2	CAL2	0.05	1	50	50	0.033	0.053	-6	12/22/2022	13:35
3	CAL3	0.10	1	50	50	0.065	0.102	-2	12/22/2022	13:36
4	CAL4	0.30	1	50	50	0.187	0.288	4	12/22/2022	13:36
5	CAL5	0.50	1	50	50	0.326	0.501	-0.2	12/22/2022	13:37
6	CAL6	1.00	1	50	50	0.654	1.002	-0.2	12/22/2022	13:37



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Iwona

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: jignesh

Run Number: LB123394

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.493	12/22/2022	13:38
2	ICB		1	50	50	0.001	0.004	12/22/2022	13:38
3	CCV1	0.50	1	50	50	0.328	0.504	12/22/2022	13:39
4	CCB1		1	50	50	0.003	0.007	12/22/2022	13:39
5	PB149859BL		1	50	50	0.002	0.006	12/22/2022	13:40
6	PB149859BS	0.50	1	50	50	0.318	0.489	12/22/2022	13:40
7	N6120-01		1	50	50	1.025	1.570	12/22/2022	13:41
8	N6172-01		1	50	50	0.628	0.963	12/22/2022	13:41
9	N6172-01DUP		1	50	50	0.629	0.964	12/22/2022	13:42
10	N6172-01MS	0.50	1	50	50	0.924	1.415	12/22/2022	13:42
11	N6172-01MSD	0.50	1	50	50	0.925	1.417	12/22/2022	13:43
12	N6174-01		1	50	50	0.872	1.336	12/22/2022	13:43
13	N6120-01		10	50	50	0.582	0.892	12/22/2022	13:44
14	N6172-01MS	0.50	2	50	50	0.466	0.715	12/22/2022	13:44
15	CCV2	0.50	1	50	50	0.324	0.498	12/22/2022	13:45
16	CCB2		1	50	50	0.002	0.006	12/22/2022	13:45
17	N6172-01MSD	0.50	2	50	50	0.465	0.713	12/22/2022	13:46
18	N6174-01		5	50	50	0.352	0.541	12/22/2022	13:46
19	CCV3	0.50	1	50	50	0.321	0.493	12/22/2022	13:47
20	CCB3		1	50	50	0.001	0.004	12/22/2022	13:47