


LB123404

SFAM01.1

INSTRUMENT ID: CV4

Linear



A= 0.0000e+000
 B= 2.4021e-004 *slope*
 C= -5.9451e-003 *intercept*
 Rho= 0.9999244
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.006	0.006	48	0.000	48					
0.05	0.050					185	0				
0.20	0.200	0.184	-0.016	790	0.0 %	790					
2.50	2.500	2.467	-0.033	10295	0.0 %	10295					
5.00	5.000	5.095	0.095	21234	0.0 %	21234					
7.50	7.500	7.456	-0.044	31063	0.0 %	31063					
10.0	10.000	9.993	-0.007	41627	0.0 %	41627					

Handwritten notes:
 100
 18
 11
 2
 1
 0

LB123404

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	48	-		0 SFAM01.1	PPB	12/22/2022 16:20	Std
	0.2 S01	790	-		0.2 SFAM01.1	PPB	12/22/2022 16:23	Std
	2.5 S02	10295	-		2.5 SFAM01.1	PPB	12/22/2022 16:25	Std
	5 S03	21234	-		5 SFAM01.1	PPB	12/22/2022 16:27	Std
	7.5 S04	31063	-		7.5 SFAM01.1	PPB	12/22/2022 16:30	Std
	10 S05	41627	-		10 SFAM01.1	PPB	12/22/2022 16:32	Std
ICV071	ICV071	17646	4.2328 -		SFAM01.1	PPB	12/22/2022 16:34	SMPL
ICB071	ICB071	69	0.0106 -		SFAM01.1	PPB	12/22/2022 16:37	SMPL
CCV027	CCV027	21184	5.0826 -		SFAM01.1	PPB	12/22/2022 16:39	SMPL
CCB027	CCB027	-2	-0.0064 -		SFAM01.1	PPB	12/22/2022 16:41	SMPL
PB149865BL	PBW865	24	-0.0002 -		SFAM01.1	PPB	12/22/2022 16:43	SMPL
PB149757TB	LEB865	12	-0.0031 -		SFAM01.1	PPB	12/22/2022 16:46	SMPL
N6045-01	MEXGJ0	9	-0.0038 -		SFAM01.1	PPB	12/22/2022 16:48	SMPL
N6045-02	MEXGJ0D	16	-0.0021 -		SFAM01.1	PPB	12/22/2022 16:50	SMPL
N6045-03	MEXGJ0S	5033	1.203 -		SFAM01.1	PPB	12/22/2022 16:55	SMPL
CCV028	CCV028	20794	4.989 -		SFAM01.1	PPB	12/22/2022 16:57	SMPL
CCB028	CCB028	-30	-0.0132 -		SFAM01.1	PPB	12/22/2022 17:00	SMPL