



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB123410

Ambient Barometric Pressure (mmHg): 734.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	84.0	12/23/2022	10:00
2	N6152-01		1	12	100.00	>212.0	>212.0	12/23/2022	10:30
3	N6168-01		1	14	100.00	>212.0	>212.0	12/23/2022	11:00
4	N6168-01DUP		1	13	100.00	>212.0	>212.0	12/23/2022	11:30
5	N6168-02		1	12	100.00	>212.0	>212.0	12/23/2022	12:00
6	N6168-03		1	12	100.00	>212.0	>212.0	12/23/2022	12:30
7	N6168-04		1	14	100.00	>212.0	>212.0	12/23/2022	13:00
8	N6168-05		1	13	100.00	>212.0	>212.0	12/23/2022	13:30
9	N6168-06		1	12	100.00	>212.0	>212.0	12/23/2022	14:00
10	N6168-07		1	12	100.00	>212.0	>212.0	12/23/2022	14:30
11	N6168-09		1	12	100.00	>212.0	>212.0	12/23/2022	15:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6123410

WorkList Name : fp12-22

WorkList ID : 166066

Department : Wet-Chemistry

Date : 12-22-2022 11:45:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/02/2023	Water	N6152-01	Flash Point	Cool 4 deg C	PSEG03	N12	GROUND-WATER	12/20/2022	1010B
12/26/2022	Water	N6168-01	Flash Point	Cool 4 deg C	PSEG03	N12	AUD-1893	12/21/2022	1010B
12/26/2022	Water	N6168-02	Flash Point	Cool 4 deg C	PSEG03	N12	AUD-1894	12/21/2022	1010B
12/26/2022	Water	N6168-03	Flash Point	Cool 4 deg C	PSEG03	N12	AUD-1895	12/21/2022	1010B
12/26/2022	Water	N6168-04	Flash Point	Cool 4 deg C	PSEG03	N12	AUD-1692	12/21/2022	1010B
12/26/2022	Water	N6168-05	Flash Point	Cool 4 deg C	PSEG03	N12	AUD-1693	12/21/2022	1010B
12/26/2022	Water	N6168-06	Flash Point	Cool 4 deg C	PSEG03	N12	AUD-1691	12/21/2022	1010B
12/26/2022	Water	N6168-07	Flash Point	Cool 4 deg C	PSEG03	N12	AUD-1741(1421)	12/21/2022	1010B
12/26/2022	Water	N6168-09	Flash Point	Cool 4 deg C	PSEG03	N12	AUD-1839-42-COMP	12/21/2022	1010B

Date/Time 12/23/2022 09:45

Raw Sample Received by: RMI CWC

Raw Sample Relinquished by: RS (Pxt-104)

Date/Time 12/23/2022 15:40

Raw Sample Received by: RS (Pxt-104)

Raw Sample Relinquished by: RMI CWC