

LB123413


7471B

INSTRUMENT ID: CV1

Linear

 $\mu$  Abs.:

53842



A= 0.0000e+000

B= 1.8677e-004

C= 1.2768e-002

Rho= 0.9997795

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.023  | 0.023  | 54    | 0.000      | 54    |       |       |       |       | -2 |
| 0.2    | 0.200  | 0.196  | -0.004 | 980   | 0.0 %      | 980   |       |       |       |       | -1 |
| 2.5    | 2.500  | 2.468  | -0.032 | 13145 | 0.0 %      | 13145 |       |       |       |       | 2  |
| 5.0    | 5.000  | 5.091  | 0.091  | 27191 | 0.0 %      | 27191 |       |       |       |       | -2 |
| 7.5    | 7.500  | 7.354  | -0.146 | 39305 | 0.0 %      | 39305 |       |       |       |       | 1  |
| 10.0   | 10.000 | 10.069 | 0.069  | 53842 | 0.0 %      | 53842 |       |       |       |       |    |

## LB123413 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Dec 2022 10:30:07

| Sample ID         | Extended ID          | $\mu$ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-------------------|----------------------|------------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1           | 50                   | 54         | -       | 0.0000   | 7471B  | PPB   | 23 Dec 2022 10:30:26 | S    | Std  |
| 0.2 - 1           | 50.2                 | 980        | -       | 0.2000   | 7471B  | PPB   | 23 Dec 2022 10:32:44 | S    | Std  |
| 2.5 - 1           | 52.5                 | 13145      | -       | 2.5000   | 7471B  | PPB   | 23 Dec 2022 10:35:01 | S    | Std  |
| 5.0 - 1           | 55                   | 27191      | -       | 5.0000   | 7471B  | PPB   | 23 Dec 2022 10:37:17 | S    | Std  |
| 7.5 - 1           | 57.5                 | 39305      | -       | 7.5000   | 7471B  | PPB   | 23 Dec 2022 10:39:34 | S    | Std  |
| 10.0 - 1          | 510                  | 53842      | -       | 10.0000  | 7471B  | PPB   | 23 Dec 2022 10:47:07 | S    | Std  |
| ICV72 - 1         | ICV72                | 22076      | 4.1358  | -        | 7471B  | PPB   | 23 Dec 2022 10:49:45 | U    | SMPL |
| ICB72 - 1         | ICB72                | 53         | 0.0227  | -        | 7471B  | PPB   | 23 Dec 2022 10:52:00 | U    | SMPL |
| CCV29 - 1         | CCV29                | 27407      | 5.1315  | -        | 7471B  | PPB   | 23 Dec 2022 10:54:14 | U    | SMPL |
| CCB29 - 1         | CCB29                | -21        | 0.0088  | -        | 7471B  | PPB   | 23 Dec 2022 10:56:29 | U    | SMPL |
| CRA - 1           | CRA                  | 981        | 0.1960  | -        | 7471B  | PPB   | 23 Dec 2022 10:58:44 | U    | SMPL |
| HighStd - 1       | HighStd              | 54224      | 10.1400 | -        | 7471B  | PPB   | 23 Dec 2022 11:01:00 | U    | SMPL |
| ChkStd - 1        | ChkStd               | 36810      | 6.8877  | -        | 7471B  | PPB   | 23 Dec 2022 11:03:15 | U    | SMPL |
| PB149868BL - 1    | PBS                  | -67        | 0.0003  | -        | 7471B  | PPB   | 23 Dec 2022 11:05:31 | U    | SMPL |
| PB149868BS - 1    | LCSS                 | 22807      | 4.2724  | -        | 7471B  | PPB   | 23 Dec 2022 11:07:48 | U    | SMPL |
| N6148-01 - 1      | COMPOSITE-3052       | 10925      | 2.0532  | -        | 7471B  | PPB   | 23 Dec 2022 11:10:05 | U    | SMPL |
| N6151-05 - 1      | FS-01-(0-2)          | 67132      | 12.5508 | -        | 7471B  | PPB   | 23 Dec 2022 11:12:21 | U    | SMPL |
| N6151-07 - 1      | FS-02-(0-2)          | 119856     | 22.3980 | -        | 7471B  | PPB   | 23 Dec 2022 11:14:38 | U    | SMPL |
| N6151-10 - 1      | WC-09-(0-2)          | 162989     | 30.4538 | -        | 7471B  | PPB   | 23 Dec 2022 11:16:56 | U    | SMPL |
| N6157-01 - 1      | 20221418-COMPOSITE-A | 22051      | 4.1312  | -        | 7471B  | PPB   | 23 Dec 2022 11:19:13 | U    | SMPL |
| CCV30 - 1         | CCV30                | 27322      | 5.1156  | -        | 7471B  | PPB   | 23 Dec 2022 11:21:31 | U    | SMPL |
| CCB30 - 1         | CCB30                | 143        | 0.0395  | -        | 7471B  | PPB   | 23 Dec 2022 11:23:46 | U    | SMPL |
| N6157-02 - 1      | 20221419-COMPOSITE-B | 26661      | 4.9922  | -        | 7471B  | PPB   | 23 Dec 2022 11:26:01 | U    | SMPL |
| N6157-03 - 1      | 20221420-COMPOSITE-C | 31354      | 5.8687  | -        | 7471B  | PPB   | 23 Dec 2022 11:28:17 | U    | SMPL |
| N6157-04 - 1      | 20221421-COMPOSITE-D | 17929      | 3.3613  | -        | 7471B  | PPB   | 23 Dec 2022 11:30:32 | U    | SMPL |
| N6160-01 - 1      | COMP-3051            | 7866       | 1.4819  | -        | 7471B  | PPB   | 23 Dec 2022 11:32:47 | U    | SMPL |
| N6161-05 - 1      | 31879                | 82         | 0.0281  | -        | 7471B  | PPB   | 23 Dec 2022 11:35:03 | U    | SMPL |
| N6161-06 - 1      | CLI-0116             | 112        | 0.0337  | -        | 7471B  | PPB   | 23 Dec 2022 11:37:20 | U    | SMPL |
| N6161-08 - 1      | N40305               | 1744       | 0.3385  | -        | 7471B  | PPB   | 23 Dec 2022 11:39:36 | U    | SMPL |
| N6163-01 - 1      | TR-06-122022         | 1115       | 0.2210  | -        | 7471B  | PPB   | 23 Dec 2022 11:41:53 | U    | SMPL |
| N6178-01 - 1      | COMP-1               | 3612       | 0.6874  | -        | 7471B  | PPB   | 23 Dec 2022 11:44:12 | U    | SMPL |
| N6178-03 - 1      | COMP-2               | 2630       | 0.5040  | -        | 7471B  | PPB   | 23 Dec 2022 11:46:27 | U    | SMPL |
| CCV31 - 1         | CCV31                | 26661      | 4.9922  | -        | 7471B  | PPB   | 23 Dec 2022 11:48:43 | U    | SMPL |
| CCB31 - 1         | CCB31                | -79        | -0.0020 | -        | 7471B  | PPB   | 23 Dec 2022 11:50:58 | U    | SMPL |
| N6180-01 - 1      | COMP-3050            | 2827       | 0.5408  | -        | 7471B  | PPB   | 23 Dec 2022 11:53:14 | U    | SMPL |
| N6183-01 - 1      | R-1                  | 3303       | 0.6297  | -        | 7471B  | PPB   | 23 Dec 2022 11:55:30 | U    | SMPL |
| N6183-02 - 1      | R-2                  | 23397      | 4.3826  | -        | 7471B  | PPB   | 23 Dec 2022 11:57:45 | U    | SMPL |
| N6183-03 - 1      | R-3                  | 5974       | 1.1285  | -        | 7471B  | PPB   | 23 Dec 2022 12:00:01 | U    | SMPL |
| N6183-03DUP - 1   | R-3DUP               | 5387       | 1.0189  | -        | 7471B  | PPB   | 23 Dec 2022 12:02:17 | U    | SMPL |
| N6183-03MS - 1    | R-3MS                | 30475      | 5.7045  | -        | 7471B  | PPB   | 23 Dec 2022 12:04:33 | U    | SMPL |
| N6183-03MSD - 1   | R-3MSD               | 29972      | 5.6106  | -        | 7471B  | PPB   | 23 Dec 2022 12:09:31 | U    | SMPL |
| N6183-03LX5 - 1   |                      | 1119       | 0.2218  | -        | 7471B  | PPB   | 23 Dec 2022 12:11:47 | U    | SMPL |
| N6183-03A - 1     |                      | 28005      | 5.2432  | -        | 7471B  | PPB   | 23 Dec 2022 12:14:04 | U    | SMPL |
| CCV32 - 1         | CCV32                | 26739      | 5.0067  | -        | 7471B  | PPB   | 23 Dec 2022 12:16:20 | U    | SMPL |
| CCB32 - 1         | CCB32                | -24        | 0.0083  | -        | 7471B  | PPB   | 23 Dec 2022 12:18:39 | U    | SMPL |
| N6088-01DLX5 - 1  |                      | 23341      | 4.3721  | -        | 7471B  | PPB   | 23 Dec 2022 12:20:54 | U    | SMPL |
| N6088-02DLX10 - 1 |                      | 27324      | 5.1160  | -        | 7471B  | PPB   | 23 Dec 2022 12:25:47 | U    | SMPL |
| N6151-05DLX5 - 1  |                      | 12592      | 2.3645  | -        | 7471B  | PPB   | 23 Dec 2022 12:28:02 | U    | SMPL |
| N6151-07DLX5 - 1  |                      | 22510      | 4.2169  | -        | 7471B  | PPB   | 23 Dec 2022 12:30:18 | U    | SMPL |
| N6151-10DLX5 - 1  |                      | 30602      | 5.7282  | -        | 7471B  | PPB   | 23 Dec 2022 12:32:34 | U    | SMPL |
| CCV33 - 1         | CCV33                | 24961      | 4.6747  | -        | 7471B  | PPB   | 23 Dec 2022 12:34:49 | U    | SMPL |
| CCB33 - 1         | CCB33                | -22        | 0.0087  | -        | 7471B  | PPB   | 23 Dec 2022 12:37:06 | U    | SMPL |