



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123417

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	12/22/2022	18:10
2	CAL2	1	Water	NA	NA	20.6	7.00	12/22/2022	18:11
3	CAL3	1	Water	NA	NA	20.7	10.03	12/22/2022	18:12
4	ICV	1	Water	NA	NA	20.7	6.99	12/22/2022	18:15
5	CCV1	1	Water	NA	NA	20.6	2.01	12/22/2022	18:20
6	N6179-02	1	Solid	20.02	20	22.4	8.04	12/22/2022	18:25
7	N6180-02	1	Solid	20.03	20	22.6	7.65	12/22/2022	18:29
8	N6196-02	1	Solid	20.02	20	21.6	8.11	12/22/2022	18:30
9	N6196-02DUP	1	Solid	20.03	20	21.7	8.12	12/22/2022	18:31
10	CCV2	1	Water	NA	NA	20.8	12.03	12/22/2022	18:35

WORKLIST(Hardcopy Internal Chain)

123417

WorkList Name : CORROSIONITY N6179

WorkList ID : 166072

Department : Wet-Chemistry

Date : 12-22-2022 12:44:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/21/2022	Solid	N6179-02	Corrosivity	Cool 4 deg C	PSEG03	N12	T-1-TRANSFORMER-DEBRIS	12/21/2022	9045D
12/21/2022	Solid	N6180-02	Corrosivity	Cool 4 deg C	SOUN03	N12	COMP-3050	12/21/2022	9045D
12/21/2022	Solid	N6196-02	Corrosivity	Cool 4 deg C	SOUN03	N13	COMP-3049	12/21/2022	9045D

Date/Time 12.22.22 18:05
 Raw Sample Received by: CP Smith
 Raw Sample Relinquished by: CP Smith

Date/Time 12.22.22 19:00
 Raw Sample Received by: CP Smith
 Raw Sample Relinquished by: 101020