



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123419

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	12/23/2022	10:25
2	CAL2	1	Water	NA	NA	20.7	7.00	12/23/2022	10:26
3	CAL3	1	Water	NA	NA	20.7	10.03	12/23/2022	10:29
4	ICV	1	Water	NA	NA	20.8	7.02	12/23/2022	10:30
5	CCV1	1	Water	NA	NA	20.7	2.01	12/23/2022	10:32
6	N6191-01	1	Water	NA	NA	20.5	4.74	12/23/2022	10:35
7	N6191-01DUP	1	Water	NA	NA	20.6	4.75	12/23/2022	10:36
8	N6191-02	1	Water	NA	NA	20.7	4.86	12/23/2022	10:40
9	N6191-03	1	Water	NA	NA	20.6	4.13	12/23/2022	10:45
10	CCV2	1	Water	NA	NA	20.7	12.03	12/23/2022	10:48

WORKLIST(Hardcopy Internal Chain)

123419

WorkList Name : ph n6191 w

WorkList ID : 166101

Department : Wet-Chemistry

Date : 12-23-2022 10:13:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/07/2022	Water	N6191-01	pH	Cool 4 deg C	STER09	N13	CELL-A	12/07/2022	9040C
12/07/2022	Water	N6191-02	pH	Cool 4 deg C	STER09	N13	CELL-B/C	12/07/2022	9040C
12/07/2022	Water	N6191-03	pH	Cool 4 deg C	STER09	N13	CELL-D	12/07/2022	9040C

Date/Time 12.23.22 10:20
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 12.23.22 13:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]