



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 12/27/2022

OVENTEMP IN Celsius(°C): 107
Time IN: 16:00
In Date: 12/23/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:22
Out Date: 12/24/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123422

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N6196-01	COMP-3049	1	1.14	8.48	9.62	8.77	90.0	
N6196-06	EPH-3049	2	1.14	8.48	9.62	8.77	90.0	
N6201-01	WALL-1	3	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N6202-01	MGC-COMP-1	4	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N6202-02	UGC-COMP-2	5	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
N6203-01	PCB-01	6	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N6203-02	PCB-02	7	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N6204-01	PCB-01	8	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N6204-02	PCB-02	9	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N6204-03	PCB-03	10	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N6204-04	PCB-04	11	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N6204-05	PCB-05	12	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N6204-06	PCB-06	13	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N6204-07	PCB-07	14	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N6204-08	PCB-08	15	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
N6205-01	CL-02-122322	16	1.15	8.80	9.95	8.7	85.8	
N6205-02	CL-02-122322-EPH-2	17	1.15	8.80	9.95	8.7	85.8	
N6208-01	HD-01-122322	18	1.13	8.84	9.97	9.1	90.2	
N6208-02	HD-01-122322-EPH-2	19	1.13	8.84	9.97	9.1	90.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

12/23/22

WorkList Name : %1-122322

WorkList ID : 166084

Department : Wet-Chemistry

Date : 12-23-2022 08:37:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/29/2022	Solid	N6196-01	Percent Solids	Cool 4 deg C	SOUN03	N13	COMP-3049	12/21/2022	Chemtech -SO
12/29/2022	Solid	N6196-06	Percent Solids	Cool 4 deg C	SOUN03	N13	EPH-3049	12/21/2022	Chemtech -SO
12/26/2022	Solid	N6201-01	Percent Solids	Cool 4 deg C	PSEG03	N13	WALL-1	12/23/2022	Chemtech -SO
12/28/2022	Solid	N6202-01	Percent Solids	Cool 4 deg C	PSEG03	N13	MGC-COMP-1	12/23/2022	Chemtech -SO
12/28/2022	Solid	N6202-02	Percent Solids	Cool 4 deg C	PSEG03	N13	UGC-COMP-2	12/23/2022	Chemtech -SO
12/30/2022	Solid	N6203-01	Percent Solids	Cool 4 deg C	CLOU03	N13	PCB-01	12/22/2022	Chemtech -SO
12/30/2022	Solid	N6203-02	Percent Solids	Cool 4 deg C	CLOU03	N13	PCB-02	12/22/2022	Chemtech -SO
12/30/2022	Solid	N6204-01	Percent Solids	Cool 4 deg C	CLOU03	N13	PCB-01	12/22/2022	Chemtech -SO
12/30/2022	Solid	N6204-02	Percent Solids	Cool 4 deg C	CLOU03	N13	PCB-02	12/22/2022	Chemtech -SO
12/30/2022	Solid	N6204-03	Percent Solids	Cool 4 deg C	CLOU03	N13	PCB-03	12/22/2022	Chemtech -SO
12/30/2022	Solid	N6204-04	Percent Solids	Cool 4 deg C	CLOU03	N13	PCB-04	12/22/2022	Chemtech -SO
12/30/2022	Solid	N6204-05	Percent Solids	Cool 4 deg C	CLOU03	N13	PCB-05	12/22/2022	Chemtech -SO
12/30/2022	Solid	N6204-06	Percent Solids	Cool 4 deg C	CLOU03	N13	PCB-06	12/22/2022	Chemtech -SO
12/30/2022	Solid	N6204-07	Percent Solids	Cool 4 deg C	CLOU03	N13	PCB-07	12/22/2022	Chemtech -SO
12/30/2022	Solid	N6204-08	Percent Solids	Cool 4 deg C	CLOU03	N13	PCB-08	12/22/2022	Chemtech -SO
12/28/2022	Solid	N6205-01	Percent Solids	Cool 4 deg C	PSEG05	N13	CL-02-122322	12/23/2022	Chemtech -SO
12/28/2022	Solid	N6205-02	Percent Solids	Cool 4 deg C	PSEG05	N13	CL-02-122322-EPH-2	12/23/2022	Chemtech -SO
12/29/2022	Solid	N6208-01	Percent Solids	Cool 4 deg C	PSEG05	N13	HD-01-122322	12/23/2022	Chemtech -SO
12/29/2022	Solid	N6208-02	Percent Solids	Cool 4 deg C	PSEG05	N13	HD-01-122322-EPH-2	12/23/2022	Chemtech -SO

JP 12.23.22 15:40 14140

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: