

LB123425


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

42307



A= 0.0000e+000

B= 2.3552e-004

C= 4.4165e-002

Rho= 0.9998631

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.062	0.062	77	0.000	77					
0.2	0.200	0.222	0.022	757	0.0 %	757					
2.5	2.500	2.369	-0.131	9872	0.0 %	9872					
5.0	5.000	5.026	0.026	21154	0.0 %	21154					
7.5	7.500	7.511	0.011	31704	0.0 %	31704					
10.0	10.000	10.008	0.008	42307	0.0 %	42307					

0/6 D
11
-5
-1
00
0

LB123425 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Dec 2022 15:23:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S6	77	-	0.0000	7470A	PPB	23 Dec 2022 15:25:25	S	Std
0.2 - 1	SP.2	757	-	0.2000	7470A	PPB	23 Dec 2022 15:27:43	S	Std
2.5 - 1	SP.5	9872	-	2.5000	7470A	PPB	23 Dec 2022 15:30:00	S	Std
5.0 - 1	SP.5	21154	-	5.0000	7470A	PPB	23 Dec 2022 15:32:16	S	Std
7.5 - 1	SP.5	31704	-	7.5000	7470A	PPB	23 Dec 2022 15:34:33	S	Std
10.0 - 1	SP.10	42307	-	10.0000	7470A	PPB	23 Dec 2022 15:36:49	S	Std
ICV73 - 1	ICV73	16544	3.9407	-	7470A	PPB	23 Dec 2022 15:40:08	U	SMPL
ICB73 - 1	ICB73	73	0.0614	-	7470A	PPB	23 Dec 2022 15:42:23	U	SMPL
CCV34 - 1	CCV34	20649	4.9075	-	7470A	PPB	23 Dec 2022 15:44:38	U	SMPL
CCB34 - 1	CCB34	50	0.0559	-	7470A	PPB	23 Dec 2022 15:46:53	U	SMPL
CRA - 1	CRA	706	0.2104	-	7470A	PPB	23 Dec 2022 15:49:08	U	SMPL
HighStd - 1	HighStd	40921	9.6820	-	7470A	PPB	23 Dec 2022 15:51:23	U	SMPL
ChkStd - 1	ChkStd	28750	6.8154	-	7470A	PPB	23 Dec 2022 15:53:39	U	SMPL
PB149890BL - 1	PBW	0	0.0442	-	7470A	PPB	23 Dec 2022 15:55:54	U	SMPL
PB149890BS - 1	LCSW	13757	3.2842	-	7470A	PPB	23 Dec 2022 15:58:10	U	SMPL
N6051-02 - 1	SP-591	115	0.0713	-	7470A	PPB	23 Dec 2022 16:00:26	U	SMPL
N6051-02DUP - 1	SP-591DUP	67	0.0599	-	7470A	PPB	23 Dec 2022 16:02:42	U	SMPL
N6051-02MS - 1	SP-591MS	15238	3.6331	-	7470A	PPB	23 Dec 2022 16:07:24	U	SMPL
N6051-02MSD - 1	SP-591MSD	14810	3.5323	-	7470A	PPB	23 Dec 2022 16:09:41	U	SMPL
CCV35 - 1	CCV35	21571	5.1246	-	7470A	PPB	23 Dec 2022 16:17:50	U	SMPL
CCB35 - 1	CCB35	48	0.0555	-	7470A	PPB	23 Dec 2022 16:20:05	U	SMPL
PB149841TB - 1	PB149841TB	114	0.0710	-	7470A	PPB	23 Dec 2022 16:22:21	U	SMPL
N6051-02LX5 - 1		2	0.0446	-	7470A	PPB	23 Dec 2022 16:24:36	U	SMPL
N6051-02A - 1		17558	4.1795	-	7470A	PPB	23 Dec 2022 16:26:51	U	SMPL
CCV36 - 1	CCV36	21427	5.0907	-	7470A	PPB	23 Dec 2022 16:29:06	U	SMPL
CCB36 - 1	CCB36	-37	0.0355	-	7470A	PPB	23 Dec 2022 16:31:22	U	SMPL