

LB123429


SFAM01.1

INSTRUMENT ID: LV1

Linear

 μ Abs.:

4166E



A= 0.0000e+000

B= 2.3871e-004 *slope*C= -1.0051e-001 *y-intercept*

Rho= 0.9993608

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% ()
0.00	0.000	0.046	0.046	615	0.000	615					1
0.05	0.050					185	0				-2
0.20	0.200	0.197	-0.003	1246	0.0 %	1246					-7
2.50	2.500	2.316	-0.184	10124	0.0 %	10124					3
5.00	5.000	5.125	0.125	21891	0.0 %	21891					2
7.50	7.500	7.670	0.170	32553	0.0 %	32553					-2
10.0	10.000	9.845	-0.155	41666	0.0 %	41666					

LB123429

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	615	-		0 SFAM01.1	PPB	12/25/2022 14:38	Std
	0.2 S01	1246	-		0.2 SFAM01.1	PPB	12/25/2022 14:41	Std
	2.5 S02	10124	-		2.5 SFAM01.1	PPB	12/25/2022 14:43	Std
	5 S03	21891	-		5 SFAM01.1	PPB	12/25/2022 14:45	Std
	7.5 S04	32553	-		7.5 SFAM01.1	PPB	12/25/2022 14:47	Std
	10 S05	41666	-		10 SFAM01.1	PPB	12/25/2022 14:50	Std
ICV074	ICV074	18156	4.2335 -		SFAM01.1	PPB	12/25/2022 14:53	SMPL
ICB704	ICB074	584	0.0389 -		SFAM01.1	PPB	12/25/2022 14:55	SMPL
CCV037	CCV037	22075	5.169 -		SFAM01.1	PPB	12/25/2022 14:57	SMPL
CCB037	CCB037	473	0.0124 -		SFAM01.1	PPB	12/25/2022 14:59	SMPL
PB149901BL	PBW901	316	-0.0251 -		SFAM01.1	PPB	12/25/2022 15:03	SMPL
N5954-05	MF5N14	498	0.0184 -		SFAM01.1	PPB	12/25/2022 15:15	SMPL
N5954-06	MF5N22	-3	-0.1012 -		SFAM01.1	PPB	12/25/2022 15:17	SMPL
N5954-07	MF5N23	47	-0.0893 -		SFAM01.1	PPB	12/25/2022 15:19	SMPL
N5954-08	MF5N24	545	0.0296 -		SFAM01.1	PPB	12/25/2022 15:22	SMPL
N5954-09	MF5N25	375	-0.011 -		SFAM01.1	PPB	12/25/2022 15:24	SMPL
N5954-10	MF5N27	63	-0.0855 -		SFAM01.1	PPB	12/25/2022 15:26	SMPL
N5954-11	MF5N29	461	0.0095 -		SFAM01.1	PPB	12/25/2022 15:28	SMPL
N5954-12	MF5N31	583	0.0387 -		SFAM01.1	PPB	12/25/2022 15:31	SMPL
N5954-13	MF5N38	519	0.0234 -		SFAM01.1	PPB	12/25/2022 15:33	SMPL
N5954-14	MF5N39	514	0.0222 -		SFAM01.1	PPB	12/25/2022 15:35	SMPL
PB149902BL	PBW902	544	0.0293 -		SFAM01.1	PPB	12/25/2022 15:37	SMPL
N5955-01	MF5N54	532	0.0265 -		SFAM01.1	PPB	12/25/2022 15:40	SMPL
N5955-02	MF5N54D	526	0.025 -		SFAM01.1	PPB	12/25/2022 15:42	SMPL
N5955-03	MF5N54S	4422	0.9551 -		SFAM01.1	PPB	12/25/2022 15:44	SMPL
N5955-04	MF5N55	516	0.0227 -		SFAM01.1	PPB	12/25/2022 15:46	SMPL
CCV038	CCV038	22357	5.2363 -		SFAM01.1	PPB	12/25/2022 15:49	SMPL
CCB308	CCB038	492	0.0169 -		SFAM01.1	PPB	12/25/2022 15:51	SMPL
N5955-05	MF5N56	111	-0.074 -		SFAM01.1	PPB	12/25/2022 15:53	SMPL
N5955-06	MF5N64	23	-0.095 -		SFAM01.1	PPB	12/25/2022 15:56	SMPL
N5955-07	MF5N65	38	-0.0914 -		SFAM01.1	PPB	12/25/2022 15:58	SMPL
N5955-08	MF5N66	15	-0.0969 -		SFAM01.1	PPB	12/25/2022 16:00	SMPL
N5955-09	MF5N67	101	-0.0764 -		SFAM01.1	PPB	12/25/2022 16:02	SMPL
N5955-10	MF5N69	82	-0.0809 -		SFAM01.1	PPB	12/25/2022 16:05	SMPL
N5955-11	MF5N71	17	-0.0965 -		SFAM01.1	PPB	12/25/2022 16:07	SMPL
N5955-12	MF5N73	169	-0.0602 -		SFAM01.1	PPB	12/25/2022 16:09	SMPL
N5955-13	MF5N80	-78	-0.1191 -		SFAM01.1	PPB	12/25/2022 16:11	SMPL
N5955-14	MF5N81	-10	-0.1029 -		SFAM01.1	PPB	12/25/2022 16:14	SMPL
PB149903BL	PBW903	501	0.0191 -		SFAM01.1	PPB	12/25/2022 16:16	SMPL
N5956-01	MF5N74	460	0.0093 -		SFAM01.1	PPB	12/25/2022 16:18	SMPL
N5956-02	MF5N74D	535	0.0272 -		SFAM01.1	PPB	12/25/2022 16:20	SMPL
N5956-03	MF5N74S	4426	0.956 -		SFAM01.1	PPB	12/25/2022 16:23	SMPL
N5956-05	MF5N42	196	-0.0537 -		SFAM01.1	PPB	12/25/2022 16:25	SMPL
N5956-06	MF5N43	199	-0.053 -		SFAM01.1	PPB	12/25/2022 16:27	SMPL
N5956-07	MF5N44	0	-0.1005 -		SFAM01.1	PPB	12/25/2022 16:30	SMPL
N5956-08	MF5N45	97	-0.0774 -		SFAM01.1	PPB	12/25/2022 16:32	SMPL

LB123429

INSTRUMENT ID : CV1

N5956-09	MF5N46	570	0.0356 -	SFAM01.1	PPB	12/25/2022 16:34 SMPL
N5956-10	MF5N47	53	-0.0879 -	SFAM01.1	PPB	12/25/2022 16:36 SMPL
CCV039	CCV039	21965	5.1427 -	SFAM01.1	PPB	12/25/2022 16:39 SMPL
CCB039	CCB039	458	0.0088 -	SFAM01.1	PPB	12/25/2022 16:41 SMPL
PB149904BL	PBW904	501	0.0191 -	SFAM01.1	PPB	12/25/2022 16:43 SMPL
N5957-01	MF5N32	516	0.0227 -	SFAM01.1	PPB	12/25/2022 16:45 SMPL
N5957-02	MF5N32D	531	0.0262 -	SFAM01.1	PPB	12/25/2022 16:48 SMPL
N5957-03	MF5N32S	4530	0.9808 -	SFAM01.1	PPB	12/25/2022 16:50 SMPL
N5957-04	MF5N00	95	-0.0778 -	SFAM01.1	PPB	12/25/2022 16:52 SMPL
N5957-05	MF5N01	106	-0.0752 -	SFAM01.1	PPB	12/25/2022 16:54 SMPL
N5957-06	MF5N02	-17	-0.1046 -	SFAM01.1	PPB	12/25/2022 16:57 SMPL
N5957-07	MF5N03	121	-0.0716 -	SFAM01.1	PPB	12/25/2022 16:59 SMPL
N5957-08	MF5N04	355	-0.0158 -	SFAM01.1	PPB	12/25/2022 17:01 SMPL
N5957-09	MF5N05	37	-0.0917 -	SFAM01.1	PPB	12/25/2022 17:03 SMPL
PB149905BL	PBW905	496	0.0179 -	SFAM01.1	PPB	12/25/2022 17:06 SMPL
N6134-01	MEXGJ1	398	-0.0055 -	SFAM01.1	PPB	12/25/2022 17:08 SMPL
N6134-02	MEXGJ1D	338	-0.0198 -	SFAM01.1	PPB	12/25/2022 17:10 SMPL
N6134-03	MEXGJ1S	4339	0.9352 -	SFAM01.1	PPB	12/25/2022 17:13 SMPL
N5954-01	MF5N12	668	0.0589 -	SFAM01.1	PPB	12/25/2022 17:15 SMPL
N5954-01	MF5N12D	613	0.0458 -	SFAM01.1	PPB	12/25/2022 17:17 SMPL
N5954-01	MF5N12S	4620	1.0023 -	SFAM01.1	PPB	12/25/2022 17:19 SMPL
N5954-04	MF5N13	615	0.0463 -	SFAM01.1	PPB	12/25/2022 17:22 SMPL
CCV040	CCV040	22486	5.2671 -	SFAM01.1	PPB	12/25/2022 17:24 SMPL
CCB040	CCB040	460	0.0093 -	SFAM01.1	PPB	12/25/2022 17:26 SMPL