

LB123438


7470A

INSTRUMENT ID: CVL

Linear

 $\mu$  Abs.:

40833



A= 0.0000e+000

B= 2.3983e-004

C= 3.0037e-004

Rho= 0.9992973

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D  |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|-----|
| 0.0    | 0.000  | 0.002 | 0.002  | 6     | 0.000      | 6     |       |       |       |       | -   |
| 0.2    | 0.200  | 0.178 | -0.022 | 742   | 0.0 %      | 742   |       |       |       |       | -11 |
| 2.5    | 2.500  | 2.412 | -0.088 | 10054 | 0.0 %      | 10054 |       |       |       |       | -4  |
| 5.0    | 5.000  | 5.074 | 0.074  | 21157 | 0.0 %      | 21157 |       |       |       |       | 1   |
| 7.5    | 7.500  | 7.741 | 0.241  | 32277 | 0.0 %      | 32277 |       |       |       |       | 3   |
| 10.0   | 10.000 | 9.793 | -0.207 | 40833 | 0.0 %      | 40833 |       |       |       |       | -2  |

## LB123438 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Dec 2022 09:51:09

| Sample ID       | Extended ID            | μ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|------------------------|--------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50                     | 6      | -       | 0.0000   | 7470A  | PPB   | 27 Dec 2022 10:27:51 | S    | Std  |
| 0.2 - 1         | 50.2                   | 742    | -       | 0.2000   | 7470A  | PPB   | 27 Dec 2022 10:30:09 | S    | Std  |
| 2.5 - 1         | 52.5                   | 10054  | -       | 2.5000   | 7470A  | PPB   | 27 Dec 2022 10:32:27 | S    | Std  |
| 5.0 - 1         | 55                     | 21157  | -       | 5.0000   | 7470A  | PPB   | 27 Dec 2022 10:34:43 | S    | Std  |
| 7.5 - 1         | 59.5                   | 32277  | -       | 7.5000   | 7470A  | PPB   | 27 Dec 2022 10:37:00 | S    | Std  |
| 10.0 - 1        | 510                    | 40833  | -       | 10.0000  | 7470A  | PPB   | 27 Dec 2022 10:39:16 | S    | Std  |
| ICV76 - 1       | ICV76                  | 17142  | 4.1114  | -        | 7470A  | PPB   | 27 Dec 2022 10:42:02 | U    | SMPL |
| ICB76 - 1       | ICB76                  | 41     | 0.0101  | -        | 7470A  | PPB   | 27 Dec 2022 10:44:17 | U    | SMPL |
| CCV41 - 1       | CCV41                  | 21574  | 5.1743  | -        | 7470A  | PPB   | 27 Dec 2022 10:46:32 | U    | SMPL |
| CCB41 - 1       | CCB41                  | -6     | -0.0011 | -        | 7470A  | PPB   | 27 Dec 2022 10:48:46 | U    | SMPL |
| CRA - 1         | CRA                    | 755    | 0.1814  | -        | 7470A  | PPB   | 27 Dec 2022 10:51:01 | U    | SMPL |
| HighStd - 1     | HighStd                | 41146  | 9.8681  | -        | 7470A  | PPB   | 27 Dec 2022 10:53:16 | U    | SMPL |
| ChkStd - 1      | ChkStd                 | 29999  | 7.1948  | -        | 7470A  | PPB   | 27 Dec 2022 10:55:32 | U    | SMPL |
| PB149891BL - 1  | PBW                    | -29    | -0.0067 | -        | 7470A  | PPB   | 27 Dec 2022 10:57:47 | U    | SMPL |
| PB149891BS - 1  | LCSW                   | 17636  | 4.2299  | -        | 7470A  | PPB   | 27 Dec 2022 11:00:03 | U    | SMPL |
| N6179-02 - 1    | T-1-TRANSFORMER-DEBRIS | 26     | 0.0065  | -        | 7470A  | PPB   | 27 Dec 2022 11:02:19 | U    | SMPL |
| N6180-02 - 1    | COMP-3050              | 17     | 0.0044  | -        | 7470A  | PPB   | 27 Dec 2022 11:04:36 | U    | SMPL |
| N6196-02 - 1    | COMP-3049              | 36     | 0.0089  | -        | 7470A  | PPB   | 27 Dec 2022 11:06:52 | U    | SMPL |
| N6196-02DUP - 1 | COMP-3049DUP           | -3     | -0.0004 | -        | 7470A  | PPB   | 27 Dec 2022 11:09:09 | U    | SMPL |
| N6196-02MS - 1  | COMP-3049MS            | 14936  | 3.5823  | -        | 7470A  | PPB   | 27 Dec 2022 11:11:24 | U    | SMPL |
| CCV42 - 1       | CCV42                  | 21579  | 5.1755  | -        | 7470A  | PPB   | 27 Dec 2022 11:13:40 | U    | SMPL |
| CCB42 - 1       | CCB42                  | -39    | -0.0091 | -        | 7470A  | PPB   | 27 Dec 2022 11:15:55 | U    | SMPL |
| N6196-02MSD - 1 | COMP-3049MSD           | 14988  | 3.5948  | -        | 7470A  | PPB   | 27 Dec 2022 11:18:10 | U    | SMPL |
| PB149894BL - 1  | PBW                    | 12     | 0.0032  | -        | 7470A  | PPB   | 27 Dec 2022 11:20:26 | U    | SMPL |
| PB149894BS - 1  | LCSW                   | 16306  | 3.9109  | -        | 7470A  | PPB   | 27 Dec 2022 11:22:42 | U    | SMPL |
| N6165-01 - 1    | 12x5                   | 492    | 0.1183  | -        | 7470A  | PPB   | 27 Dec 2022 11:24:57 | U    | SMPL |
| N6165-02 - 1    | 12x6                   | 169    | 0.0408  | -        | 7470A  | PPB   | 27 Dec 2022 11:27:13 | U    | SMPL |
| N6165-03 - 1    | 12x7                   | 109    | 0.0264  | -        | 7470A  | PPB   | 27 Dec 2022 11:29:29 | U    | SMPL |
| N6165-04 - 1    | 12x8                   | 62     | 0.0152  | -        | 7470A  | PPB   | 27 Dec 2022 11:31:45 | U    | SMPL |
| N6165-05 - 1    | 12x9                   | 748    | 0.1797  | -        | 7470A  | PPB   | 27 Dec 2022 11:34:02 | U    | SMPL |
| N6165-06 - 1    | 12x10                  | 1005   | 0.2413  | -        | 7470A  | PPB   | 27 Dec 2022 11:36:19 | U    | SMPL |
| N6165-07 - 1    | 12x11                  | 1122   | 0.2694  | -        | 7470A  | PPB   | 27 Dec 2022 11:38:34 | U    | SMPL |
| CCV43 - 1       | CCV43                  | 21673  | 5.1980  | -        | 7470A  | PPB   | 27 Dec 2022 11:40:50 | U    | SMPL |
| CCB43 - 1       | CCB43                  | -9     | -0.0019 | -        | 7470A  | PPB   | 27 Dec 2022 11:43:06 | U    | SMPL |
| N6165-08 - 1    | 12x12                  | 209    | 0.0504  | -        | 7470A  | PPB   | 27 Dec 2022 11:45:21 | U    | SMPL |
| N6165-08DUP - 1 | 12x12DUP               | 166    | 0.0401  | -        | 7470A  | PPB   | 27 Dec 2022 11:47:37 | U    | SMPL |
| N6165-08MS - 1  | 12x12MS                | 16512  | 3.9603  | -        | 7470A  | PPB   | 27 Dec 2022 11:49:52 | U    | SMPL |
| N6165-08MSD - 1 | 12x12MSD               | 16353  | 3.9222  | -        | 7470A  | PPB   | 27 Dec 2022 11:52:08 | U    | SMPL |
| PB149838TB - 1  | PB149838TB             | -17    | -0.0038 | -        | 7470A  | PPB   | 27 Dec 2022 11:54:24 | U    | SMPL |
| PB149840TB - 1  | PB149840TB             | 54     | 0.0133  | -        | 7470A  | PPB   | 27 Dec 2022 11:56:40 | U    | SMPL |
| N6196-02LX5 - 1 |                        | 21     | 0.0053  | -        | 7470A  | PPB   | 27 Dec 2022 11:58:56 | U    | SMPL |
| N6196-02A - 1   |                        | 17204  | 4.1263  | -        | 7470A  | PPB   | 27 Dec 2022 12:01:13 | U    | SMPL |
| N6165-08LX5 - 1 |                        | 185    | 0.0447  | -        | 7470A  | PPB   | 27 Dec 2022 12:03:30 | U    | SMPL |
| N6165-08A - 1   |                        | 17174  | 4.1191  | -        | 7470A  | PPB   | 27 Dec 2022 12:05:46 | U    | SMPL |
| CCV44 - 1       | CCV44                  | 21821  | 5.2335  | -        | 7470A  | PPB   | 27 Dec 2022 12:08:01 | U    | SMPL |
| CCB44 - 1       | CCB44                  | -3     | -0.0004 | -        | 7470A  | PPB   | 27 Dec 2022 12:10:17 | U    | SMPL |