

LB123444


7471B

INSTRUMENT ID: CV4

Linear

 μ Abs.:

52554



A= 0.0000e+000

B= 1.8787e-004

C= 4.1115e-002

Rho= 0.9996808

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.043	0.043	12	0.000	12					2
0.2	0.200	0.212	0.012	909	0.0 %	909					6
2.5	2.500	2.418	-0.082	12653	0.0 %	12653					-3
5.0	5.000	4.934	-0.066	26044	0.0 %	26044					-1
7.5	7.500	7.678	0.178	40651	0.0 %	40651					2
10.0	10.000	9.914	-0.086	52554	0.0 %	52554					-1

LB123444 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Dec 2022 12:35:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	12	-	0.0000	7471B	PPB	27 Dec 2022 12:38:23	S	Std
0.2 - 1	502	909	-	0.2000	7471B	PPB	27 Dec 2022 12:40:40	S	Std
2.5 - 1	545	12653	-	2.5000	7471B	PPB	27 Dec 2022 12:42:57	S	Std
5.0 - 1	55	26044	-	5.0000	7471B	PPB	27 Dec 2022 12:45:14	S	Std
7.5 - 1	595	40651	-	7.5000	7471B	PPB	27 Dec 2022 12:47:31	S	Std
10.0 - 1	510	52554	-	10.0000	7471B	PPB	27 Dec 2022 12:49:46	S	Std
ICV77 - 1	ICV77	21457	4.0722	-	7471B	PPB	27 Dec 2022 12:52:33	U	SMPL
ICB77 - 1	ICB77	17	0.0443	-	7471B	PPB	27 Dec 2022 12:54:48	U	SMPL
CCV45 - 1	CCV45	26393	4.9995	-	7471B	PPB	27 Dec 2022 12:57:03	U	SMPL
CCB45 - 1	CCB45	-6	0.0400	-	7471B	PPB	27 Dec 2022 12:59:18	U	SMPL
CRA - 1	CRA	918	0.2136	-	7471B	PPB	27 Dec 2022 13:01:33	U	SMPL
HighStd - 1	HighStd	51839	9.7801	-	7471B	PPB	27 Dec 2022 13:03:48	U	SMPL
ChkStd - 1	ChkStd	37248	7.0389	-	7471B	PPB	27 Dec 2022 13:06:02	U	SMPL
PB149908BL - 1	PBS	-76	0.0268	-	7471B	PPB	27 Dec 2022 13:08:19	U	SMPL
PB149908BS - 1	LCSS	17953	3.4139	-	7471B	PPB	27 Dec 2022 13:10:35	U	SMPL
N6190-01 - 1	LOCATION-1	2508	0.5123	-	7471B	PPB	27 Dec 2022 13:12:51	U	SMPL
N6190-02 - 1	LOCATION-2	1491	0.3212	-	7471B	PPB	27 Dec 2022 13:15:08	U	SMPL
N6190-03 - 1	LOCATION-3	341	0.1052	-	7471B	PPB	27 Dec 2022 13:17:25	U	SMPL
N6194-01 - 1	CONC-GIS-1	1142	0.2557	-	7471B	PPB	27 Dec 2022 13:19:41	U	SMPL
N6196-01 - 1	COMP-3049	241	0.0864	-	7471B	PPB	27 Dec 2022 13:21:57	U	SMPL
CCV46 - 1	CCV46	26244	4.9715	-	7471B	PPB	27 Dec 2022 13:24:12	U	SMPL
CCB46 - 1	CCB46	-56	0.0306	-	7471B	PPB	27 Dec 2022 13:26:28	U	SMPL
N6205-01 - 1	CL-02-122322	1632	0.3477	-	7471B	PPB	27 Dec 2022 13:28:43	U	SMPL
N6208-01 - 1	HD-01-122322	1387	0.3017	-	7471B	PPB	27 Dec 2022 13:30:58	U	SMPL
N6208-01DUP - 1	HD-01-122322DUP	1461	0.3156	-	7471B	PPB	27 Dec 2022 13:33:14	U	SMPL
N6208-01MS - 1	HD-01-122322MS	19509	3.7062	-	7471B	PPB	27 Dec 2022 13:35:30	U	SMPL
N6208-01MSD - 1	HD-01-122322MSD	19924	3.7842	-	7471B	PPB	27 Dec 2022 13:37:45	U	SMPL
N6208-01LX5 - 1		111	0.0620	-	7471B	PPB	27 Dec 2022 13:40:02	U	SMPL
N6208-01A - 1		21476	4.0758	-	7471B	PPB	27 Dec 2022 13:42:18	U	SMPL
CCV47 - 1	CCV47	26515	5.0225	-	7471B	PPB	27 Dec 2022 13:44:35	U	SMPL
CCB47 - 1	CCB47	-26	0.0362	-	7471B	PPB	27 Dec 2022 13:46:52	U	SMPL