



Analytical Summary Report

Analysis Method: 7196A
Parameter: Hexavalent Chromium
Run Number: LB123448

ANALYST: Sohil
SUPERVISOR REVIEW BY: Iwona
pH Meter ID: WC pH Meter-1

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include Calibration Std. hexchrome 0.1 ppm, 0.05 ppm, 0.01 ppm, 0 ppm, hexavalent chromium color reagent, 5N sulfuric acid, Hexchrome Cleaning Solution, Calibration Std Hexachrome 0.025 ppm, Hexavalent Chromium ICV-LCS Std, Calibration and CCV std HexChrome 0.5PPM, Calibration std HexChrome 1.0PPM.

Intercept: -0.0007 Slope: 0.7808 Regression: 0.999998

Table with 13 columns: Seq, Lab ID, True Value (mg/l), DF, Initial Vol (ml), Final Vol (ml), pH HN03, pH H2SO4, Absorb.at 540nm (Backgrnd, Color), Absorbance Difference, Result (mg/L), %D, Anal Date, Anal Time. Contains 7 rows of calibration data.



Analytical Summary Report

Analysis Method: 7196A

ANALYST: Sohil

Parameter: Hexavalent Chromium

SUPERVISOR REVIEW BY: Iwona

Run Number: LB123448

pH Meter ID: WC pH Meter-1

Seq	Lab ID	True Value	DF	Initial Vol (ml/gm)	Final Vol (ml)	pH HN03	pH H2SO4	Absorb.at540nm		Absorbance Difference	Intermediate Result (mg/L)	Anal Date	Anal Time
								Backgrnd	Color				
1	ICV	0.5	1	100	100		2.00	0.000	0.390	0.390	0.500	12/27/2022	11:22
2	ICB		1	100	100		1.96	0.000	0.000	0.000	0.001	12/27/2022	11:23
3	CCV1	0.5	1	100	100		2.04	0.000	0.392	0.392	0.503	12/27/2022	11:24
4	CCB1		1	100	100		1.88	0.000	0.000	0.000	0.001	12/27/2022	11:25
5	RL Check	0.01	1	100	100		2.05	0.000	0.008	0.008	0.011	12/27/2022	11:26
6	LB123448BL		1	100	100		1.90	0.000	0.000	0.000	0.001	12/27/2022	11:27
7	LB123448BS	0.5	1	100	100		1.88	0.000	0.392	0.392	0.503	12/27/2022	11:28
8	N6088-03		1	100	100		2.06	0.000	0.000	0.000	0.001	12/27/2022	11:29
9	N6088-03DU		1	100	100		2.04	0.000	0.000	0.000	0.001	12/27/2022	11:30
10	CCV2	0.5	1	100	100		2.06	0.000	0.395	0.395	0.507	12/27/2022	11:31
11	CCB2		1	100	100		2.00	0.0000	0.000	0.000	0.001	12/27/2022	11:32