



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123451

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	12/28/2022	14:50
2	N6211-02		1	12	100.00	>212.0	>212.0	12/28/2022	15:15
3	N6217-02		1	14	100.00	>212.0	>212.0	12/28/2022	15:45
4	N6217-02DUP		1	13	100.00	>212.0	>212.0	12/28/2022	16:15
5	N6221-02		1	14	100.00	>212.0	>212.0	12/28/2022	16:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66123454

WorkList Name : fp12-28

WorkList ID : 166157

Department : Wet-Chemistry

Date : 12-28-2022 07:55:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/03/2023	Water	N6211-02	Flash Point	Cool 4 deg C	PSEG03	N21	329	12/27/2022	1010B
01/03/2023	Water	N6217-02	Flash Point	Cool 4 deg C	PSEG03	N21	D2813	12/29/2022	1010B
01/03/2023	Water	N6221-02	Flash Point	Cool 4 deg C	PSEG03	N21	184	12/27/2022	1010B

Date/Time 12/28/2022 14:45
 Raw Sample Received by: RM CWD
 Raw Sample Relinquished by: RJ CWD

Date/Time 12/28/2022 16:50
 Raw Sample Received by: RJ CWD
 Raw Sample Relinquished by: RM CWD