

LB123462


7470A

INSTRUMENT ID: CVL

Linear

μ Abs.:

40512



A= 0.0000e+000

B= 2.4230e-004

C= 7.0213e-003

Rho= 0.9988051

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.074	0.074	276	0.000	276					13
0.2	0.200	0.226	0.026	902	0.0 %	902					12
2.5	2.500	2.204	-0.296	9066	0.0 %	9066					4
5.0	5.000	5.197	0.197	21421	0.0 %	21421					2
7.5	7.500	7.676	0.176	31651	0.0 %	31651					0
10.0	10.000	9.823	-0.177	40512	0.0 %	40512					

LB123462 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Dec 2022 14:03:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	56	276	-	0.0000	7470A	PPB	28 Dec 2022 14:14:09	S	Std
0.2 - 1	58-2	902	-	0.2000	7470A	PPB	28 Dec 2022 14:16:26	S	Std
2.5 - 1	52-5	9066	-	2.5000	7470A	PPB	28 Dec 2022 14:18:51	S	Std
5.0 - 1	55	21421	-	5.0000	7470A	PPB	28 Dec 2022 14:23:51	S	Std
7.5 - 1	57-5	31651	-	7.5000	7470A	PPB	28 Dec 2022 14:28:42	S	Std
10.0 - 1	51-1	40512	-	10.0000	7470A	PPB	28 Dec 2022 14:33:33	S	Std
ICV78 - 1	ICV78	17058	4.1402	-	7470A	PPB	28 Dec 2022 14:36:38	U	SMPL
ICB78 - 1	ICB78	189	0.0528	-	7470A	PPB	28 Dec 2022 14:38:53	U	SMPL
CCV48 - 1	CCV48	21926	5.3198	-	7470A	PPB	28 Dec 2022 14:41:08	U	SMPL
CCB48 - 1	CCB48	99	0.0310	-	7470A	PPB	28 Dec 2022 14:43:23	U	SMPL
CRA - 1	CRA	722	0.1820	-	7470A	PPB	28 Dec 2022 14:45:37	U	SMPL
HighStd - 1	HighStd	40421	9.8012	-	7470A	PPB	28 Dec 2022 14:47:52	U	SMPL
ChkStd - 1	ChkStd	28922	7.0149	-	7470A	PPB	28 Dec 2022 14:52:42	U	SMPL
PB149936BL - 1	PBW	6	0.0085	-	7470A	PPB	28 Dec 2022 14:54:57	U	SMPL
PB149936BS - 1	LCSW	14065	3.4150	-	7470A	PPB	28 Dec 2022 14:57:12	U	SMPL
N6207-01 - 1	EB-1	88	0.0283	-	7470A	PPB	28 Dec 2022 14:59:29	U	SMPL
N6207-02 - 1	EB-2	81	0.0266	-	7470A	PPB	28 Dec 2022 15:01:45	U	SMPL
N6207-03 - 1	EB-3	-10	0.0046	-	7470A	PPB	28 Dec 2022 15:04:02	U	SMPL
N6207-04 - 1	EB-4	49	0.0189	-	7470A	PPB	28 Dec 2022 15:06:19	U	SMPL
N6207-04DUP - 1	EB-4DUP	12	0.0099	-	7470A	PPB	28 Dec 2022 15:08:34	U	SMPL
CCV49 - 1	CCV49	21794	5.2878	-	7470A	PPB	28 Dec 2022 15:10:49	U	SMPL
CCB49 - 1	CCB49	101	0.0315	-	7470A	PPB	28 Dec 2022 15:13:04	U	SMPL
N6207-04MS - 1	EB-4MS	13831	3.3583	-	7470A	PPB	28 Dec 2022 15:15:20	U	SMPL
N6207-04LX5 - 1		-27	0.0005	-	7470A	PPB	28 Dec 2022 15:19:51	U	SMPL
N6207-04A - 1		17338	4.2081	-	7470A	PPB	28 Dec 2022 15:22:07	U	SMPL
PB149893TB - 1	PB149893TB	31	0.0145	-	7470A	PPB	28 Dec 2022 15:24:22	U	SMPL
N6207-04MSD - 1	EB-4MSD	13329	3.2367	-	7470A	PPB	28 Dec 2022 15:33:37	U	SMPL
CCV50 - 1	CCV50	21507	5.2182	-	7470A	PPB	28 Dec 2022 15:36:07	U	SMPL
CCB50 - 1	CCB50	34	0.0153	-	7470A	PPB	28 Dec 2022 15:38:24	U	SMPL