

7471B


LB123467

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46294



A= 0.0000e+000

B= 2.1051e-004

C= 1.8570e-002

Rho= 0.9981059

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.028	0.028	45	0.000	45					-
0.2	0.200	0.210	0.010	911	0.0 %	911					5
2.5	2.500	2.196	-0.304	10343	0.0 %	10343					-12
5.0	5.000	5.381	0.381	25476	0.0 %	25476					8
7.5	7.500	7.620	0.120	36112	0.0 %	36112					2
10.0	10.000	9.764	-0.236	46294	0.0 %	46294					-2

LB123467 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Dec 2022 16:23:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	45	-	0.0000	7471B	PPB	28 Dec 2022 16:35:54	S	Std
0.2 - 1	50.2	911	-	0.2000	7471B	PPB	28 Dec 2022 16:38:11	S	Std
2.5 - 1	52.5	10343	-	2.5000	7471B	PPB	28 Dec 2022 16:40:28	S	Std
5.0 - 1	55	25476	-	5.0000	7471B	PPB	28 Dec 2022 16:42:45	S	Std
7.5 - 1	57.5	36112	-	7.5000	7471B	PPB	28 Dec 2022 16:47:40	S	Std
10.0 - 1	50	46294	-	10.0000	7471B	PPB	28 Dec 2022 16:52:28	S	Std
ICV79 - 1	ICV79	19636	4.1521	-	7471B	PPB	28 Dec 2022 16:55:18	U	SMPL
ICB79 - 1	ICB79	45	0.0280	-	7471B	PPB	28 Dec 2022 16:57:32	U	SMPL
CCV51 - 1	CCV51	23943	5.0588	-	7471B	PPB	28 Dec 2022 16:59:48	U	SMPL
CCB51 - 1	CCB51	7	0.0200	-	7471B	PPB	28 Dec 2022 17:02:03	U	SMPL
CRA - 1	CRA	794	0.1857	-	7471B	PPB	28 Dec 2022 17:04:18	U	SMPL
HighStd - 1	HighStd	46530	9.8135	-	7471B	PPB	28 Dec 2022 17:06:32	U	SMPL
ChkStd - 1	ChkStd	32863	6.9365	-	7471B	PPB	28 Dec 2022 17:08:48	U	SMPL
PB149971BL - 1	PBS	-58	0.0064	-	7471B	PPB	28 Dec 2022 17:11:05	U	SMPL
PB149971BS - 1	LCSS	16981	3.5932	-	7471B	PPB	28 Dec 2022 17:16:00	U	SMPL
N6199-01 - 1	TR-604-COMP-1	722	0.1706	-	7471B	PPB	28 Dec 2022 17:18:16	U	SMPL
N6200-01 - 1	PL-502A-COMP-3	499	0.1236	-	7471B	PPB	28 Dec 2022 17:20:32	U	SMPL
N6212-01 - 1	TP-1	2291	0.5008	-	7471B	PPB	28 Dec 2022 17:22:49	U	SMPL
N6218-01 - 1	CHRT-26179	182	0.0569	-	7471B	PPB	28 Dec 2022 17:25:07	U	SMPL
N6218-03 - 1	CHRT-26223	2592	0.5642	-	7471B	PPB	28 Dec 2022 17:27:24	U	SMPL
CCV52 - 1	CCV52	23982	5.0670	-	7471B	PPB	28 Dec 2022 17:29:39	U	SMPL
CCB52 - 1	CCB52	-16	0.0152	-	7471B	PPB	28 Dec 2022 17:31:54	U	SMPL
N6218-05 - 1	CHRT-26598	1525	0.3396	-	7471B	PPB	28 Dec 2022 17:34:09	U	SMPL
N6219-01 - 1	RB-21147	4242	0.9115	-	7471B	PPB	28 Dec 2022 17:36:24	U	SMPL
N6219-01DUP - 1	RB-21147DUP	3623	0.7812	-	7471B	PPB	28 Dec 2022 17:38:40	U	SMPL
N6219-01MS - 1	RB-21147MS	19887	4.2049	-	7471B	PPB	28 Dec 2022 17:40:56	U	SMPL
N6219-01MSD - 1	RB-21147MSD	20841	4.4058	-	7471B	PPB	28 Dec 2022 17:43:12	U	SMPL
N6219-04 - 1	72-12016	3388	0.7318	-	7471B	PPB	28 Dec 2022 17:45:28	U	SMPL
N6220-01 - 1	TP-1	9649	2.0498	-	7471B	PPB	28 Dec 2022 17:47:45	U	SMPL
N6220-04 - 1	TP-CR	7425	1.5816	-	7471B	PPB	28 Dec 2022 17:50:02	U	SMPL
N6219-01LX5 - 1		747	0.1758	-	7471B	PPB	28 Dec 2022 17:52:19	U	SMPL
N6219-01A - 1		21767	4.6007	-	7471B	PPB	28 Dec 2022 17:54:35	U	SMPL
CCV53 - 1	CCV53	23951	5.0605	-	7471B	PPB	28 Dec 2022 17:56:50	U	SMPL
CCB53 - 1	CCB53	-35	0.0112	-	7471B	PPB	28 Dec 2022 17:59:06	U	SMPL