

LB123470


SFAM01.1

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

44045



A= 0.0000e+000

B= 2.2712e-004 slope

C= 5.9066e-003 y-intercept

Rho= 0.9999697

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.013	0.013	32	0.000	32					
0.05	0.050					185	0				
0.20	0.200	0.210	0.010	899	0.0 %	899					
2.50	2.500	2.451	-0.049	10766	0.0 %	10766					
5.00	5.000	5.040	0.040	22166	0.0 %	22166					
7.50	7.500	7.476	-0.024	32891	0.0 %	32891					
10.0	10.000	10.009	0.009	44045	0.0 %	44045					

1511-2000

LB123470
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	32	-		0 SFAM01.1	PPB	12/28/2022 18:40	Std
	0.2 S01	899	-		0.2 SFAM01.1	PPB	12/28/2022 18:43	Std
	2.5 S02	10766	-		2.5 SFAM01.1	PPB	12/28/2022 18:45	Std
	5 S03	22166	-		5 SFAM01.1	PPB	12/28/2022 18:47	Std
	7.5 S04	32891	-		7.5 SFAM01.1	PPB	12/28/2022 18:49	Std
	10 S05	44045	-		10 SFAM01.1	PPB	12/28/2022 18:52	Std
ICV080	ICV080	18277	4.157 -		SFAM01.1	PPB	12/28/2022 18:55	SMPL
ICB080	ICB080	60	0.0195 -		SFAM01.1	PPB	12/28/2022 18:57	SMPL
CCV054	CCV054	22379	5.0886 -		SFAM01.1	PPB	12/28/2022 18:59	SMPL
CCB054	CCB054	12	0.0086 -		SFAM01.1	PPB	12/28/2022 19:01	SMPL
PB149956BL	PBS956	157	0.0416 -		SFAM01.1	PPB	12/28/2022 19:04	SMPL
N6077-01	GBP37	723	0.1701 -		SFAM01.1	PPB	12/28/2022 19:06	SMPL
N6077-02	GBP38	628	0.1485 -		SFAM01.1	PPB	12/28/2022 19:08	SMPL
N6077-03	GBP39	1940	0.4465 -		SFAM01.1	PPB	12/28/2022 19:10	SMPL
N6077-04	GBP40	539	0.1283 -		SFAM01.1	PPB	12/28/2022 19:13	SMPL
N6077-05	GBP41	779	0.1828 -		SFAM01.1	PPB	12/28/2022 19:15	SMPL
N6077-06	GBP41D	754	0.1772 -		SFAM01.1	PPB	12/28/2022 19:17	SMPL
N6077-07	GBP41S	11724	2.6687 -		SFAM01.1	PPB	12/28/2022 19:19	SMPL
N6077-10	GBP44	809	0.1896 -		SFAM01.1	PPB	12/28/2022 19:22	SMPL
N6077-12	GBP46	530	0.1263 -		SFAM01.1	PPB	12/28/2022 19:24	SMPL
N6077-13	GBP47	938	0.2189 -		SFAM01.1	PPB	12/28/2022 19:26	SMPL
N6077-15	GBP49	539	0.1283 -		SFAM01.1	PPB	12/28/2022 19:28	SMPL
N6077-16	GBP50	452	0.1086 -		SFAM01.1	PPB	12/28/2022 19:31	SMPL
N6077-17	GBP51	463	0.1111 -		SFAM01.1	PPB	12/28/2022 19:33	SMPL
N6077-19	GBP53	923	0.2155 -		SFAM01.1	PPB	12/28/2022 19:35	SMPL
N6077-20	GBP54	951	0.2219 -		SFAM01.1	PPB	12/28/2022 19:38	SMPL
PB149958BL	PBS958	20	0.0104 -		SFAM01.1	PPB	12/28/2022 19:40	SMPL
N6106-02	GBP19	216	0.055 -		SFAM01.1	PPB	12/28/2022 19:42	SMPL
N6106-03	GBP19D	236	0.0595 -		SFAM01.1	PPB	12/28/2022 19:44	SMPL
N6106-04	GBP19S	11405	2.5962 -		SFAM01.1	PPB	12/28/2022 19:47	SMPL
CCV055	CCV055	22246	5.0584 -		SFAM01.1	PPB	12/28/2022 19:49	SMPL
CCB055	CCB055	8	0.0077 -		SFAM01.1	PPB	12/28/2022 19:51	SMPL
N6106-05	GBP20	586	0.139 -		SFAM01.1	PPB	12/28/2022 19:53	SMPL
N6106-06	GBP21	1772	0.4084 -		SFAM01.1	PPB	12/28/2022 19:56	SMPL
N6106-07	GBP22	180	0.0468 -		SFAM01.1	PPB	12/28/2022 19:58	SMPL
N6106-08	GBP23	150	0.04 -		SFAM01.1	PPB	12/28/2022 20:00	SMPL
N6106-09	GBP24	362	0.0881 -		SFAM01.1	PPB	12/28/2022 20:02	SMPL
N6106-11	GBP26	170	0.0445 -		SFAM01.1	PPB	12/28/2022 20:05	SMPL
N6106-12	GBP27	424	0.1022 -		SFAM01.1	PPB	12/28/2022 20:07	SMPL
N6106-14	GBP29	838	0.1962 -		SFAM01.1	PPB	12/28/2022 20:09	SMPL
N6106-15	GBP30	979	0.2283 -		SFAM01.1	PPB	12/28/2022 20:12	SMPL
N6106-16	GBP31	888	0.2076 -		SFAM01.1	PPB	12/28/2022 20:14	SMPL
N6106-17	GBP32	3696	0.8453 -		SFAM01.1	PPB	12/28/2022 20:16	SMPL
N6106-19	GBP34	1549	0.3577 -		SFAM01.1	PPB	12/28/2022 20:18	SMPL
N6106-21	GBP36	280	0.0695 -		SFAM01.1	PPB	12/28/2022 20:21	SMPL
CCV056	CCV056	22498	5.1156 -		SFAM01.1	PPB	12/28/2022 20:23	SMPL

LB123470

INSTRUMENT ID : CV1

CCB056

CCB056

0 0.0059 -

SFAM01.1 PPB

12/28/2022 20:25 SMPL