



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/3/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:15
In Date: 12/29/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:05
Out Date: 12/30/2022
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123483

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
N6113-01	GBP99	1	1.12	2.50	3.62	3.57	98.0	
N6113-02	GBPA0	2	1.00	1.00	2.00	2.00	100.0	
N6113-03	GBPA1	3	1.13	8.66	9.79	9.65	98.4	
N6113-04	GBPA2	4	1.14	8.48	9.62	9.44	97.9	
N6113-05	GBPA3	5	1.14	4.11	5.25	5.12	96.8	
N6113-06	GBPA4	6	1.13	8.37	9.5	9.42	99.0	
N6113-07	GBPA5	7	1.13	8.48	9.61	9.47	98.3	
N6113-08	GBPA6	8	1.14	7.97	9.11	9.05	99.2	
N6113-09	GBPA7	9	1.13	8.50	9.63	9.58	99.4	
N6113-10	GBPA7D	10	1.13	8.50	9.63	9.58	99.4	
N6113-11	GBPA7S	11	1.13	8.50	9.63	9.58	99.4	
N6113-12	GBPA8	12	1.14	8.73	9.87	9.74	98.5	
N6113-13	GBPA9	13	1.14	8.54	9.68	9.66	99.8	
N6113-14	GBPB0	14	1.15	8.37	9.52	9.48	99.5	
N6113-15	GBPB1	15	1.15	8.66	9.81	9.68	98.5	
N6113-16	GBPB2	16	1.00	1.00	2.00	2.00	100.0	
N6113-17	GBPB3	17	1.00	1.00	2.00	2.00	100.0	
N6113-18	GBPB4	18	1.13	8.68	9.81	9.63	97.9	
N6113-19	GBPB5	19	1.14	8.56	9.7	9.67	99.6	
N6113-20	GBPB6	20	1.13	8.51	9.64	9.63	99.9	
N6113-21	GBPB7	21	1.13	8.48	9.61	9.58	99.6	
N6113-22	GBPH5	22	1.00	1.00	2.00	2.00	100.0	
N6113-23	GBPH9	23	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$