

LB123489


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

4341E



A= 0.0000e+000

B= 2.3110e-004 slope

C= 3.6649e-002 y-intercept

Rho= 0.9999026

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.067	0.067	131	0.000	131					=
0.05	0.050					185					-6
0.20	0.200	0.188	-0.012	655	0.0 %	655	0				-1
2.50	2.500	2.468	-0.032	10521	0.0 %	10521					-1
5.00	5.000	4.975	-0.025	21369	0.0 %	21369					-1
7.50	7.500	7.432	-0.068	31999	0.0 %	31999					-1
10.0	10.000	10.070	0.070	43418	0.0 %	43418					+1

LB123489

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	131	-		0 SFAM01.1	PPB	12/29/2022 15:14	Std
	0.2 S01	655	-		0.2 SFAM01.1	PPB	12/29/2022 15:16	Std
	2.5 S02	10521	-		2.5 SFAM01.1	PPB	12/29/2022 15:18	Std
	5 S03	21369	-		5 SFAM01.1	PPB	12/29/2022 15:21	Std
	7.5 S04	31999	-		7.5 SFAM01.1	PPB	12/29/2022 15:23	Std
	10 S05	43418	-		10 SFAM01.1	PPB	12/29/2022 15:29	Std
ICV.82	ICV.82	16994	3.9639 -		SFAM01.1	PPB	12/29/2022 15:33	SMPL
ICB.82	ICB.82	196	0.0819 -		SFAM01.1	PPB	12/29/2022 15:36	SMPL
CCV.63	CCV.63	21248	4.947 -		SFAM01.1	PPB	12/29/2022 15:38	SMPL
CCB.63	CCB.63	127	0.066 -		SFAM01.1	PPB	12/29/2022 15:40	SMPL
PB149996BL	PBW996	107	0.0614 -		SFAM01.1	PPB	12/29/2022 15:42	SMPL
N6079-01	YBKA2	70676	16.3697 -		SFAM01.1	PPB	12/29/2022 15:45	SMPL
N6079-02	YBKA4	132173	30.5816 -		SFAM01.1	PPB	12/29/2022 15:47	SMPL
N6079-03	YBKA6	60485	14.0146 -		SFAM01.1	PPB	12/29/2022 15:49	SMPL
N6079-04	YBKA8	90828	21.0268 -		SFAM01.1	PPB	12/29/2022 15:51	SMPL
N6079-05	YBKB2	412	0.1319 -		SFAM01.1	PPB	12/29/2022 15:54	SMPL
N6079-06	YBKB4	5354	1.2739 -		SFAM01.1	PPB	12/29/2022 15:56	SMPL
N6079-07	YBKB6	4929	1.1757 -		SFAM01.1	PPB	12/29/2022 15:58	SMPL
N6079-08	YBKB8	6361	1.5067 -		SFAM01.1	PPB	12/29/2022 16:01	SMPL
N6079-09	YBKCO	3348	0.8104 -		SFAM01.1	PPB	12/29/2022 16:03	SMPL
N6079-10	YBKf4	32575	7.5647 -		SFAM01.1	PPB	12/29/2022 16:05	SMPL
N6079-11	YBKC4	14113	3.2981 -		SFAM01.1	PPB	12/29/2022 16:07	SMPL
N6079-12	YBKC6	558	0.1656 -		SFAM01.1	PPB	12/29/2022 16:10	SMPL
N6079-13	YBKE2	46867	10.8675 -		SFAM01.1	PPB	12/29/2022 16:12	SMPL
N6079-14	YBKE2D	48702	11.2916 -		SFAM01.1	PPB	12/29/2022 16:14	SMPL
N6079-15	YBKE2S	47128	10.9278 -		SFAM01.1	PPB	12/29/2022 16:16	SMPL
N6079-16	YBKf6	45355	10.5181 -		SFAM01.1	PPB	12/29/2022 16:19	SMPL
N6079-17	YBKf8	45010	10.4384 -		SFAM01.1	PPB	12/29/2022 16:21	SMPL
N6079-18	YBKGO	87872	20.3437 -		SFAM01.1	PPB	12/29/2022 16:23	SMPL
N6079-19	YBKG2	88287	20.4396 -		SFAM01.1	PPB	12/29/2022 16:25	SMPL
CCV064	CCV064	21104	4.9137 -		SFAM01.1	PPB	12/29/2022 16:28	SMPL
CCB064	CCB064	236	0.0912 -		SFAM01.1	PPB	12/29/2022 16:30	SMPL
N6079-20	YBKG4	38548	8.945 -		SFAM01.1	PPB	12/29/2022 16:32	SMPL
N6079-21	YBKJ1	22565	5.2514 -		SFAM01.1	PPB	12/29/2022 16:35	SMPL
PB149997BL	PBW997	-43	0.0267 -		SFAM01.1	PPB	12/29/2022 16:37	SMPL
N6080-01	YBKC8	3165	0.7681 -		SFAM01.1	PPB	12/29/2022 16:39	SMPL
N6080-02	YBKC8D	3180	0.7715 -		SFAM01.1	PPB	12/29/2022 16:41	SMPL
N6080-03	YBKC8S	7073	1.6712 -		SFAM01.1	PPB	12/29/2022 16:44	SMPL
N6080-04	YBKDO	1338	0.3459 -		SFAM01.1	PPB	12/29/2022 16:46	SMPL
N6080-05	YBKD2	1161	0.305 -		SFAM01.1	PPB	12/29/2022 16:48	SMPL
N6080-06	YBKD4	1179	0.3091 -		SFAM01.1	PPB	12/29/2022 16:50	SMPL
N6080-07	YBKD8	104453	24.1756 -		SFAM01.1	PPB	12/29/2022 16:53	SMPL
N6080-08	YBKE0	143652	33.2344 -		SFAM01.1	PPB	12/29/2022 16:55	SMPL
N6080-09	YBKE4	278419	64.3788 -		SFAM01.1	PPB	12/29/2022 16:57	SMPL
N6080-10	YBKf2	6899	1.631 -		SFAM01.1	PPB	12/29/2022 17:00	SMPL
N6080-11	YBKJ7	17296	4.0337 -		SFAM01.1	PPB	12/29/2022 17:02	SMPL

LB123489

INSTRUMENT ID : CV1

N6080-12	YBKFO	2721	0.6655 -	SFAM01.1	PPB	12/29/2022 17:04	SMPL
N6080-13	YBKJ3	745	0.2088 -	SFAM01.1	PPB	12/29/2022 17:06	SMPL
N6080-14	YBKJ5	715	0.2019 -	SFAM01.1	PPB	12/29/2022 17:09	SMPL
PB149998BL	PBW998	-118	0.0094 -	SFAM01.1	PPB	12/29/2022 17:11	SMPL
N6081-01	YBKA3	1335	0.3452 -	SFAM01.1	PPB	12/29/2022 17:13	SMPL
N6081-02	YBKA5	910	0.2469 -	SFAM01.1	PPB	12/29/2022 17:15	SMPL
CCV065	CCV065	21325	4.9648 -	SFAM01.1	PPB	12/29/2022 17:18	SMPL
CCB065	CCB065	-99	0.0138 -	SFAM01.1	PPB	12/29/2022 17:20	SMPL
N6081-03	YBKA7	836	0.2298 -	SFAM01.1	PPB	12/29/2022 17:22	SMPL
N6081-04	YBKA9	3144	0.7632 -	SFAM01.1	PPB	12/29/2022 17:25	SMPL
N6081-05	YBKB3	541	0.1617 -	SFAM01.1	PPB	12/29/2022 17:27	SMPL
N6081-06	YBKB5	595	0.1742 -	SFAM01.1	PPB	12/29/2022 17:29	SMPL
N6081-07	YBKB7	641	0.1848 -	SFAM01.1	PPB	12/29/2022 17:31	SMPL
N6081-08	YBKB9	629	0.182 -	SFAM01.1	PPB	12/29/2022 17:34	SMPL
N6081-09	YBKC1	618	0.1795 -	SFAM01.1	PPB	12/29/2022 17:36	SMPL
N6081-10	YBK F5	471	0.1455 -	SFAM01.1	PPB	12/29/2022 17:38	SMPL
N6081-11	YBKC5	829	0.2282 -	SFAM01.1	PPB	12/29/2022 17:40	SMPL
N6081-12	YBKC7	4917	1.173 -	SFAM01.1	PPB	12/29/2022 17:43	SMPL
N6081-13	YBKE3	619	0.1797 -	SFAM01.1	PPB	12/29/2022 17:45	SMPL
N6081-14	YBKE3D	614	0.1785 -	SFAM01.1	PPB	12/29/2022 17:47	SMPL
N6081-15	YBKE3S	4928	1.1755 -	SFAM01.1	PPB	12/29/2022 17:49	SMPL
N6081-16	YBK F7	1449	0.3715 -	SFAM01.1	PPB	12/29/2022 17:52	SMPL
N6081-17	YBK F9	1465	0.3752 -	SFAM01.1	PPB	12/29/2022 17:54	SMPL
N6081-18	YBKG1	1049	0.2791 -	SFAM01.1	PPB	12/29/2022 17:56	SMPL
N6081-19	YBKG3	1016	0.2714 -	SFAM01.1	PPB	12/29/2022 17:59	SMPL
N6081-20	YBKG5	2781	0.6793 -	SFAM01.1	PPB	12/29/2022 18:01	SMPL
N6081-21	YBKJ2	928	0.2511 -	SFAM01.1	PPB	12/29/2022 18:03	SMPL
PB149999BL	PBW999	8	0.0385 -	SFAM01.1	PPB	12/29/2022 18:05	SMPL
CCV066	CCV066	21419	4.9865 -	SFAM01.1	PPB	12/29/2022 18:08	SMPL
CCB066	CCB066	-32	0.0293 -	SFAM01.1	PPB	12/29/2022 18:10	SMPL
N6082-01	YBKC9	514	0.1554 -	SFAM01.1	PPB	12/29/2022 18:12	SMPL
N6082-02	YBKC9D	466	0.1443 -	SFAM01.1	PPB	12/29/2022 18:14	SMPL
N6082-03	YBKC9S	4558	1.09 -	SFAM01.1	PPB	12/29/2022 18:17	SMPL
N6082-04	YBKD1	186	0.0796 -	SFAM01.1	PPB	12/29/2022 18:19	SMPL
N6082-05	YBKD3	440	0.1383 -	SFAM01.1	PPB	12/29/2022 18:21	SMPL
N6082-06	YBKD5	216	0.0866 -	SFAM01.1	PPB	12/29/2022 18:24	SMPL
N6082-07	YBKD9	559	0.1658 -	SFAM01.1	PPB	12/29/2022 18:26	SMPL
N6082-08	YBKE1	796	0.2206 -	SFAM01.1	PPB	12/29/2022 18:28	SMPL
N6082-09	YBKE5	524	0.1577 -	SFAM01.1	PPB	12/29/2022 18:30	SMPL
N6082-10	YBK F3	302	0.1064 -	SFAM01.1	PPB	12/29/2022 18:33	SMPL
N6082-11	YBKJ8	330	0.1129 -	SFAM01.1	PPB	12/29/2022 18:35	SMPL
N6082-12	YBK F1	113	0.0628 -	SFAM01.1	PPB	12/29/2022 18:37	SMPL
N6082-13	YBKG6	76	0.0542 -	SFAM01.1	PPB	12/29/2022 18:39	SMPL
N6082-14	YBKJ4	120	0.0644 -	SFAM01.1	PPB	12/29/2022 18:42	SMPL
N6082-15	YBKJ6	224	0.0884 -	SFAM01.1	PPB	12/29/2022 18:44	SMPL
CCV067	CCV067	21202	4.9364 -	SFAM01.1	PPB	12/29/2022 18:46	SMPL
CCB067	CCB067	-32	0.0293 -	SFAM01.1	PPB	12/29/2022 18:49	SMPL

LB123489

INSTRUMENT ID : CV1

N6079-01DLX5		13851	3.2376 -	SFAM01.1	PPB	12/29/2022 18:51	SMPL
N6079-02DLX5		25992	6.0434 -	SFAM01.1	PPB	12/29/2022 18:53	SMPL
N6079-03DLX5		11307	2.6497 -	SFAM01.1	PPB	12/29/2022 18:55	SMPL
N6079-04DLX5		17070	3.9815 -	SFAM01.1	PPB	12/29/2022 18:58	SMPL
N6079-13DLX2		24811	5.7704 -	SFAM01.1	PPB	12/29/2022 19:00	SMPL
N6079-14DLX2		24672	5.7383 -	SFAM01.1	PPB	12/29/2022 19:02	SMPL
N6079-15DLX2		24278	5.6473 -	SFAM01.1	PPB	12/29/2022 19:04	SMPL
N6079-16DLX2		22461	5.2273 -	SFAM01.1	PPB	12/29/2022 19:07	SMPL
N6079-17DLX2		22581	5.2551 -	SFAM01.1	PPB	12/29/2022 19:09	SMPL
N6079-18DLX5		15754	3.6774 -	SFAM01.1	PPB	12/29/2022 19:11	SMPL
N6079-19DLX5		10483	2.4593 -	SFAM01.1	PPB	12/29/2022 19:14	SMPL
N6080-07DLX5		19955	4.6482 -	SFAM01.1	PPB	12/29/2022 19:16	SMPL
N6080-08DLX5		23748	5.5248 -	SFAM01.1	PPB	12/29/2022 19:18	SMPL
N6080-09DLX10		21118	4.917 -	SFAM01.1	PPB	12/29/2022 19:20	SMPL
CCV068	CCV068	21107	4.9144 -	SFAM01.1	PPB	12/29/2022 19:23	SMPL
CCB068	CCB068	-314	-0.0359 -	SFAM01.1	PPB	12/29/2022 19:25	SMPL