

LB123502


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42111



Max. Conc.: 10.000 PPB

A= 0.0000e+000

B= 2.3557e-004

C= 3.4172e-002

Rho= 0.9998660

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.030	0.030	-19	0.000	-19					
0.2	0.200	0.186	-0.014	645	0.0 %	645					
2.5	2.500	2.496	-0.004	10452	0.0 %	10452					
5.0	5.000	4.922	-0.078	20748	0.0 %	20748					
7.5	7.500	7.612	0.112	32169	0.0 %	32169					
10.0	10.000	9.954	-0.046	42111	0.0 %	42111					

1701-20-0

LB123502 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Dec 2022 12:05:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.0000	7470A	PPB	30 Dec 2022 12:18:17	S	Std
0.2 - 1	50.2	645	-	0.2000	7470A	PPB	30 Dec 2022 12:22:49	S	Std
2.5 - 1	52.5	10452	-	2.5000	7470A	PPB	30 Dec 2022 12:25:07	S	Std
5.0 - 1	55	20748	-	5.0000	7470A	PPB	30 Dec 2022 12:27:24	S	Std
7.5 - 1	57.5	32169	-	7.5000	7470A	PPB	30 Dec 2022 12:29:40	S	Std
10.0 - 1	510	42111	-	10.0000	7470A	PPB	30 Dec 2022 12:31:56	S	Std
ICV83 - 1	ICV83	16490	3.9187	-	7470A	PPB	30 Dec 2022 12:40:20	U	SMPL
ICB83 - 1	ICB83	42	0.0441	-	7470A	PPB	30 Dec 2022 12:42:35	U	SMPL
CCV69 - 1	CCV69	20280	4.8115	-	7470A	PPB	30 Dec 2022 12:44:50	U	SMPL
CCB69 - 1	CCB69	-5	0.0330	-	7470A	PPB	30 Dec 2022 12:47:05	U	SMPL
CRA - 1	CRA	523	0.1574	-	7470A	PPB	30 Dec 2022 12:49:22	U	SMPL
HighStd - 1	HighStd	42192	9.9732	-	7470A	PPB	30 Dec 2022 12:51:37	U	SMPL
ChkStd - 1	ChkStd	30145	7.1353	-	7470A	PPB	30 Dec 2022 12:56:31	U	SMPL
PB150005BL - 1	PBW	-6	0.0328	-	7470A	PPB	30 Dec 2022 12:58:46	U	SMPL
PB150005BS - 1	LCSW	18002	4.2748	-	7470A	PPB	30 Dec 2022 13:01:02	U	SMPL
N6228-02 - 1	TP-1	16	0.0379	-	7470A	PPB	30 Dec 2022 13:03:18	U	SMPL
N6228-05 - 1	TP-2	35	0.0424	-	7470A	PPB	30 Dec 2022 13:05:34	U	SMPL
N6228-08 - 1	TP-3	21	0.0391	-	7470A	PPB	30 Dec 2022 13:07:51	U	SMPL
N6228-11 - 1	TP-4	75	0.0518	-	7470A	PPB	30 Dec 2022 13:10:08	U	SMPL
N6231-02 - 1	TP-3S	2	0.0346	-	7470A	PPB	30 Dec 2022 13:12:23	U	SMPL
CCV70 - 1	CCV70	19156	4.5467	-	7470A	PPB	30 Dec 2022 13:14:38	U	SMPL
CCB70 - 1	CCB70	-24	0.0285	-	7470A	PPB	30 Dec 2022 13:16:53	U	SMPL
N6231-05 - 1	TP-2S	9	0.0363	-	7470A	PPB	30 Dec 2022 13:19:09	U	SMPL
N6231-05DUP - 1	TP-2SDUP	20	0.0389	-	7470A	PPB	30 Dec 2022 13:21:24	U	SMPL
N6231-05MS - 1	TP-2SMS	15535	3.6937	-	7470A	PPB	30 Dec 2022 13:23:39	U	SMPL
N6231-05MSD - 1	TP-2SMSD	17051	4.0508	-	7470A	PPB	30 Dec 2022 13:25:55	U	SMPL
N6231-05LX5 - 1		-32	0.0266	-	7470A	PPB	30 Dec 2022 13:28:11	U	SMPL
N6231-05A - 1		16595	3.9434	-	7470A	PPB	30 Dec 2022 13:30:27	U	SMPL
PB149967TB - 1	PB149967TB	4	0.0351	-	7470A	PPB	30 Dec 2022 13:32:43	U	SMPL
CCV71 - 1	CCV71	20541	4.8729	-	7470A	PPB	30 Dec 2022 13:35:00	U	SMPL
CCB71 - 1	CCB71	-54	0.0215	-	7470A	PPB	30 Dec 2022 13:37:17	U	SMPL