

L3123505


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

43535



A= 0.0000e+000

B= 2.2896e-004 *slight*

C= -1.4803e-003 *intercept*

Rho= 0.9999301

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.013	0.013	64	0.000	64					
0.05	0.050					185					
0.20	0.200	0.219	0.019	965	0.0 %	965	0				
2.50	2.500	2.454	-0.046	10725	0.0 %	10725					
5.00	5.000	4.966	-0.034	21697	0.0 %	21697					
7.50	7.500	7.581	0.081	33115	0.0 %	33115					
10.0	10.000	9.966	-0.034	43535	0.0 %	43535					

Handwritten notes:
 10
 114123
 mb

LB123505
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		64	-		0 SFAM01.1	PPB	12/30/2022 16:15	Std
0.2		965	-		0.2 SFAM01.1	PPB	12/30/2022 16:20	Std
2.5		10725	-		2.5 SFAM01.1	PPB	12/30/2022 16:22	Std
5		21697	-		5 SFAM01.1	PPB	12/30/2022 16:24	Std
7.5		33115	-		7.5 SFAM01.1	PPB	12/30/2022 16:27	Std
10		43535	-		10 SFAM01.1	PPB	12/30/2022 16:29	Std
ICV084	ICV084	17771	4.0674	-	SFAM01.1	PPB	12/30/2022 16:32	SMPL
ICB084	ICB084	83	0.0175	-	SFAM01.1	PPB	12/30/2022 16:34	SMPL
CCV072	CCV072	21698	4.9665	-	SFAM01.1	PPB	12/30/2022 16:36	SMPL
CCB072	CCB072	6	-0.0001	-	SFAM01.1	PPB	12/30/2022 16:39	SMPL
PB149957BL	PBS957	28	0.0049	-	SFAM01.1	PPB	12/30/2022 16:41	SMPL
N6083-01	YBKG7	1323559	HIGH	-	SFAM01.1	PPB	12/30/2022 16:43	
N6083-02	YBKG7D	1315539	HIGH	-	SFAM01.1	PPB	12/30/2022 16:45	
N6083-03	YBKG7S	1390951	HIGH	-	SFAM01.1	PPB	12/30/2022 16:47	
N6083-04	YBKG8	648471	HIGH	-	SFAM01.1	PPB	12/30/2022 16:49	
N6083-05	YBKG9	666954	HIGH	-	SFAM01.1	PPB	12/30/2022 16:51	
N6083-06	YBKHO	791520	HIGH	-	SFAM01.1	PPB	12/30/2022 16:52	
N6083-07	YBKH2	188841	43.2358	-	SFAM01.1	PPB	12/30/2022 16:54	SMPL
N6083-08	YBKH3	111795	25.5953	-	SFAM01.1	PPB	12/30/2022 16:57	SMPL
N6083-09	YBKH4	116007	26.5596	-	SFAM01.1	PPB	12/30/2022 16:59	SMPL
N6083-10	YBKH5	287208	65.7581	-	SFAM01.1	PPB	12/30/2022 17:01	SMPL
N6083-11	YBKH6	444750	HIGH	-	SFAM01.1	PPB	12/30/2022 17:03	
N6083-12	YBKH7	314121	HIGH	-	SFAM01.1	PPB	12/30/2022 17:05	
N6083-13	YBKH8	280477	64.2169	-	SFAM01.1	PPB	12/30/2022 17:07	SMPL
N6083-14	YBKH9	175337	40.1439	-	SFAM01.1	PPB	12/30/2022 17:10	SMPL
N6083-15	YBKJ0	331461	75.8903	-	SFAM01.1	PPB	12/30/2022 17:12	SMPL
PB149959BL	PBS959	-265	-0.0622	-	SFAM01.1	PPB	12/30/2022 17:17	SMPL
N6108-02	GBP57	301	0.0674	-	SFAM01.1	PPB	12/30/2022 17:21	SMPL
N6108-03	GBP58	2499	0.5707	-	SFAM01.1	PPB	12/30/2022 17:24	SMPL
CCV073	CCV073	20096	4.5997	-	SFAM01.1	PPB	12/30/2022 17:26	SMPL
CCB073	CCB073	-357	-0.0832	-	SFAM01.1	PPB	12/30/2022 17:31	SMPL
N6108-04	GBP59	116	0.0251	-	SFAM01.1	PPB	12/30/2022 17:34	SMPL
N6108-05	GBP59D	165	0.0363	-	SFAM01.1	PPB	12/30/2022 17:36	SMPL
N6108-06	GBP59S	12467	2.853	-	SFAM01.1	PPB	12/30/2022 17:38	SMPL
N6108-07	GBP60	418	0.0942	-	SFAM01.1	PPB	12/30/2022 17:40	SMPL
N6108-08	GBP61	2056	0.4693	-	SFAM01.1	PPB	12/30/2022 17:43	SMPL
N6108-09	GBP62	75	0.0157	-	SFAM01.1	PPB	12/30/2022 17:45	SMPL
N6108-10	GBP63	147	0.0322	-	SFAM01.1	PPB	12/30/2022 17:47	SMPL
N6108-12	GBP65	268	0.0599	-	SFAM01.1	PPB	12/30/2022 17:52	SMPL
N6108-13	GBP66	1840	0.4198	-	SFAM01.1	PPB	12/30/2022 17:54	SMPL
N6108-14	GBP67	1667	0.3802	-	SFAM01.1	PPB	12/30/2022 17:56	SMPL
N6108-15	GBP68	1285	0.2927	-	SFAM01.1	PPB	12/30/2022 17:58	SMPL
N6108-16	GBP69	1184	0.2696	-	SFAM01.1	PPB	12/30/2022 18:01	SMPL
N6108-17	GBP70	1691	0.3857	-	SFAM01.1	PPB	12/30/2022 18:03	SMPL
N6108-18	GBP71	981	0.2231	-	SFAM01.1	PPB	12/30/2022 18:05	SMPL
N6108-19	GBP72	760	0.1725	-	SFAM01.1	PPB	12/30/2022 18:07	SMPL

LB123505
INSTRUMENT ID : CV1

N6108-20	GBP73	1299	0.2959 -	SFAM01.1	PPB	12/30/2022 18:10 SMPL
N6108-21	GBP74	935	0.2126 -	SFAM01.1	PPB	12/30/2022 18:12 SMPL
N6108-22	GBP96	984	0.2238 -	SFAM01.1	PPB	12/30/2022 18:14 SMPL
PB149960BL	PBS960	33	0.0061 -	SFAM01.1	PPB	12/30/2022 18:17 SMPL
CCV074	CCV074	21748	4.978 -	SFAM01.1	PPB	12/30/2022 18:19 SMPL
CCB074	CCB074	-14	-0.0047 -	SFAM01.1	PPB	12/30/2022 18:21 SMPL
N6109-01	GBP75	988	0.2247 -	SFAM01.1	PPB	12/30/2022 18:23 SMPL
N6109-02	GBP76	255	0.0569 -	SFAM01.1	PPB	12/30/2022 18:26 SMPL
N6109-03	GBP77	450	0.1016 -	SFAM01.1	PPB	12/30/2022 18:28 SMPL
N6109-04	GBP78	1002	0.2279 -	SFAM01.1	PPB	12/30/2022 18:30 SMPL
N6109-05	GBP78D	1050	0.2389 -	SFAM01.1	PPB	12/30/2022 18:32 SMPL
N6109-06	GBP78S	12886	2.9489 -	SFAM01.1	PPB	12/30/2022 18:35 SMPL
N6109-07	GBP79	1810	0.4129 -	SFAM01.1	PPB	12/30/2022 18:37 SMPL
N6109-08	GBP80	582	0.1318 -	SFAM01.1	PPB	12/30/2022 18:39 SMPL
N6109-09	GBP81	1735	0.3958 -	SFAM01.1	PPB	12/30/2022 18:42 SMPL
N6109-10	GBP82	2247	0.513 -	SFAM01.1	PPB	12/30/2022 18:44 SMPL
N6109-11	GBP83	2097	0.4787 -	SFAM01.1	PPB	12/30/2022 18:46 SMPL
N6109-12	GBP84	1626	0.3708 -	SFAM01.1	PPB	12/30/2022 18:48 SMPL
N6109-15	GBP87	1871	0.4269 -	SFAM01.1	PPB	12/30/2022 18:55 SMPL
N6109-16	GBP88	1835	0.4187 -	SFAM01.1	PPB	12/30/2022 18:57 SMPL
N6109-17	GBP89	1300	0.2962 -	SFAM01.1	PPB	12/30/2022 19:00 SMPL
N6109-18	GBP90	877	0.1993 -	SFAM01.1	PPB	12/30/2022 19:02 SMPL
N6109-19	GBP91	1454	0.3314 -	SFAM01.1	PPB	12/30/2022 19:04 SMPL
N6109-20	GBP92	916	0.2082 -	SFAM01.1	PPB	12/30/2022 19:07 SMPL
CCV075	CCV075	20890	4.7815 -	SFAM01.1	PPB	12/30/2022 19:09 SMPL
CCB075	CCB075	-19	-0.0058 -	SFAM01.1	PPB	12/30/2022 19:11 SMPL
N6109-21	GBP93	824	0.1872 -	SFAM01.1	PPB	12/30/2022 19:13 SMPL
PB149961BL	PBS961	13	0.0015 -	SFAM01.1	PPB	12/30/2022 19:18 SMPL
N6113-01	GBP99	4593	1.0501 -	SFAM01.1	PPB	12/30/2022 19:20 SMPL
N6113-03	GBPA1	1981	0.4521 -	SFAM01.1	PPB	12/30/2022 19:25 SMPL
N6113-04	GBPA2	1989	0.4539 -	SFAM01.1	PPB	12/30/2022 19:27 SMPL
N6113-05	GBPA3	2597	0.5931 -	SFAM01.1	PPB	12/30/2022 19:29 SMPL
N6113-06	GBPA4	796	0.1808 -	SFAM01.1	PPB	12/30/2022 19:32 SMPL
N6113-07	GBPA5	1144	0.2605 -	SFAM01.1	PPB	12/30/2022 19:34 SMPL
N6113-08	GBPA6	6733	1.5401 -	SFAM01.1	PPB	12/30/2022 19:36 SMPL
N6113-09	GBPA7	2179	0.4974 -	SFAM01.1	PPB	12/30/2022 19:38 SMPL
N6113-10	GBPA7D	2131	0.4864 -	SFAM01.1	PPB	12/30/2022 19:41 SMPL
N6113-11	GBPA7S	14031	3.2111 -	SFAM01.1	PPB	12/30/2022 19:43 SMPL
N6113-12	GBPA8	4597	1.0511 -	SFAM01.1	PPB	12/30/2022 19:45 SMPL
N6113-13	GBPA9	469	0.1059 -	SFAM01.1	PPB	12/30/2022 19:47 SMPL
N6113-14	GBPB0	445	0.1004 -	SFAM01.1	PPB	12/30/2022 19:50 SMPL
N6113-15	GBPB1	1125	0.2561 -	SFAM01.1	PPB	12/30/2022 19:52 SMPL
N6113-16	GBPB2	2082	0.4752 -	SFAM01.1	PPB	12/30/2022 19:54 SMPL
CCV076	CCV076	21339	4.8843 -	SFAM01.1	PPB	12/30/2022 19:59 SMPL
CCB076	CCB076	-41	-0.0109 -	SFAM01.1	PPB	12/30/2022 20:01 SMPL
N6113-18	GBPB4	1697	0.3871 -	SFAM01.1	PPB	12/30/2022 20:03 SMPL
N6113-19	GBPB5	603	0.1366 -	SFAM01.1	PPB	12/30/2022 20:06 SMPL

LB123505

INSTRUMENT ID : CV1

N6113-20	GBPB6	758	0.1721 -	SFAM01.1	PPB	12/30/2022 20:08	SMPL
N6113-21	GBPB7	678	0.1538 -	SFAM01.1	PPB	12/30/2022 20:10	SMPL
PB149962BL	PBS962	10	0.0008 -	SFAM01.1	PPB	12/30/2022 20:17	SMPL
N6114-01	GBPB8	3304	0.755 -	SFAM01.1	PPB	12/30/2022 20:19	SMPL
N6114-02	GBPB9	1019	0.2318 -	SFAM01.1	PPB	12/30/2022 20:22	SMPL
N6114-03	GBPC0	707	0.1604 -	SFAM01.1	PPB	12/30/2022 20:24	SMPL
N6114-04	GBPC1	686	0.1556 -	SFAM01.1	PPB	12/30/2022 20:26	SMPL
N6114-05	GBPC1D	696	0.1579 -	SFAM01.1	PPB	12/30/2022 20:28	SMPL
N6114-06	GBPC1S	12397	2.837 -	SFAM01.1	PPB	12/30/2022 20:31	SMPL
N6114-07	GBPC2	777	0.1764 -	SFAM01.1	PPB	12/30/2022 20:33	SMPL
N6114-08	GBPC3	327	0.0734 -	SFAM01.1	PPB	12/30/2022 20:35	SMPL
N6114-09	GBPC4	977	0.2222 -	SFAM01.1	PPB	12/30/2022 20:37	SMPL
N6114-10	GBPC5	243	0.0542 -	SFAM01.1	PPB	12/30/2022 20:40	SMPL
N6114-11	GBPC6	448	0.1011 -	SFAM01.1	PPB	12/30/2022 20:42	SMPL
N6114-12	GBPC7	1823	0.4159 -	SFAM01.1	PPB	12/30/2022 20:44	SMPL
N6114-13	GBPC8	5008	1.1452 -	SFAM01.1	PPB	12/30/2022 20:47	SMPL
CCV077	CCV077	21385	4.8949 -	SFAM01.1	PPB	12/30/2022 20:49	SMPL
CCB077	CCB077	0	-0.0015 -	SFAM01.1	PPB	12/30/2022 20:51	SMPL
N6114-15	GBPDO	10101	2.3113 -	SFAM01.1	PPB	12/30/2022 20:56	SMPL
N6114-16	GBPD1	1893	0.4319 -	SFAM01.1	PPB	12/30/2022 20:58	SMPL
N6114-17	GBPD2	1222	0.2783 -	SFAM01.1	PPB	12/30/2022 21:00	SMPL
N6114-18	GBPD3	1137	0.2588 -	SFAM01.1	PPB	12/30/2022 21:03	SMPL
N6114-19	GBPD4	1054	0.2398 -	SFAM01.1	PPB	12/30/2022 21:05	SMPL
N6114-20	GBPD5	2253	0.5144 -	SFAM01.1	PPB	12/30/2022 21:07	SMPL
N6114-21	GBPD6	721	0.1636 -	SFAM01.1	PPB	12/30/2022 21:09	SMPL
PB149963BL	PBS963	78	0.0164 -	SFAM01.1	PPB	12/30/2022 21:14	SMPL
N6115-01	GBPD7	828	0.1881 -	SFAM01.1	PPB	12/30/2022 21:16	SMPL
N6115-02	GBPD8	1627	0.371 -	SFAM01.1	PPB	12/30/2022 21:19	SMPL
N6115-03	GBPD9	1546	0.3525 -	SFAM01.1	PPB	12/30/2022 21:21	SMPL
N6115-04	GBPD9D	1322	0.3012 -	SFAM01.1	PPB	12/30/2022 21:23	SMPL
N6115-05	GBPD9S	13377	3.0613 -	SFAM01.1	PPB	12/30/2022 21:25	SMPL
N6115-06	GBPE0	918	0.2087 -	SFAM01.1	PPB	12/30/2022 21:28	SMPL
N6115-07	GBPE1	995	0.2263 -	SFAM01.1	PPB	12/30/2022 21:30	SMPL
N6115-08	GBPE2	289	0.0647 -	SFAM01.1	PPB	12/30/2022 21:32	SMPL
N6115-09	GBPE3	158	0.0347 -	SFAM01.1	PPB	12/30/2022 21:34	SMPL
N6115-10	GBPE4	195	0.0432 -	SFAM01.1	PPB	12/30/2022 21:37	SMPL
CCV078	CCV078	21573	4.9379 -	SFAM01.1	PPB	12/30/2022 21:39	SMPL
CCB078	CCB078	-10	-0.0038 -	SFAM01.1	PPB	12/30/2022 21:41	SMPL
N6115-11	GBPE5	2301	0.5254 -	SFAM01.1	PPB	12/30/2022 21:44	SMPL
N6115-12	GBPE6	233	0.0519 -	SFAM01.1	PPB	12/30/2022 21:46	SMPL
N6115-13	GBPE7	318	0.0713 -	SFAM01.1	PPB	12/30/2022 21:48	SMPL
N6115-14	GBPE8	1943	0.4434 -	SFAM01.1	PPB	12/30/2022 21:50	SMPL
N6115-15	GBPE9	1177	0.268 -	SFAM01.1	PPB	12/30/2022 21:53	SMPL
N6115-16	GBPF0	1042	0.2371 -	SFAM01.1	PPB	12/30/2022 21:55	SMPL
N6115-17	GBPF1	112	0.0242 -	SFAM01.1	PPB	12/30/2022 21:57	SMPL
N6115-18	GBPF2	145	0.0317 -	SFAM01.1	PPB	12/30/2022 22:00	SMPL
N6115-19	GBPF3	506	0.1144 -	SFAM01.1	PPB	12/30/2022 22:02	SMPL

LB123505

INSTRUMENT ID : CV1

N6115-20	GBPF4	145	0.0317 -	SFAM01.1	PPB	12/30/2022 22:04	SMPL
N6115-21	GBPF5	136	0.0297 -	SFAM01.1	PPB	12/30/2022 22:06	SMPL
N6115-22	GBPH7	6037	1.3808 -	SFAM01.1	PPB	12/30/2022 22:09	SMPL
PB149964BL	PBS964	16	0.0022 -	SFAM01.1	PPB	12/30/2022 22:11	SMPL
N6116-01	GBPF6	173	0.0381 -	SFAM01.1	PPB	12/30/2022 22:13	SMPL
N6116-02	GBPF7	601	0.1361 -	SFAM01.1	PPB	12/30/2022 22:16	SMPL
N6116-04	GBPF9	1653	0.377 -	SFAM01.1	PPB	12/30/2022 22:20	SMPL
N6116-05	GBPF9D	1644	0.3749 -	SFAM01.1	PPB	12/30/2022 22:26	SMPL
N6116-07	GBPG0	866	0.1968 -	SFAM01.1	PPB	12/30/2022 22:30	SMPL
CCV079	CCV079	22184	5.0778 -	SFAM01.1	PPB	12/30/2022 22:36	SMPL
CCB079	CCB079	0	-0.0015 -	SFAM01.1	PPB	12/30/2022 22:38	SMPL
N6116-08	GBPG1	1990	0.4542 -	SFAM01.1	PPB	12/30/2022 22:40	SMPL
N6116-09	GBPG2	1166	0.2655 -	SFAM01.1	PPB	12/30/2022 22:42	SMPL
N6116-10	GBPG3	409	0.0922 -	SFAM01.1	PPB	12/30/2022 22:45	SMPL
N6116-11	GBPG4	103	0.0221 -	SFAM01.1	PPB	12/30/2022 22:47	SMPL
N6116-12	GBPG5	257	0.0574 -	SFAM01.1	PPB	12/30/2022 22:49	SMPL
N6116-13	GBPG6	88	0.0187 -	SFAM01.1	PPB	12/30/2022 22:52	SMPL
N6116-14	GBPG7	104	0.0223 -	SFAM01.1	PPB	12/30/2022 22:54	SMPL
N6116-15	GBPG8	333	0.0748 -	SFAM01.1	PPB	12/30/2022 22:56	SMPL
N6116-16	GBPG9	1771	0.404 -	SFAM01.1	PPB	12/30/2022 22:59	SMPL
N6116-17	GBPH0	3810	0.8709 -	SFAM01.1	PPB	12/30/2022 23:01	SMPL
N6116-18	GBPH1	1080	0.2458 -	SFAM01.1	PPB	12/30/2022 23:03	SMPL
N6116-19	GBPH2	294	0.0658 -	SFAM01.1	PPB	12/30/2022 23:05	SMPL
N6116-20	GBPH3	195	0.0432 -	SFAM01.1	PPB	12/30/2022 23:08	SMPL
N6116-21	GBPH4	818	0.1858 -	SFAM01.1	PPB	12/30/2022 23:10	SMPL
N6116-22	GBPH8	663	0.1503 -	SFAM01.1	PPB	12/30/2022 23:12	SMPL
N6116-06	GBPF9S	11352	2.5977 -	SFAM01.1	PPB	12/30/2022 23:15	SMPL
CCV080	CCV080	23653	5.4141 -	SFAM01.1	PPB	12/30/2022 23:17	SMPL
CCB080	CCB080	-28	-0.0079 -	SFAM01.1	PPB	12/30/2022 23:19	SMPL