

LB123508


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4938E



A= 0.0000e+000

B= 2.0206e-004 slope

C= -1.6432e-002 = intercept

Rho= 0.9999258

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.006	0.006	111	> 0.000	111					
0.05	0.050					185	0				
0.20	0.200	0.205	0.005	1096	0.0 %	1096					
2.50	2.500	2.431	-0.069	12111	0.0 %	12111					
5.00	5.000	5.074	0.074	25191	0.0 %	25191					
7.50	7.500	7.522	0.022	37306	0.0 %	37306					
10.0	10.000	9.963	-0.037	49388	0.0 %	49388					

LB123506

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	111	-		0 SFAM01.1	PPB	12/30/2022 23:43	Std
	0.2 S01	1096	-		0.2 SFAM01.1	PPB	12/30/2022 23:46	Std
	2.5 S02	12111	-		2.5 SFAM01.1	PPB	12/30/2022 23:48	Std
	5 S03	25191	-		5 SFAM01.1	PPB	12/30/2022 23:50	Std
	7.5 S04	37306	-		7.5 SFAM01.1	PPB	12/30/2022 23:53	Std
	10 S05	49388	-		10 SFAM01.1	PPB	12/30/2022 23:55	Std
ICV085	ICV085	20090	4.043	-	SFAM01.1	PPB	12/30/2022 23:58	SMPL
ICB085	ICB085	63	-0.0037	-	SFAM01.1	PPB	12/31/2022 0:00	SMPL
CCV081	CCV81	25250	5.0856	-	SFAM01.1	PPB	12/31/2022 0:02	SMPL
CCB081	CCB81	15	-0.0134	-	SFAM01.1	PPB	12/31/2022 0:04	SMPL
N6083-04DLX10	YBKG8	48583	9.8003	-	SFAM01.1	PPB	12/31/2022 0:18	SMPL
N6083-05DLX10	YBKG9	48216	9.7261	-	SFAM01.1	PPB	12/31/2022 0:21	SMPL
N6083-06DLX10	YBKHO	108984	22.0049	-	SFAM01.1	PPB	12/31/2022 0:23	SMPL
N6083-07DLX10	YBKH2	10922	2.1905	-	SFAM01.1	PPB	12/31/2022 0:25	SMPL
N6083-08DLX5	YBKH3	10026	2.0094	-	SFAM01.1	PPB	12/31/2022 0:28	SMPL
N6083-09DLX5	YBKH4	17737	3.5675	-	SFAM01.1	PPB	12/31/2022 0:30	SMPL
N6083-10DLX10	YBKH5	15472	3.1098	-	SFAM01.1	PPB	12/31/2022 0:32	SMPL
N6083-11DLX10	YBKH6	34007	6.855	-	SFAM01.1	PPB	12/31/2022 0:34	SMPL
N6083-12DLX10	YBKH7	17549	3.5295	-	SFAM01.1	PPB	12/31/2022 0:37	SMPL
N6083-13DLX10	YBKH8	14104	2.8334	-	SFAM01.1	PPB	12/31/2022 0:44	SMPL
N6083-14DLX10	YBKH9	8864	1.7746	-	SFAM01.1	PPB	12/31/2022 0:46	SMPL
N6083-15DLX10	YBKJO	12565	2.5225	-	SFAM01.1	PPB	12/31/2022 0:48	SMPL
N6083-01DLX10		331823	HIGH	-	SFAM01.1	PPB	12/31/2022 0:51	
N6083-02DLX10		335418	HIGH	-	SFAM01.1	PPB	12/31/2022 0:52	
N6083-03DLX10		343439	HIGH	-	SFAM01.1	PPB	12/31/2022 0:54	
CCV082	CCV082	25679	5.1723	-	SFAM01.1	PPB	12/31/2022 0:59	SMPL
CCB082	CCB082	27	-0.011	-	SFAM01.1	PPB	12/31/2022 1:01	SMPL
N6083-01DL2X100		29201	5.8839	-	SFAM01.1	PPB	12/31/2022 1:03	SMPL
N6083-02DL2X100		32987	6.6489	-	SFAM01.1	PPB	12/31/2022 1:08	SMPL
N6083-03DL2X100		37797	7.6209	-	SFAM01.1	PPB	12/31/2022 1:11	SMPL
N6083-04DL2X25		21934	4.4156	-	SFAM01.1	PPB	12/31/2022 1:13	SMPL
N6083-05DL2X25		43768	8.8274	-	SFAM01.1	PPB	12/31/2022 1:15	SMPL
N6083-06DL2X50		20527	4.1313	-	SFAM01.1	PPB	12/31/2022 1:18	SMPL
N6083-07DL2X5		42161	8.5026	-	SFAM01.1	PPB	12/31/2022 1:26	SMPL
N6083-08DL2X5		37731	7.6075	-	SFAM01.1	PPB	12/31/2022 1:28	SMPL
N6083-14DLX5		38773	7.8181	-	SFAM01.1	PPB	12/31/2022 1:34	SMPL
N6083-05DL2X50		4120	0.8161	-	SFAM01.1	PPB	12/31/2022 1:36	SMPL
CCV083	CCV083	25333	5.1024	-	SFAM01.1	PPB	12/31/2022 1:38	SMPL
CCB083	CCB083	-100	-0.0366	-	SFAM01.1	PPB	12/31/2022 1:41	SMPL