

LB123508


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

46836



A= 0.0000e+000

B= 2.1089e-004 *slope*C= 1.0560e-002 *intercept*

Rho= 0.9996718

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.00	0.000	0.037	0.037	127	0.000	127					=
0.05	0.050					185	0				-20
0.20	0.200	0.160	-0.040	708	0.0 %	708					-1
2.50	2.500	2.485	-0.015	11735	0.0 %	11735					-1
5.00	5.000	4.944	-0.056	23392	0.0 %	23392					2
7.50	7.500	7.686	0.186	36396	0.0 %	36396					-1
10.0	10.000	9.888	-0.112	46836	0.0 %	46836					

LB123508

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	127	-		0 SFAM01.1	PPB	12/31/2022 2:04	Std
	0.2 S01	708	-		0.2 SFAM01.1	PPB	12/31/2022 2:06	Std
	2.5 S02	11735	-		2.5 SFAM01.1	PPB	12/31/2022 2:08	Std
	5 S03	23392	-		5 SFAM01.1	PPB	12/31/2022 2:11	Std
	7.5 S04	36396	-		7.5 SFAM01.1	PPB	12/31/2022 2:13	Std
	10 S05	46836	-		10 SFAM01.1	PPB	12/31/2022 2:18	Std
ICV085	ICV085	19654	4.1554 -		SFAM01.1	PPB	12/31/2022 2:20	SMPL
ICB085	ICB085	149	0.042 -		SFAM01.1	PPB	12/31/2022 2:23	SMPL
CCV084	CCV084	23999	5.0717 -		SFAM01.1	PPB	12/31/2022 2:25	SMPL
CCB084	CCB084	93	0.0302 -		SFAM01.1	PPB	12/31/2022 2:27	SMPL
N6079-19DLX5	YBKG2	19338	4.0887 -		SFAM01.1	PPB	12/31/2022 2:29	SMPL
N6081-12	YBKC7	404	0.0958 -		SFAM01.1	PPB	12/31/2022 2:32	SMPL
CCV085	CCV085	24166	5.1069 -		SFAM01.1	PPB	12/31/2022 2:34	SMPL
CCB085	CCB085	42	0.0194 -		SFAM01.1	PPB	12/31/2022 2:36	SMPL