



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123516

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	01/03/2023	14:15
2	O1006-01		1	12	100.00	>212.0	>212.0	01/03/2023	14:40
3	O1006-01DUP		1	14	100.00	>212.0	>212.0	01/03/2023	15:10
4	O1006-02		1	13	100.00	>212.0	>212.0	01/03/2023	15:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp12-29

WorkList ID : 166211

Department : Wet-Chemistry

Date : 12-29-2022 08:42:09

66123516

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/10/2023	Water	O1006-01	Flash Point	Cool 4 deg C	PSEG03	N23	33888	01/03/2023	1010B
01/10/2023	Water	O1006-02	Flash Point	Cool 4 deg C	PSEG03	N23	33891-TOTE	01/03/2023	1010B

Date/Time 01/03/2023 14:00
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RS (BCE-WWC)

Date/Time 01/03/2023 16:00
 Raw Sample Received by: RS (BCE-WWC)
 Raw Sample Relinquished by: RM CWC