



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123518

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	01/03/2023	14:10
2	CAL2	1	Water	NA	NA	20.8	7.00	01/03/2023	14:11
3	CAL3	1	Water	NA	NA	20.8	10.03	01/03/2023	14:15
4	ICV	1	Water	NA	NA	20.7	6.99	01/03/2023	14:20
5	CCV1	1	Water	NA	NA	20.8	2.01	01/03/2023	14:22
6	N6240-01	1	Water	NA	NA	20.3	11.87	01/03/2023	14:25
7	N6240-02	1	Water	NA	NA	20.7	11.88	01/03/2023	14:28
8	N6240-03	1	Water	NA	NA	20.1	10.99	01/03/2023	14:30
9	N6240-04	1	Water	NA	NA	20.1	9.01	01/03/2023	14:33
10	N6240-05	1	Water	NA	NA	20.7	10.08	01/03/2023	14:36
11	O1006-01	1	Water	NA	NA	20.1	5.64	01/03/2023	14:40
12	O1006-02	1	Water	NA	NA	20.2	5.87	01/03/2023	14:45
13	O1006-02DUP	1	Water	NA	NA	20.3	5.88	01/03/2023	14:46
14	CCV2	1	Water	NA	NA	20.8	12.03	01/03/2023	14:50

W223518

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o1006 WorkList ID : 166273 Department : Wet-Chemistry Date : 01-03-2023 13:44:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
12/28/2022	Water	N6240-01	pH	Cool 4 deg C	ATCE02	N22	T4	12/28/2022	9040C
12/28/2022	Water	N6240-02	pH	Cool 4 deg C	ATCE02	N22	T5	12/28/2022	9040C
12/28/2022	Water	N6240-03	pH	Cool 4 deg C	ATCE02	N22	T6	12/28/2022	9040C
12/28/2022	Water	N6240-04	pH	Cool 4 deg C	ATCE02	N22	C1	12/28/2022	9040C
12/28/2022	Water	N6240-05	pH	Cool 4 deg C	ATCE02	N22	C2	12/28/2022	9040C
01/03/2023	Water	O1006-01	pH	Cool 4 deg C	PSEG03	N23	33888	01/03/2023	9040C
01/03/2023	Water	O1006-02	pH	Cool 4 deg C	PSEG03	N23	33891-TOTE	01/03/2023	9040C

Date/Time 01/03/23 13:50
 Raw Sample Received by: DP(WC)
 Raw Sample Relinquished by: QP SN

Date/Time 01/03/23 15:00
 Raw Sample Received by: QP SN
 Raw Sample Relinquished by: DP(WC)