


LB123519

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.8459e-004

C= 7.5825e-002

Rho= 0.9998949

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.086	0.086	57	0.000	57					0.00
0.2	0.200	0.171	-0.029	517	0.0 %	517					-15
2.5	2.500	2.450	-0.050	12861	0.0 %	12861					-2
5.0	5.000	4.938	-0.062	26338	0.0 %	26338					-1
7.5	7.500	7.550	0.050	40488	0.0 %	40488					1
10.0	10.000	10.005	0.005	53793	0.0 %	53793					0

LB123519 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Jan 2023 14:37:48

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	57	-	0.0000	7471B	PPB	03 Jan 2023 15:24:24	S	Std
0.2 - 1	50.2	517	-	0.2000	7471B	PPB	03 Jan 2023 15:26:42	S	Std
2.5 - 1	52.5	12861	-	2.5000	7471B	PPB	03 Jan 2023 15:28:59	S	Std
5.0 - 1	55	26338	-	5.0000	7471B	PPB	03 Jan 2023 15:31:16	S	Std
7.5 - 1	57.5	40488	-	7.5000	7471B	PPB	03 Jan 2023 15:33:32	S	Std
10.0 - 1	510	53793	-	10.0000	7471B	PPB	03 Jan 2023 15:35:48	S	Std
ICV87 - 1	ICV87	21936	4.1250	-	7471B	PPB	03 Jan 2023 15:40:59	U	SMPL
ICB87 - 1	ICB87	24	0.0803	-	7471B	PPB	03 Jan 2023 15:43:14	U	SMPL
CCV87 - 1	CCV87	26221	4.9160	-	7471B	PPB	03 Jan 2023 15:45:29	U	SMPL
CCB87 - 1	CCB87	-27	0.0708	-	7471B	PPB	03 Jan 2023 15:47:44	U	SMPL
CRA - 1	CRA	497	0.1676	-	7471B	PPB	03 Jan 2023 15:49:59	U	SMPL
HighStd - 1	HighStd	53588	9.9676	-	7471B	PPB	03 Jan 2023 15:52:14	U	SMPL
ChkStd - 1	ChkStd	37260	6.9537	-	7471B	PPB	03 Jan 2023 15:54:30	U	SMPL
PB150016BL - 1	PBS	-15	0.0731	-	7471B	PPB	03 Jan 2023 15:56:47	U	SMPL
PB150016BS - 1	LCSS	21003	3.9528	-	7471B	PPB	03 Jan 2023 15:59:03	U	SMPL
N6224-01 - 1	OK-3-122822	6978	1.3639	-	7471B	PPB	03 Jan 2023 16:01:19	U	SMPL
N6225-01 - 1	EO-03-122822	998	0.2600	-	7471B	PPB	03 Jan 2023 16:03:36	U	SMPL
N6228-01 - 1	TP-1	31556	5.9008	-	7471B	PPB	03 Jan 2023 16:05:53	U	SMPL
N6228-04 - 1	TP-2	64287	11.9426	-	7471B	PPB	03 Jan 2023 16:08:09	U	SMPL
N6228-07 - 1	TP-3	35908	6.7041	-	7471B	PPB	03 Jan 2023 16:10:25	U	SMPL
CCV88 - 1	CCV88	26802	5.0232	-	7471B	PPB	03 Jan 2023 16:12:41	U	SMPL
CCB88 - 1	CCB88	14	0.0784	-	7471B	PPB	03 Jan 2023 16:14:57	U	SMPL
N6228-10 - 1	TP-4	42415	7.9052	-	7471B	PPB	03 Jan 2023 16:17:13	U	SMPL
N6229-01 - 1	NB-8-122822	4444	0.8961	-	7471B	PPB	03 Jan 2023 16:19:28	U	SMPL
N6230-01 - 1	C-1	1661	0.3824	-	7471B	PPB	03 Jan 2023 16:21:43	U	SMPL
N6231-01 - 1	TP-3S	1127	0.2839	-	7471B	PPB	03 Jan 2023 16:23:59	U	SMPL
N6231-04 - 1	TP-2S	1321	0.3197	-	7471B	PPB	03 Jan 2023 16:26:15	U	SMPL
N6233-01 - 1	RO-124019	1104	0.2796	-	7471B	PPB	03 Jan 2023 16:28:31	U	SMPL
N6234-01 - 1	72-11944	135	0.1007	-	7471B	PPB	03 Jan 2023 16:30:48	U	SMPL
N6235-01 - 1	OR-2-122922	934	0.2482	-	7471B	PPB	03 Jan 2023 16:33:04	U	SMPL
N6237-01 - 1	TR-05-122922	4631	0.9307	-	7471B	PPB	03 Jan 2023 16:35:21	U	SMPL
N6239-01 - 1	TP-1S	495	0.1672	-	7471B	PPB	03 Jan 2023 16:37:37	U	SMPL
CCV89 - 1	CCV89	26821	5.0267	-	7471B	PPB	03 Jan 2023 16:39:53	U	SMPL
CCB89 - 1	CCB89	-58	0.0651	-	7471B	PPB	03 Jan 2023 16:42:08	U	SMPL
N6239-01DUP - 1	TP-1SDUP	485	0.1654	-	7471B	PPB	03 Jan 2023 16:44:24	U	SMPL
N6239-01MS - 1	TP-1SMS	21461	4.0373	-	7471B	PPB	03 Jan 2023 16:46:40	U	SMPL
N6239-01MSD - 1	TP-1MSD	21268	4.0017	-	7471B	PPB	03 Jan 2023 16:48:55	U	SMPL
N6239-04 - 1	TP-C1	881	0.2384	-	7471B	PPB	03 Jan 2023 16:51:11	U	SMPL
N6241-01 - 1	SU-2-122922	1341	0.3234	-	7471B	PPB	03 Jan 2023 16:53:27	U	SMPL
N6242-04 - 1	MM-4	239	0.1199	-	7471B	PPB	03 Jan 2023 16:55:43	U	SMPL
N6248-01 - 1	TP-15	725849	HIGH	-	7471B	PPB	03 Jan 2023 16:57:59	X	
N6248-12 - 1	TP-20	78260	14.5219	-	7471B	PPB	03 Jan 2023 16:59:49	U	SMPL
N6239-01LX5 - 1		27	0.0808	-	7471B	PPB	03 Jan 2023 17:02:09	U	SMPL
N6239-01A - 1		22069	4.1495	-	7471B	PPB	03 Jan 2023 17:04:26	U	SMPL
CCV90 - 1	CCV90	27059	5.0707	-	7471B	PPB	03 Jan 2023 17:06:42	U	SMPL
CCB90 - 1	CCB90	23	0.0801	-	7471B	PPB	03 Jan 2023 17:08:57	U	SMPL
N6228-04DLX2 - 1		32623	6.0977	-	7471B	PPB	03 Jan 2023 17:11:13	U	SMPL
N6248-01DLX10 - 1		121837	22.5657	-	7471B	PPB	03 Jan 2023 17:13:28	U	SMPL
N6248-12DLX5 - 1		15614	2.9580	-	7471B	PPB	03 Jan 2023 17:15:44	U	SMPL
N6248-01DL2X50 - 1		17882	3.3767	-	7471B	PPB	03 Jan 2023 17:18:02	U	SMPL
CCV91 - 1	CCV91	25823	4.8425	-	7471B	PPB	03 Jan 2023 17:20:17	U	SMPL
CCB91 - 1	CCB91	-51	0.0664	-	7471B	PPB	03 Jan 2023 17:22:35	U	SMPL