

LB123527


7470A

INSTRUMENT ID: CVL

Linear

μ Abs.:

47304



A= 0.0000e+000

B= 2.1331e-004

C= -1.2950e-002

Rho= 0.9997604

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Y/D
0.0	0.000	-0.011	-0.011	11	0.000	11					
0.2	0.200	0.171	-0.029	862	0.0 %	862					15
2.5	2.500	2.624	0.124	12364	0.0 %	12364					5
5.0	5.000	4.961	-0.039	23318	0.0 %	23318					1
7.5	7.500	7.377	-0.123	34643	0.0 %	34643					2
10.0	10.000	10.077	0.077	47304	0.0 %	47304					1

LB123527 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Jan 2023 13:05:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	11	-	0.0000	7470A	PPB	04 Jan 2023 13:08:59	S	Std
0.2 - 1	50.2	862	-	0.2000	7470A	PPB	04 Jan 2023 13:15:55	S	Std
2.5 - 1	52.5	12364	-	2.5000	7470A	PPB	04 Jan 2023 13:18:12	S	Std
5.0 - 1	55	23318	-	5.0000	7470A	PPB	04 Jan 2023 13:20:29	S	Std
7.5 - 1	57.5	34643	-	7.5000	7470A	PPB	04 Jan 2023 13:22:45	S	Std
10.0 - 1	51V	47304	-	10.0000	7470A	PPB	04 Jan 2023 13:27:33	S	Std
ICV89 - 1	ICV89	18064	3.8403	-	7470A	PPB	04 Jan 2023 13:30:18	U	SMPL
ICB89 - 1	ICB89	78	0.0037	-	7470A	PPB	04 Jan 2023 13:32:33	U	SMPL
CCV95 - 1	CCV95	22937	4.8797	-	7470A	PPB	04 Jan 2023 13:34:48	U	SMPL
CCB95 - 1	CCB95	-16	-0.0164	-	7470A	PPB	04 Jan 2023 13:37:03	U	SMPL
CRA - 1	CRA	822	0.1624	-	7470A	PPB	04 Jan 2023 13:39:18	U	SMPL
HighStd - 1	HighStd	44881	9.5606	-	7470A	PPB	04 Jan 2023 13:41:33	U	SMPL
ChkStd - 1	ChkStd	31326	6.6692	-	7470A	PPB	04 Jan 2023 13:43:49	U	SMPL
PB150042BL - 1	PBW	25	-0.0076	-	7470A	PPB	04 Jan 2023 13:48:27	U	SMPL
PB150042BS - 1	LCSW	18478	3.9286	-	7470A	PPB	04 Jan 2023 13:50:43	U	SMPL
N6233-02 - 1	RO-124019	419	0.0764	-	7470A	PPB	04 Jan 2023 13:53:00	U	SMPL
N6234-02 - 1	72-11944	393	0.0709	-	7470A	PPB	04 Jan 2023 13:55:16	U	SMPL
N6239-02 - 1	TP-1S	238	0.0378	-	7470A	PPB	04 Jan 2023 13:57:33	U	SMPL
N6239-05 - 1	TP-C1	341	0.0598	-	7470A	PPB	04 Jan 2023 13:59:50	U	SMPL
N6248-02 - 1	TP-15	333	0.0581	-	7470A	PPB	04 Jan 2023 14:02:05	U	SMPL
CCV96 - 1	CCV96	22427	4.7710	-	7470A	PPB	04 Jan 2023 14:04:21	U	SMPL
CCB96 - 1	CCB96	-67	-0.0272	-	7470A	PPB	04 Jan 2023 14:06:37	U	SMPL
N6248-13 - 1	TP-20	268	0.0442	-	7470A	PPB	04 Jan 2023 14:08:52	U	SMPL
O1001-02 - 1	TP-C2	174	0.0242	-	7470A	PPB	04 Jan 2023 14:11:08	U	SMPL
O1001-05 - 1	TP-C3	201	0.0299	-	7470A	PPB	04 Jan 2023 14:13:23	U	SMPL
O1009-02 - 1	TP-4	286	0.0481	-	7470A	PPB	04 Jan 2023 14:15:39	U	SMPL
O1009-02DUP - 1	TP-4DUP	200	0.0297	-	7470A	PPB	04 Jan 2023 15:02:15	U	SMPL
O1009-02MS - 1	TP-4MS	19241	4.0914	-	7470A	PPB	04 Jan 2023 15:04:31	U	SMPL
O1009-02MSD - 1	TP-4MSD	18606	3.9559	-	7470A	PPB	04 Jan 2023 15:06:48	U	SMPL
O1009-02LX5 - 1		0	-0.0130	-	7470A	PPB	04 Jan 2023 15:09:04	U	SMPL
O1009-02A - 1		17887	3.8025	-	7470A	PPB	04 Jan 2023 15:11:22	U	SMPL
PB150012TB - 1	PB150012TB	-30	-0.0193	-	7470A	PPB	04 Jan 2023 15:13:37	U	SMPL
CCV97 - 1	CCV97	22717	4.8328	-	7470A	PPB	04 Jan 2023 15:15:53	U	SMPL
CCB97 - 1	CCB97	-44	-0.0223	-	7470A	PPB	04 Jan 2023 15:18:08	U	SMPL