



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/5/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 15:40
In Date: 01/04/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:33
Out Date: 01/05/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123531

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1017-01	MDBXZ1	1	1.13	8.61	9.74	9.63	98.7	
O1017-02	MDBXZ2	2	1.13	8.66	9.79	9.65	98.4	
O1017-03	MDBXZ3	3	1.17	5.87	7.04	6.95	98.5	
O1017-04	MDBXZ4	4	1.12	8.66	9.78	9.64	98.4	
O1017-05	MDBXZ5	5	1.18	8.46	9.64	9.11	93.7	
O1017-06	MDBXZ7	6	1.19	8.47	9.66	9.51	98.2	
O1017-07	MDBXZ8	7	1.18	8.79	9.97	9.58	95.6	
O1017-08	MDBXZ9	8	1.19	8.53	9.72	9.56	98.1	
O1017-09	MDBYA0	9	1.18	8.49	9.67	9.52	98.2	
O1017-10	MDBYA1	10	1.18	8.73	9.91	9.73	97.9	
O1017-11	MDBYA2	11	1.17	8.50	9.67	9.56	98.7	
O1017-12	MDBYA3	12	1.19	8.65	9.84	9.7	98.4	
O1017-13	MDBYA6	13	1.18	8.48	9.66	9.6	99.3	
O1017-14	MDBYA7	14	1.13	8.40	9.53	9.45	99.0	
O1017-15	MDBYA8	15	1.19	8.66	9.85	9.75	98.8	
O1017-16	MDBYB2	16	1.18	8.72	9.9	9.72	97.9	
O1017-17	MDBYB3	17	1.12	8.82	9.94	9.75	97.8	
O1017-18	MDBYB4	18	1.16	8.58	9.74	9.54	97.7	
O1017-19	MDBYB5	19	1.18	8.54	9.72	9.58	98.4	
O1017-20	MDBYB8	20	1.13	8.78	9.91	9.8	98.7	
O1017-21	MDBYB8D	21	1.13	8.78	9.91	9.8	98.7	
O1017-22	MDBYB8S	22	1.13	8.78	9.91	9.8	98.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$