

LB123532


SFAM01.1

INSTRUMENT ID: CV4

Linear

 μ Abs.:

46950



A= 0.0000e+000

B= 2.1547e-004 slope

C= -2.3681e-002 y-intercept

Rho= 0.9998454

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.067	0.067	420	0.000	420					=
0.05	0.050					185	0				11
0.20	0.200	0.222	0.022	1141	0.0 %	1141					-2
2.50	2.500	2.459	-0.041	11521	0.0 %	11521					-2
5.00	5.000	4.920	-0.080	22944	0.0 %	22944					-1
7.50	7.500	7.440	-0.060	34638	0.0 %	34638					1
10.0	10.000	10.093	0.093	46950	0.0 %	46950					

LB123532
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	420	-		0 SFAM01.1	PPB	1/4/2023 16:01	Std
	0.2 S01	1141	-		0.2 SFAM01.1	PPB	1/4/2023 16:04	Std
	2.5 S02	11521	-		2.5 SFAM01.1	PPB	1/4/2023 16:06	Std
	5 S03	22944	-		5 SFAM01.1	PPB	1/4/2023 16:08	Std
	7.5 S04	34638	-		7.5 SFAM01.1	PPB	1/4/2023 16:10	Std
	10 S05	46950	-		10 SFAM01.1	PPB	1/4/2023 16:13	Std
ICV090	ICV090	19274	4.1293	-	SFAM01.1	PPB	1/4/2023 16:16	SMPL
ICB090	ICB090	480	0.0797	-	SFAM01.1	PPB	1/4/2023 16:18	SMPL
CCV001	CCV001	23157	4.9659	-	SFAM01.1	PPB	1/4/2023 16:20	SMPL
CCB001	CCB001	408	0.0642	-	SFAM01.1	PPB	1/4/2023 16:22	SMPL
PB149958BL	PBS958	418	0.0664	-	SFAM01.1	PPB	1/4/2023 16:25	SMPL
N6106-02	GBP19	227	0.0252	-	SFAM01.1	PPB	1/4/2023 16:27	SMPL
N6106-03	GBP19D	206	0.0207	-	SFAM01.1	PPB	1/4/2023 16:29	SMPL
N6106-04	GBP19S	11416	2.4361	-	SFAM01.1	PPB	1/4/2023 16:32	SMPL
N6106-05	GBP20	511	0.0864	-	SFAM01.1	PPB	1/4/2023 16:34	SMPL
N6106-06	GBP21	1733	0.3497	-	SFAM01.1	PPB	1/4/2023 16:36	SMPL
N6106-07	GBP22	138	0.0061	-	SFAM01.1	PPB	1/4/2023 16:38	SMPL
N6106-08	GBP23	123	0.0028	-	SFAM01.1	PPB	1/4/2023 16:41	SMPL
N6106-09	GBP24	365	0.055	-	SFAM01.1	PPB	1/4/2023 16:43	SMPL
N6106-11	GBP26	194	0.0181	-	SFAM01.1	PPB	1/4/2023 16:45	SMPL
N6106-12	GBP27	430	0.069	-	SFAM01.1	PPB	1/4/2023 16:47	SMPL
N6106-14	GBP29	849	0.1593	-	SFAM01.1	PPB	1/4/2023 16:50	SMPL
N6106-15	GBP30	994	0.1905	-	SFAM01.1	PPB	1/4/2023 16:52	SMPL
N6106-16	GBP31	895	0.1692	-	SFAM01.1	PPB	1/4/2023 16:54	SMPL
N6106-17	GBP32	3780	0.7908	-	SFAM01.1	PPB	1/4/2023 16:56	SMPL
N6106-19	GBP34	1574	0.3155	-	SFAM01.1	PPB	1/4/2023 16:59	SMPL
N6106-21	GBP36	293	0.0395	-	SFAM01.1	PPB	1/4/2023 17:01	SMPL
N6106-23	GBP98	21032	4.508	-	SFAM01.1	PPB	1/4/2023 17:03	SMPL
PB149961BL	PBS961	388	0.0599	-	SFAM01.1	PPB	1/4/2023 17:05	SMPL
N6113-01	GBP99	4421	0.9289	-	SFAM01.1	PPB	1/4/2023 17:08	SMPL
CCV002	CCV002	23026	4.9377	-	SFAM01.1	PPB	1/4/2023 17:10	SMPL
CCB002	CCB002	403	0.0632	-	SFAM01.1	PPB	1/4/2023 17:12	SMPL
N6113-03	GBPA1	2131	0.4355	-	SFAM01.1	PPB	1/4/2023 17:17	SMPL
N6113-04	GBPA2	2125	0.4342	-	SFAM01.1	PPB	1/4/2023 17:19	SMPL
N6113-05	GBPA3	2497	0.5143	-	SFAM01.1	PPB	1/4/2023 17:21	SMPL
N6113-06	GBPA4	841	0.1575	-	SFAM01.1	PPB	1/4/2023 17:24	SMPL
N6113-07	GBPA5	1188	0.2323	-	SFAM01.1	PPB	1/4/2023 17:26	SMPL
N6113-08	GBPA6	6996	1.4837	-	SFAM01.1	PPB	1/4/2023 17:28	SMPL
N6113-09	GBPA7	2281	0.4678	-	SFAM01.1	PPB	1/4/2023 17:30	SMPL
N6113-10	GBPA7D	1653	0.3325	-	SFAM01.1	PPB	1/4/2023 17:33	SMPL
N6113-11	GBPA7S	14402	3.0795	-	SFAM01.1	PPB	1/4/2023 17:35	SMPL
N6113-12	GBPA8	4962	1.0455	-	SFAM01.1	PPB	1/4/2023 17:37	SMPL
N6113-13	GBPA9	479	0.0795	-	SFAM01.1	PPB	1/4/2023 17:40	SMPL
N6113-14	GBPB0	469	0.0774	-	SFAM01.1	PPB	1/4/2023 17:42	SMPL
N6113-15	GBPB1	1201	0.2351	-	SFAM01.1	PPB	1/4/2023 17:44	SMPL
N6113-16	GBPB2	2112	0.4314	-	SFAM01.1	PPB	1/4/2023 17:46	SMPL

LB123532

INSTRUMENT ID : CV1

N6113-18	GBPB4	1791	0.3622 -	SFAM01.1	PPB	1/4/2023 17:51 SMPL
N6113-19	GBPB5	629	0.1118 -	SFAM01.1	PPB	1/4/2023 17:53 SMPL
N6113-20	GBPB6	777	0.1437 -	SFAM01.1	PPB	1/4/2023 17:55 SMPL
N6113-21	GBPB7	714	0.1302 -	SFAM01.1	PPB	1/4/2023 17:58 SMPL
CCV003	CCV003	23027	4.9379 -	SFAM01.1	PPB	1/4/2023 18:00 SMPL
CCB003	CCB003	380	0.0582 -	SFAM01.1	PPB	1/4/2023 18:02 SMPL
N6113-23	GBPH9	21864	4.6873 -	SFAM01.1	PPB	1/4/2023 18:07 SMPL
CCV004	CCV004	23050	4.9429 -	SFAM01.1	PPB	1/4/2023 18:23 SMPL
CCB004	CCB004	413	0.0653 -	SFAM01.1	PPB	1/4/2023 18:25 SMPL