



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123535

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.9	4.01	01/04/2023	17:00
2	CAL2	1	Water	NA	NA	20.9	7.00	01/04/2023	17:01
3	CAL3	1	Water	NA	NA	20.8	10.03	01/04/2023	17:02
4	ICV	1	Water	NA	NA	20.8	6.99	01/04/2023	17:05
5	CCV1	1	Water	NA	NA	20.8	2.01	01/04/2023	18:03
6	O1011-01	1	Water	NA	NA	20.5	7.06	01/04/2023	18:05
7	O1011-01DUP	1	Water	NA	NA	20.6	7.07	01/04/2023	18:06
8	O1019-01	1	Water	NA	NA	20.8	6.13	01/04/2023	18:09
9	O1019-02	1	Water	NA	NA	20.8	6.77	01/04/2023	18:11
10	O1019-03	1	Water	NA	NA	20.2	6.64	01/04/2023	18:15
11	O1028-14	1	Water	NA	NA	20.2	11.27	01/04/2023	18:20
12	O1028-15	1	Water	NA	NA	20.3	11.07	01/04/2023	18:23
13	CCV2	1	Water	NA	NA	20.6	12.03	01/04/2023	18:25

WORKLIST(Hardcopy Internal Chain)

123535

WorkList Name : PH O1011 WorkList ID : 166309 Department : Wet-Chemistry Date : 01-04-2023 16:41:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/04/2023	Water	O1011-01	pH	Cool 4 deg C	PSEG03	N23	NWB-1865	01/04/2023	9040C
01/04/2023	Water	O1019-01	pH	Cool 4 deg C	PSEG03	N23	33881	01/04/2023	9040C
01/04/2023	Water	O1019-02	pH	Cool 4 deg C	PSEG03	N23	33885	01/04/2023	9040C
01/04/2023	Water	O1019-03	pH	Cool 4 deg C	PSEG03	N23	33886	01/04/2023	9040C
01/04/2023	Water	O1028-14	pH	Cool 4 deg C	PSEG03	N23	MW-1D-PURGEWATER-A	01/04/2023	9040C
01/04/2023	Water	O1028-15	pH	Cool 4 deg C	PSEG03	N23	MW-1D-PURGEWATER-B	01/04/2023	9040C

Date/Time 01/04/23 16:50
 Raw Sample Received by: J.C. (swi)
 Raw Sample Relinquished by: J.C. (swi)

Date/Time 01/04/23 18:30
 Raw Sample Received by: J.C. (swi)
 Raw Sample Relinquished by: J.C. (swi)