

LB123537


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

53453



A= 0.0000e+000

B= 1.8778e-004

C= 1.2169e-002

Rho= 0.9998018

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%
0.0	0.000	0.072	0.072	318	0.000	318					2
0.2	0.200	0.205	0.005	1027	0.0 %	1027					0
2.5	2.500	2.493	-0.007	13210	0.0 %	13210					0
5.0	5.000	4.847	-0.153	25746	0.0 %	25746					3
7.5	7.500	7.534	0.034	40060	0.0 %	40060					0
10.0	10.000	10.049	0.049	53453	0.0 %	53453					0

LB123537 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 05 Jan 2023 08:57:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	318	-	0.0000	7471B	PPB	05 Jan 2023 08:57:51	S	Std
0.2 - 1	50.2	1027	-	0.2000	7471B	PPB	05 Jan 2023 09:00:08	S	Std
2.5 - 1	52.5	13210	-	2.5000	7471B	PPB	05 Jan 2023 09:02:25	S	Std
5.0 - 1	55	25746	-	5.0000	7471B	PPB	05 Jan 2023 09:04:42	S	Std
7.5 - 1	57.5	40060	-	7.5000	7471B	PPB	05 Jan 2023 09:06:58	S	Std
10.0 - 1	511	53453	-	10.0000	7471B	PPB	05 Jan 2023 09:09:14	S	Std
ICV91 - 1	ICV91	21000	3.9555	-	7471B	PPB	05 Jan 2023 09:11:57	U	SMPL
ICB91 - 1	ICB91	55	0.0225	-	7471B	PPB	05 Jan 2023 09:14:12	U	SMPL
CCV05 - 1	CCV05	25718	4.8414	-	7471B	PPB	05 Jan 2023 09:16:26	U	SMPL
CCB05 - 1	CCB05	22	0.0163	-	7471B	PPB	05 Jan 2023 09:18:41	U	SMPL
CRA - 1	CRA	993	0.1986	-	7471B	PPB	05 Jan 2023 09:20:57	U	SMPL
HighStd - 1	HighStd	53074	9.9782	-	7471B	PPB	05 Jan 2023 09:23:13	U	SMPL
ChkStd - 1	ChkStd	38331	7.2098	-	7471B	PPB	05 Jan 2023 09:25:29	U	SMPL
PB150050BL - 1	PB150050BL	395	0.0863	-	7471B	PPB	05 Jan 2023 09:27:45	U	SMPL
PB150050BS - 1	PB150050BS	22447	4.2272	-	7471B	PPB	05 Jan 2023 09:32:38	U	SMPL
N6244-01 - 1	PE-11	28998	5.4573	-	7471B	PPB	05 Jan 2023 09:34:53	U	SMPL
N6244-02 - 1	PE-12	2472	0.4763	-	7471B	PPB	05 Jan 2023 09:37:10	U	SMPL
N6244-03 - 1	PE-13	5599	1.0635	-	7471B	PPB	05 Jan 2023 09:39:27	U	SMPL
N6244-04 - 1	PE-14	27478	5.1719	-	7471B	PPB	05 Jan 2023 09:41:43	U	SMPL
N6244-05 - 1	PE-15	424	0.0918	-	7471B	PPB	05 Jan 2023 09:43:59	U	SMPL
CCV06 - 1	CCV06	24980	4.7028	-	7471B	PPB	05 Jan 2023 09:46:14	U	SMPL
CCB06 - 1	CCB06	-48	0.0032	-	7471B	PPB	05 Jan 2023 09:48:30	U	SMPL
N6244-06 - 1	PE-16	53271	10.0152	-	7471B	PPB	05 Jan 2023 09:50:45	U	SMPL
N6244-07 - 1	PE-17	26207	4.9332	-	7471B	PPB	05 Jan 2023 09:53:01	U	SMPL
N6244-08 - 1	PE-18	56790	10.6759	-	7471B	PPB	05 Jan 2023 09:55:17	U	SMPL
N6245-01 - 1	WASTE-COMP	38492	7.2400	-	7471B	PPB	05 Jan 2023 09:57:33	U	SMPL
O1001-01 - 1	TP-C2	898	0.1808	-	7471B	PPB	05 Jan 2023 09:59:49	U	SMPL
O1001-01MS - 1	TP-C2MS	23696	4.4617	-	7471B	PPB	05 Jan 2023 10:04:22	U	SMPL
O1001-01MSD - 1	TP-C2MSD	22508	4.2386	-	7471B	PPB	05 Jan 2023 10:06:39	U	SMPL
O1001-01DUP - 1	TP-C2DUP	878	0.1770	-	7471B	PPB	05 Jan 2023 10:09:11	U	SMPL
O1001-04 - 1	TP-C3	2629	0.5058	-	7471B	PPB	05 Jan 2023 10:12:24	U	SMPL
O1004-01 - 1	CONTAINMENT-STONES	226	0.0546	-	7471B	PPB	05 Jan 2023 10:14:39	U	SMPL
CCV07 - 1	CCV07	25656	4.8297	-	7471B	PPB	05 Jan 2023 10:16:54	U	SMPL
CCB07 - 1	CCB07	-89	-0.0045	-	7471B	PPB	05 Jan 2023 10:19:10	U	SMPL
O1005-01 - 1	PL-01-10323	738	0.1507	-	7471B	PPB	05 Jan 2023 10:21:25	U	SMPL
O1008-01 - 1	EO-1-010323	5327	1.0124	-	7471B	PPB	05 Jan 2023 10:23:41	U	SMPL
O1001-01LX5 - 1		96	0.0302	-	7471B	PPB	05 Jan 2023 10:25:57	U	SMPL
O1001-01A - 1		21604	4.0689	-	7471B	PPB	05 Jan 2023 10:28:12	U	SMPL
N6038-14DLX5 - 1		16822	3.1709	-	7471B	PPB	05 Jan 2023 10:32:45	U	SMPL
N6244-06DLX2 - 1		26892	5.0618	-	7471B	PPB	05 Jan 2023 10:35:03	U	SMPL
N6244-08DLX2 - 1		27796	5.2316	-	7471B	PPB	05 Jan 2023 10:37:20	U	SMPL
CCV08 - 1	CCV08	24760	4.6615	-	7471B	PPB	05 Jan 2023 10:39:37	U	SMPL
CCB08 - 1	CCB08	13	0.0146	-	7471B	PPB	05 Jan 2023 10:41:53	U	SMPL