



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123538

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	01/05/2023	08:50
2	O1011-01		1	12	100.00	>212.0	>212.0	01/05/2023	09:20
3	O1011-01DUP		1	13	100.00	>212.0	>212.0	01/05/2023	09:50
4	O1019-01		1	12	100.00	>212.0	>212.0	01/05/2023	10:20
5	O1019-02		1	14	100.00	>212.0	>212.0	01/05/2023	10:50
6	O1019-03		1	12	100.00	>212.0	>212.0	01/05/2023	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6123538

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp1-4

WorkList ID : 166300

Department : Wet-Chemistry

Date : 01-04-2023 11:30:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/11/2023	Water	O1011-01	Flash Point	Cool 4 deg C	PSEG03	N23	NWB-1865	01/04/2023	1010B
01/09/2023	Water	O1019-01	Flash Point	Cool 4 deg C	PSEG03	N23	33881	01/04/2023	1010B
01/09/2023	Water	O1019-02	Flash Point	Cool 4 deg C	PSEG03	N23	33885	01/04/2023	1010B
01/09/2023	Water	O1019-03	Flash Point	Cool 4 deg C	PSEG03	N23	33886	01/04/2023	1010B

Date/Time 01/05/2023 08.40
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RS 1614-1615

Date/Time 01/05/2023 12.00
 Raw Sample Received by: RS C 1614-1615
 Raw Sample Relinquished by: RM CWC