

=====13123551=====

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

1/5/2023 14:51

Test: Total CN

Sample Id	^{mg/L} Result	Dil. 1 +	Response	Errors
ICV1	96.165	0.0	0.063	
ICB1	1.259	0.0	0.001	
CCV1	253.366	0.0	0.165	
CCB1	1.152	0.0	0.001	
PB150071BL	1.098	0.0	0.001	
O1023-02	1.471	0.0	0.001	
O1023-02DUP	1.132	0.0	0.001	
O1025-02	1.499	0.0	0.001	
CCV2	247.672	0.0	0.161	
CCB2	1.320	0.0	0.001	

N	10
Mean	60.613
SD	104.3793
CV%	172.20

Aquakem v. 7.2AQ1

Results from time period:

Thu Jan 05 14:41:43 2023

Thu Jan 05 14:46:41 2023

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	1.0892	µg/l	1/5/2023 10:56:53	
5.0PPBCN	A	Total CN	P	5.2172	µg/l	1/5/2023 10:56:54	
10PPBCN	A	Total CN	P	9.3413	µg/l	1/5/2023 10:56:55	
50PPBCN	A	Total CN	P	49.914	µg/l	1/5/2023 10:56:56	
100PPBCN	A	Total CN	P	97.1273	µg/l	1/5/2023 10:56:57	
250PPBCN	A	Total CN	P	253.4338	µg/l	1/5/2023 10:56:58	
500PPBCN	A	Total CN	P	498.8773	µg/l	1/5/2023 10:56:59	
ICV1	S	Total CN	P	96.1646	µg/l	1/5/2023 14:41:44	
ICB1	S	Total CN	P	1.2593	µg/l	1/5/2023 14:41:46	
CCV1	S	Total CN	P	253.3665	µg/l	1/5/2023 14:41:48	
CCB1	S	Total CN	P	1.1524	µg/l	1/5/2023 14:41:49	
PB150071BL	S	Total CN	P	1.098	µg/l	1/5/2023 14:41:52	
O1023-02	S	Total CN	P	1.471	µg/l	1/5/2023 14:46:37	
O1023-02DUP	S	Total CN	P	1.1315	µg/l	1/5/2023 14:46:38	
O1025-02	S	Total CN	P	1.4993	µg/l	1/5/2023 14:46:39	
CCV2	S	Total CN	P	247.6721	µg/l	1/5/2023 14:46:40	
CCB2	S	Total CN	P	1.3197	µg/l	1/5/2023 14:46:41	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

1/5/2023 10:57


Test Total CN

Accepted 1/5/2023 10:57

Factor 1544
Bias 0.001

Coeff. of det. 0.999888

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.001	1.0892	0.0000	—
2	5.0PPBCN	0.004	5.2172	5.0000	4.3
3	10PPBCN	0.007	9.3413	10.0000	-6.6
4	50PPBCN	0.033	49.9140	50.0000	-0.2
5	100PPBCN	0.063	97.1273	100.0000	-2.9
6	250PPBCN	0.165	253.4338	250.0000	1.4
7	500PPBCN	0.324	498.8773	500.0000	-0.2

1/5/23