

LB123555


7471B

INSTRUMENT ID: CX1

Linear

μ Abs.:

51549



A= 0.0000e+000

B= 1.9215e-004

C= -3.1514e-002

Rho= 0.9997689

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.016	-0.016	81	0.000	81					-
0.2	0.200	0.161	-0.039	1002	0.0 %	1002					-20
2.5	2.500	2.486	-0.014	13100	0.0 %	13100					-1
5.0	5.000	5.079	0.079	26597	0.0 %	26597					2
7.5	7.500	7.616	0.116	39802	0.0 %	39802					2
10.0	10.000	9.874	-0.126	51549	0.0 %	51549					-1

LB123555 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 06 Jan 2023 08:16:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	81	-	0.0000	7471B	PPB	06 Jan 2023 08:20:37	S	Std
0.2 - 1	50-2	1002	-	0.2000	7471B	PPB	06 Jan 2023 08:22:55	S	Std
2.5 - 1	52.5	13100	-	2.5000	7471B	PPB	06 Jan 2023 08:25:12	S	Std
5.0 - 1	55	26597	-	5.0000	7471B	PPB	06 Jan 2023 08:27:29	S	Std
7.5 - 1	57.5	39802	-	7.5000	7471B	PPB	06 Jan 2023 08:29:45	S	Std
10.0 - 1	58	51549	-	10.0000	7471B	PPB	06 Jan 2023 08:36:12	S	Std
ICV92 - 1	ICV92	21589	4.1168	-	7471B	PPB	06 Jan 2023 08:40:28	U	SMPL
ICB92 - 1	ICB92	29	-0.0259	-	7471B	PPB	06 Jan 2023 08:42:43	U	SMPL
CCV09 - 1	CCV09	26763	5.1110	-	7471B	PPB	06 Jan 2023 08:44:57	U	SMPL
CCB09 - 1	CCB09	-37	-0.0386	-	7471B	PPB	06 Jan 2023 08:47:13	U	SMPL
CRA - 1	CRA	1211	0.2012	-	7471B	PPB	06 Jan 2023 08:55:29	U	SMPL
HighStd - 1	HighStd	51791	9.9202	-	7471B	PPB	06 Jan 2023 08:57:44	U	SMPL
ChkStd - 1	ChkStd	36066	6.8986	-	7471B	PPB	06 Jan 2023 09:00:00	U	SMPL
PB150017BL - 1	PBS	-66	-0.0423	-	7471B	PPB	06 Jan 2023 09:02:17	U	SMPL
PB150017BS - 1	LCSS	20284	3.8661	-	7471B	PPB	06 Jan 2023 09:04:34	U	SMPL
N6243-01 - 1	PE-1	11039	2.0896	-	7471B	PPB	06 Jan 2023 09:14:06	U	SMPL
N6243-01DUP - 1	PE-1DUP	11234	2.1271	-	7471B	PPB	06 Jan 2023 09:16:25	U	SMPL
N6243-01MS - 1	PE-1MS	33547	6.4146	-	7471B	PPB	06 Jan 2023 09:18:43	U	SMPL
N6243-01MSD - 1	PE-1MSD	32267	6.1686	-	7471B	PPB	06 Jan 2023 09:21:00	U	SMPL
N6243-02 - 1	PE-2	9036	1.7048	-	7471B	PPB	06 Jan 2023 09:23:39	U	SMPL
CCV10 - 1	CCV10	26673	5.0937	-	7471B	PPB	06 Jan 2023 09:25:55	U	SMPL
CCB10 - 1	CCB10	-47	-0.0405	-	7471B	PPB	06 Jan 2023 09:28:11	U	SMPL
N6243-03 - 1	PE-3	12318	2.3354	-	7471B	PPB	06 Jan 2023 09:30:27	U	SMPL
N6243-04 - 1	PE-4	12925	2.4520	-	7471B	PPB	06 Jan 2023 09:32:42	U	SMPL
N6243-05 - 1	PE-5	35201	6.7324	-	7471B	PPB	06 Jan 2023 09:34:58	U	SMPL
N6243-06 - 1	PE-6	11435	2.1657	-	7471B	PPB	06 Jan 2023 09:37:14	U	SMPL
N6243-07 - 1	PE-7	5782	1.0795	-	7471B	PPB	06 Jan 2023 09:39:30	U	SMPL
N6243-08 - 1	PE-8	27209	5.1967	-	7471B	PPB	06 Jan 2023 09:41:46	U	SMPL
N6243-09 - 1	PE-9	59770	11.4533	-	7471B	PPB	06 Jan 2023 09:44:03	U	SMPL
N6243-10 - 1	PE-10	24072	4.5939	-	7471B	PPB	06 Jan 2023 09:46:19	U	SMPL
O1009-01 - 1	TP-4	7705	1.4490	-	7471B	PPB	06 Jan 2023 09:48:37	U	SMPL
O1027-01 - 1	COMP-1-1-4-2023	2173	0.3860	-	7471B	PPB	06 Jan 2023 09:50:53	U	SMPL
CCV11 - 1	CCV11	26677	5.0945	-	7471B	PPB	06 Jan 2023 09:53:08	U	SMPL
CCB11 - 1	CCB11	-36	-0.0384	-	7471B	PPB	06 Jan 2023 09:55:24	U	SMPL
O1029-01 - 1	TP-205	4022	0.7413	-	7471B	PPB	06 Jan 2023 09:57:40	U	SMPL
O1029-04 - 1	TP-204	4079	0.7523	-	7471B	PPB	06 Jan 2023 09:59:55	U	SMPL
O1030-01 - 1	TP-5	12293	2.3306	-	7471B	PPB	06 Jan 2023 10:02:11	U	SMPL
O1033-01 - 1	VNJ-228	2320	0.4143	-	7471B	PPB	06 Jan 2023 10:04:27	U	SMPL
N6243-01LX5 - 1		2237	0.3983	-	7471B	PPB	06 Jan 2023 10:06:42	U	SMPL
N6243-01A - 1		32960	6.3018	-	7471B	PPB	06 Jan 2023 10:08:59	U	SMPL
N6243-09DLX2 - 1		29939	5.7213	-	7471B	PPB	06 Jan 2023 10:11:15	U	SMPL
CCV12 - 1	CCV12	26677	5.0945	-	7471B	PPB	06 Jan 2023 10:13:32	U	SMPL
CCB12 - 1	CCB12	1	-0.0313	-	7471B	PPB	06 Jan 2023 10:15:50	U	SMPL