



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123557

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	01/06/2023	09:00
2	CAL2	1	Water	NA	NA	20.7	7.00	01/06/2023	09:01
3	CAL3	1	Water	NA	NA	20.7	10.03	01/06/2023	09:02
4	ICV	1	Water	NA	NA	20.7	6.99	01/06/2023	09:05
5	CCV1	1	Water	NA	NA	20.7	2.01	01/06/2023	09:06
6	O1037-01	1	Water	NA	NA	20.1	3.76	01/06/2023	09:10
7	O1037-02	1	Water	NA	NA	20.3	3.66	01/06/2023	09:11
8	O1038-02	1	Water	NA	NA	20.4	10.26	01/06/2023	09:15
9	O1038-03	1	Water	NA	NA	20.5	10.26	01/06/2023	09:16
10	O1038-04	1	Water	NA	NA	20.7	10.25	01/06/2023	09:19
11	O1038-05	1	Water	NA	NA	20.8	10.20	01/06/2023	09:20
12	O1038-06	1	Water	NA	NA	20.5	8.78	01/06/2023	09:23
13	O1038-07	1	Water	NA	NA	20.3	7.15	01/06/2023	09:25
14	O1038-08	1	Water	NA	NA	20.2	7.33	01/06/2023	09:26
15	O1038-09	1	Water	NA	NA	20.3	7.44	01/06/2023	09:30
16	CCV2	1	Water	NA	NA	20.8	12.02	01/06/2023	09:33
17	O1038-10	1	Water	NA	NA	20.4	7.27	01/06/2023	09:35
18	O1044-01	1	Water	NA	NA	20.6	7.14	01/06/2023	09:40
19	O1045-01	1	Water	NA	NA	20.6	7.49	01/06/2023	09:43
20	O1046-01	1	Water	NA	NA	20.6	7.52	01/06/2023	09:45

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	O1046-01DUP	1	Water	NA	NA	20.7	7.53	01/06/2023	09:46
22	CCV3	1	Water	NA	NA	20.7	2.01	01/06/2023	09:50

W 123558

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O1045 WorkList ID : 166359 Department : Wet-Chemistry Date : 01-06-2023 08:45:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/04/2023	Water	O1037-01	pH	A	Cool 4 deg C	STER16	N23	MAIN-CELL	01/04/2023 9040C
01/04/2023	Water	O1037-02	pH	A	Cool 4 deg C	STER16	N23	EX-CELL	01/04/2023 9040C
01/04/2023	Water	O1038-02	pH	A	Cool 4 deg C	ATCE02	N31	1X1-GREEN-POWER	01/04/2023 9040C
01/04/2023	Water	O1038-03	pH	A	Cool 4 deg C	ATCE02	N31	1X2-GREEN-POWER	01/04/2023 9040C
01/04/2023	Water	O1038-04	pH	A	Cool 4 deg C	ATCE02	N31	1X3-GREEN-POWER	01/04/2023 9040C
01/04/2023	Water	O1038-05	pH	A	Cool 4 deg C	ATCE02	N31	1X4-GREEN-POWER	01/04/2023 9040C
01/04/2023	Water	O1038-06	pH	A	Cool 4 deg C	ATCE02	N31	1X5-GREEN-POWER	01/04/2023 9040C
01/04/2023	Water	O1038-07	pH	A	Cool 4 deg C	ATCE02	N31	1X6-GREEN-POWER	01/04/2023 9040C
01/04/2023	Water	O1038-08	pH	A	Cool 4 deg C	ATCE02	N31	1X7-GREEN-POWER	01/04/2023 9040C
01/04/2023	Water	O1038-09	pH	A	Cool 4 deg C	ATCE02	N31	1X8-GREEN-POWER	01/04/2023 9040C
01/04/2023	Water	O1038-10	pH	A	Cool 4 deg C	ATCE02	N31	1X9-GREEN-POWER	01/04/2023 9040C
01/04/2023	Water	O1044-01	pH	D	Cool 4 deg C	ENV170	N31	COMP-SAMPLE-TANK-10-34-35	01/04/2023 9040C
01/04/2023	Water	O1045-01	pH	D	Cool 4 deg C	ENV170	N31	COMP-SAMPLE-TANK-11-12-31-32-33	01/04/2023 9040C
01/04/2023	Water	O1046-01	pH	D	Cool 4 deg C	ENV170	N31	COMP-SAMPLE-API-3	01/04/2023 9040C

Date/Time 01/06/23 08:50
 Raw Sample Received by: J P (wpc)
 Raw Sample Relinquished by: J P (wpc)

Date/Time 01-06-23 11:00
 Raw Sample Received by: J P (wpc)
 Raw Sample Relinquished by: J P (wpc)