



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/9/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 17:15
In Date: 01/06/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 01/07/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123562

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1047-01	TP-6	1	0.11	9.64	9.75	8.92	91.4	
O1047-03	TP-6-EPH-2	2	0.11	9.64	9.75	8.92	91.4	
O1051-01	72-11933	7	1.15	8.58	9.73	9.51	97.4	
O1051-02	72-11933-EPH-2	8	1.13	8.66	9.79	9.52	96.9	
O1053-01	TP-201	3	1.15	8.80	9.95	8.14	79.4	
O1053-03	TP-201-EPH-2	4	1.14	8.82	9.96	8.15	79.5	
O1054-01	COMP-1	10	1.00	1.00	2.00	2.00	100.0	DEBRIS
O1054-02	COMP-2	11	1.00	1.00	2.00	2.00	100.0	DEBRIS
O1054-03	2016	12	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
O1054-04	2029	13	1.00	1.00	2.00	2.00	100.0	DEBRIS
O1061-01	SLUDGE-4	14	1.14	8.79	9.93	6.88	65.3	
O1061-02	SLUDGE-5	15	1.16	8.80	9.96	5.77	52.4	
O1061-03	SLUDGE-6	16	1.19	8.74	9.93	6.23	57.7	
O1061-04	SLUDGE-7	17	1.16	8.56	9.72	5.75	53.6	
O1061-05	SLUDGE-COMP	18	1.12	8.52	9.64	5.69	53.6	
O1061-06	1623	19	1.00	1.00	2.00	2.00	100.0	DEBRIS
O1061-07	1624	20	1.00	1.00	2.00	2.00	100.0	DEBRIS
O1061-08	1623-1624	21	1.00	1.00	2.00	2.00	100.0	DEBRIS
O1061-09	12123	22	1.00	1.00	2.00	2.00	100.0	DEBRIS
O1065-01	VNJ-210	9	1.14	8.76	9.9	8.04	78.8	
O1067-01	31920	23	1.00	1.00	2.00	2.00	100.0	DEBRIS
O1067-02	7822	24	1.15	8.54	9.69	9.49	97.7	
O1067-03	31921	25	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
O1067-05	31500	26	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
O1067-07	31495	27	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
O1067-09	31499	28	1.00	1.00	2.00	2.00	100.0	OIL SAMPLE
O1068-01	S-1	5	1.13	8.78	9.91	8.51	84.1	
O1068-02	S-2	6	1.16	8.58	9.74	8.00	79.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$