



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123565

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	01/06/2023	14:20
2	O1055-01		1	12	100.00	>212.0	>212.0	01/06/2023	14:45
3	O1055-01DUP		1	12	100.00	>212.0	>212.0	01/06/2023	15:15
4	O1055-03		1	14	100.00	>212.0	>212.0	01/06/2023	15:45
5	O1055-05		1	13	100.00	>212.0	>212.0	01/06/2023	16:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66123565

WorkList Name : fp1-6

WorkList ID : 166375

Department : Wet-Chemistry

Date : 01-06-2023 11:40:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/13/2023	Water	O1055-01	Flash Point	Cool 4 deg C	TULL02	N31	MH-J	01/06/2023	1010B
01/13/2023	Water	O1055-03	Flash Point	Cool 4 deg C	TULL02	N31	MH-F	01/06/2023	1010B
01/13/2023	Water	O1055-05	Flash Point	Cool 4 deg C	TULL02	N31	MH-G	01/06/2023	1010B

Date/Time 01/06/2023 14:15
 Raw Sample Received by: RM CWO
 Raw Sample Relinquished by: RJ CWO

Date/Time 01/06/2023 16:50
 Raw Sample Received by: RJ CWO
 Raw Sample Relinquished by: RM CWO