


LB123566

SFAM01.1 INSTRUMENT ID: CV1

Linear

 μ Abs.:

44980



A= 0.0000e+000

B= 2.2141e-004 slope

C= 6.7371e-003 y-intercept

Rho= 0.9999749

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	9.D
0.00	0.000	0.010	0.010	13	0.000	13					
0.05	0.050					185	0				-10
0.20	0.200	0.181	-0.019	787	0.0 %	787					-10
2.50	2.500	2.480	-0.020	11169	0.0 %	11169					-10
5.00	5.000	5.035	0.035	22709	0.0 %	22709					-10
7.50	7.500	7.529	0.029	33975	0.0 %	33975					-10
10.0	10.000	9.966	-0.034	44980	0.0 %	44980					-10

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0		13	-		0 SFAM01.1	PPB	1/6/2023 15:24	Std
0.2		787	-		0.2 SFAM01.1	PPB	1/6/2023 15:26	Std
2.5		11169	-		2.5 SFAM01.1	PPB	1/6/2023 15:29	Std
5		22709	-		5 SFAM01.1	PPB	1/6/2023 15:31	Std
7.5		33975	-		7.5 SFAM01.1	PPB	1/6/2023 15:33	Std
10		44980	-		10 SFAM01.1	PPB	1/6/2023 15:41	Std
ICV094	ICV094	18142	4.0236	-	SFAM01.1	PPB	1/6/2023 15:43	SMPL
ICB094	ICB094	50	0.0178	-	SFAM01.1	PPB	1/6/2023 15:46	SMPL
CCV017	CCV017	22623	5.0157	-	SFAM01.1	PPB	1/6/2023 15:48	SMPL
CCB017	CCB017	-34	-0.0008	-	SFAM01.1	PPB	1/6/2023 15:50	SMPL
PB150102BL	PBW102	-29	0.0003	-	SFAM01.1	PPB	1/6/2023 15:52	SMPL
N6158-01	GBQ57	45	0.0167	-	SFAM01.1	PPB	1/6/2023 15:55	SMPL
N6158-02	GBQ58	-6	0.0054	-	SFAM01.1	PPB	1/6/2023 15:57	SMPL
N6158-03	GBQ59	-21	0.0021	-	SFAM01.1	PPB	1/6/2023 15:59	SMPL
N6158-04	GBQ60	-41	-0.0023	-	SFAM01.1	PPB	1/6/2023 16:01	SMPL
N6158-05	GBQ61	-16	0.0032	-	SFAM01.1	PPB	1/6/2023 16:04	SMPL
N6158-06	GBQ62	-33	-0.0006	-	SFAM01.1	PPB	1/6/2023 16:06	SMPL
N6158-07	GBQ63	-39	-0.0019	-	SFAM01.1	PPB	1/6/2023 16:08	SMPL
N6158-08	GBQ64	3	0.0074	-	SFAM01.1	PPB	1/6/2023 16:10	SMPL
N6158-09	GBQ65	20	0.0112	-	SFAM01.1	PPB	1/6/2023 16:13	SMPL
N6158-10	GBQ66	15	0.0101	-	SFAM01.1	PPB	1/6/2023 16:15	SMPL
N6158-11	GBQ67	380	0.0909	-	SFAM01.1	PPB	1/6/2023 16:17	SMPL
N6158-12	GBQ68	35	0.0145	-	SFAM01.1	PPB	1/6/2023 16:19	SMPL
N6158-13	GBQ69	89	0.0264	-	SFAM01.1	PPB	1/6/2023 16:22	SMPL
N6158-14	GBQ70	58	0.0196	-	SFAM01.1	PPB	1/6/2023 16:24	SMPL
N6158-15	GBQ71	50	0.0178	-	SFAM01.1	PPB	1/6/2023 16:26	SMPL
N6158-16	GBQ72	76	0.0236	-	SFAM01.1	PPB	1/6/2023 16:29	SMPL
N6158-17	GBQ73	60	0.02	-	SFAM01.1	PPB	1/6/2023 16:31	SMPL
N6158-18	GBQ73D	60	0.02	-	SFAM01.1	PPB	1/6/2023 16:33	SMPL
N6158-19	GBQ73S	4218	0.9406	-	SFAM01.1	PPB	1/6/2023 16:35	SMPL
CCV018	CCV018	22639	5.0193	-	SFAM01.1	PPB	1/6/2023 16:38	SMPL
CCB018	CCB018	21	0.0114	-	SFAM01.1	PPB	1/6/2023 16:40	SMPL
N6158-20	GBQ74	55	0.0189	-	SFAM01.1	PPB	1/6/2023 16:42	SMPL
N6158-21	GBQ75	66	0.0214	-	SFAM01.1	PPB	1/6/2023 16:44	SMPL
PB150103BL	PBW103	43	0.0163	-	SFAM01.1	PPB	1/6/2023 16:47	SMPL
N6159-01	GBQ76	63	0.0207	-	SFAM01.1	PPB	1/6/2023 16:49	SMPL
N6159-02	GBQ77	53	0.0185	-	SFAM01.1	PPB	1/6/2023 16:51	SMPL
N6159-03	GBQ78	52	0.0183	-	SFAM01.1	PPB	1/6/2023 16:53	SMPL
N6159-04	GBQ79	26	0.0125	-	SFAM01.1	PPB	1/6/2023 16:56	SMPL
N6159-05	GBQ80	44	0.0165	-	SFAM01.1	PPB	1/6/2023 16:58	SMPL
N6159-06	GBQ81	26	0.0125	-	SFAM01.1	PPB	1/6/2023 17:00	SMPL
N6159-07	GBQ82	21	0.0114	-	SFAM01.1	PPB	1/6/2023 17:03	SMPL
N6159-09	GBQ84	4	0.0076	-	SFAM01.1	PPB	1/6/2023 17:07	SMPL
N6159-10	GBQ85	16	0.0103	-	SFAM01.1	PPB	1/6/2023 17:09	SMPL
N6159-11	GBQ85D	41	0.0158	-	SFAM01.1	PPB	1/6/2023 17:12	SMPL
N6159-12	GBQ85S	4505	1.0042	-	SFAM01.1	PPB	1/6/2023 17:14	SMPL

LB123566

INSTRUMENT ID : CV1

N6159-08	GBQ83	22707	5.0343 -	SFAM01.1	PPB	1/6/2023 17:18 SMPL
CCV019	CCV019	22294	4.9429 -	SFAM01.1	PPB	1/6/2023 17:21 SMPL
CCB019	CCB019	4	0.0076 -	SFAM01.1	PPB	1/6/2023 17:23 SMPL