



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123576

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	01/06/2023	16:15
2	CAL2	1	Water	NA	NA	20.7	7.00	01/06/2023	16:16
3	CAL3	1	Water	NA	NA	20.8	10.02	01/06/2023	16:20
4	ICV	1	Water	NA	NA	20.7	6.99	01/06/2023	16:22
5	CCV1	1	Water	NA	NA	20.7	2.01	01/06/2023	16:25
6	O1052-01	1	Water	NA	NA	20.4	6.38	01/06/2023	16:30
7	O1055-01	1	Water	NA	NA	20.1	8.27	01/06/2023	16:33
8	O1055-03	1	Water	NA	NA	20.4	8.20	01/06/2023	16:35
9	O1055-05	1	Water	NA	NA	20.6	7.80	01/06/2023	16:40
10	O1061-10	1	Water	NA	NA	20.9	7.34	01/06/2023	16:45
11	O1061-11	1	Water	NA	NA	21.3	7.42	01/06/2023	16:46
12	O1066-01	1	Water	NA	NA	20.9	7.56	01/06/2023	16:50
13	O1066-01DUP	1	Water	NA	NA	20.8	7.57	01/06/2023	16:51
14	CCV2	1	Water	NA	NA	20.7	12.02	01/06/2023	17:00

WB 123576

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O1055 WorkList ID : 166377 Department : Wet-Chemistry Date : 01-06-2023 13:48:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/06/2023	Water	O1052-01	pH <u>R</u>	Cool 4 deg C	PSEG03	N31	UST-DRUM	01/06/2023	9040C
01/06/2023	Water	O1055-01	pH <u>E</u>	Cool 4 deg C	TULL02	N31	MH-J	01/06/2023	9040C
01/06/2023	Water	O1055-03	pH <u>E</u>	Cool 4 deg C	TULL02	N31	MH-F	01/06/2023	9040C
01/06/2023	Water	O1055-05	pH <u>E</u>	Cool 4 deg C	TULL02	N31	MH-G	01/06/2023	9040C
01/06/2023	Water	O1061-10	pH <u>C</u>	Cool 4 deg C	PSEG03	N31	TOTE-2	01/06/2023	9040C
01/06/2023	Water	O1061-11	pH <u>C</u>	Cool 4 deg C	PSEG03	N31	TOTE-3	01/06/2023	9040C
01/06/2023	Water	O1066-01	pH <u>K</u>	Cool 4 deg C	PSEG03	N31	OILY-WATER-1	01/06/2023	9040C

Date/Time 01-06-23 16:10
Raw Sample Received by: JP (WCI)
Raw Sample Relinquished by: RJ (ETA-106)

Date/Time 01-06-23 18:00
Raw Sample Received by: RJ (ETA-106)
Raw Sample Relinquished by: JP (WCI)