



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB123577

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 99.3
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	01/08/2023	11:15
2	CAL2	1	Water	NA	NA	20.8	7.00	01/08/2023	11:16
3	CAL3	1	Water	NA	NA	20.8	10.03	01/08/2023	11:16
4	ICV	1	Water	NA	NA	20.8	6.99	01/08/2023	11:20
5	CCV1	1	Water	NA	NA	20.8	2.01	01/08/2023	11:25
6	O1054-05	1	Water	NA	NA	21.2	7.79	01/08/2023	11:30
7	O1054-06	1	Water	NA	NA	22.1	7.30	01/08/2023	11:32
8	O1054-07	1	Water	NA	NA	22.2	6.51	01/08/2023	11:35
9	O1054-07DUP	1	Water	NA	NA	22.3	6.52	01/08/2023	11:36
10	CCV2	1	Water	NA	NA	20.9	12.02	01/08/2023	11:39

WORKLIST(Hardcopy Internal Chain)

1235ff

WorkList Name : CORROSIONITY O1054 V WorkList ID : 166389 Department : Wet-Chemistry Date : 01-08-2023 11:03:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/06/2023	Water	O1054-05	Corrosivity	Cool 4 deg C	PSEG03	N31	TOTE-COMP	01/06/2023	9040C
01/06/2023	Water	O1054-06	Corrosivity	Cool 4 deg C	PSEG03	N31	COMP-03	01/06/2023	9040C
01/06/2023	Water	O1054-07	Corrosivity	Cool 4 deg C	PSEG03	N31	2026	01/06/2023	9040C

Date/Time 01-08-23 11:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 01-08-23 12:30
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]