




Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB123578
Analysis Date: 01/08/2023
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 01/08/2023
Extraction IN Time: 10:10
Extraction OUT Time: 10:45
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB123578BL	LB123578BL	WATER	1.3	1000	100	3.0529	3.0529	3.02	3.0529	3.0529	0.0000	0
2	LB123578BS	LB123578BS	WATER	1.3	1000	100	3.1957	3.1957	3.01	3.2124	3.2124	0.0167	16.7
3	LB123578BSD	LB123578BSD	WATER	1.3	1000	100	3.1477	3.1477	3.03	3.1645	3.1645	0.0168	16.8
4	O1052-01	UST-DRUM	WATER	1.6	1000	100	3.0723	3.0723	24.05	3.2652	3.2652	0.1929	192.9
5	O1066-01	OILY-WATER-1	WATER	1.3	1000	100	3.1338	3.1338	3.02	3.1354	3.1354	0.0016	1.6



Test: TPH

Analysis Date: 01/08/2023

Chemical Name	Chemical Lot #
HEXANE	W2969
pH Paper 0-14	M4909
Sodium Sulfate	EP2279
1:1 HCL	WP10035
Silica Gel	W2839
Sand	NA

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP100036
LCSWD	5.00 ML	WP100037
MS/MSD	NA	NA

Analytical Balance ID # : WC SC-6

0.0020 gram Balance:	<u>0.0019</u>	(0.0018-0.0022)	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1 :	<u>12:26</u>
1.0000 gram Balance:	<u>1.0004</u>	(0.9950-1.0050)	In Time1:	<u>11:30</u>		
Bal Check Time:	<u>10:30</u>		Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1:	<u>13:00</u>
			Out Time1:	<u>12:25</u>		

0.0020 gram Balance:	<u>0.0021</u>	(0.0018-0.0022)	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2 :	<u>14:01</u>
1.0000 gram Balance:	<u>1.0003</u>	(0.9950-1.0050)	In Time2:	<u>13:30</u>		
Bal Check Time:	<u>14:45</u>		Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2:	<u>14:37</u>
			Out Time2:	<u>14:00</u>		

123578

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH O1066 WorkList ID : 166388 Department : Wet-Chemistry Date : 01-08-2023 09:39:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/12/2023	Water	O1052-01	TPH	Conc H2SO4 to	PSEG03	N31	UST-DRUM	01/06/2023	1664A
01/11/2023	Water	O1066-01	TPH	Conc H2SO4 to	PSEG03	N31	OILY-WATER-1	01/06/2023	1664A

Date/Time 01-08-23 09:40
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 01-08-23 12:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]