



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123584

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	01/09/2023	08:40
2	O1052-01		1	12	100.00	>212.0	>212.0	01/09/2023	09:10
3	O1052-01DUP		1	12	100.00	>212.0	>212.0	01/09/2023	09:40
4	O1054-05		1	14	100.00	>212.0	>212.0	01/09/2023	10:10
5	O1054-06		1	13	100.00	>212.0	>212.0	01/09/2023	10:40
6	O1054-07		1	13	100.00	>212.0	>212.0	01/09/2023	11:10
7	O1061-12		1	14	100.00	>212.0	>212.0	01/09/2023	11:40
8	O1066-01		1	14	100.00	>212.0	>212.0	01/09/2023	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6123584

WorkList Name : fp1-9

WorkList ID : 166393

Department : Wet-Chemistry

Date : 01-09-2023 08:25:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/12/2023	Water	O1052-01	✓ Flash Point	Cool 4 deg C	PSEG03	N31	UST-DRUM	01/06/2023	1010B
01/11/2023	Water	O1054-05	Δ Flash Point	Cool 4 deg C	PSEG03	N31	TOTE-COMP	01/06/2023	1010B
01/11/2023	Water	O1054-06	Δ Flash Point	Cool 4 deg C	PSEG03	N31	COMP-03	01/06/2023	1010B
01/11/2023	Water	O1054-07	Δ Flash Point	Cool 4 deg C	PSEG03	N31	2026	01/06/2023	1010B
01/13/2023	Water	O1061-12	Δ Flash Point	Cool 4 deg C	PSEG03	N31	TOTE-COMP	01/06/2023	1010B
01/11/2023	Water	O1066-01	Δ Flash Point	Cool 4 deg C	PSEG03	N31	OILY-WATER-1	01/06/2023	1010B

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

01/09/2023 08:30
Rm CWC
M (GOC)

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

01/09/2023 12:15
Rm CWC
Rm CWC