

LB123587


7471B

INSTRUMENT ID: CY1

Linear

μ Abs.:

49477



A= 0.0000e+000

B= 2.0079e-004

C= -4.4324e-003

Rho= 0.9999023

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.015	0.015	99	0.000	99					4/11
0.2	0.200	0.183	-0.017	935	0.0 %	935					-9
2.5	2.500	2.455	-0.045	12251	0.0 %	12251					-2
5.0	5.000	5.028	0.028	25063	0.0 %	25063					-1
7.5	7.500	7.588	0.088	37814	0.0 %	37814					-1
10.0	10.000	9.930	-0.070	49477	0.0 %	49477					-1

LB123587 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Jan 2023 11:02:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	99	-	0.0000	7471B	PPB	09 Jan 2023 11:02:49	S	Std
0.2 - 1	50.2	935	-	0.2000	7471B	PPB	09 Jan 2023 11:05:06	S	Std
2.5 - 1	52.5	12251	-	2.5000	7471B	PPB	09 Jan 2023 11:07:23	S	Std
5.0 - 1	55	25063	-	5.0000	7471B	PPB	09 Jan 2023 11:09:40	S	Std
7.5 - 1	57.5	37814	-	7.5000	7471B	PPB	09 Jan 2023 11:11:56	S	Std
10.0 - 1	51.0	49477	-	10.0000	7471B	PPB	09 Jan 2023 11:16:44	S	Std
ICV95 - 1	ICV95	20848	4.1816	-	7471B	PPB	09 Jan 2023 11:19:31	U	SMPL
ICB95 - 1	ICB95	-4	-0.0052	-	7471B	PPB	09 Jan 2023 11:21:46	U	SMPL
CCV20 - 1	CCV20	25279	5.0712	-	7471B	PPB	09 Jan 2023 11:24:01	U	SMPL
CCB20 - 1	CCB20	-57	-0.0159	-	7471B	PPB	09 Jan 2023 11:26:16	U	SMPL
CRA - 1	CRA	866	0.1694	-	7471B	PPB	09 Jan 2023 11:28:31	U	SMPL
HighStd - 1	HighStd	50622	10.1598	-	7471B	PPB	09 Jan 2023 11:30:46	U	SMPL
ChkStd - 1	ChkStd	35395	7.1024	-	7471B	PPB	09 Jan 2023 11:33:01	U	SMPL
PB150123BL - 1	PBS	-83	-0.0211	-	7471B	PPB	09 Jan 2023 11:35:18	U	SMPL
PB150123BS - 1	LCSS	20537	4.1191	-	7471B	PPB	09 Jan 2023 11:37:34	U	SMPL
O1043-01 - 1	TP-203	504	0.0968	-	7471B	PPB	09 Jan 2023 11:39:51	U	SMPL
O1043-04 - 1	TP-202	736	0.1433	-	7471B	PPB	09 Jan 2023 11:42:07	U	SMPL
O1047-01 - 1	TP-6	636	0.1233	-	7471B	PPB	09 Jan 2023 11:44:24	U	SMPL
O1047-01DUP - 1	TP-6DUP	570	0.1100	-	7471B	PPB	09 Jan 2023 11:46:42	U	SMPL
O1047-01MS - 1	TP-6MS	18081	3.6260	-	7471B	PPB	09 Jan 2023 11:48:57	U	SMPL
CCV21 - 1	CCV21	25183	5.0520	-	7471B	PPB	09 Jan 2023 11:51:12	U	SMPL
CCB21 - 1	CCB21	-77	-0.0199	-	7471B	PPB	09 Jan 2023 11:53:28	U	SMPL
O1047-01MSD - 1	TP-6MSD	19413	3.8934	-	7471B	PPB	09 Jan 2023 11:55:43	U	SMPL
O1048-01 - 1	CL-01-010523	2197	0.4367	-	7471B	PPB	09 Jan 2023 11:57:59	U	SMPL
O1053-01 - 1	TP-201	777	0.1516	-	7471B	PPB	09 Jan 2023 12:00:14	U	SMPL
O1047-01LX5 - 1		102	0.0160	-	7471B	PPB	09 Jan 2023 12:02:30	U	SMPL
O1047-01A - 1		19872	3.9856	-	7471B	PPB	09 Jan 2023 12:04:46	U	SMPL
CCV22 - 1	CCV22	25384	5.0923	-	7471B	PPB	09 Jan 2023 12:07:03	U	SMPL
CCB22 - 1	CCB22	-26	-0.0097	-	7471B	PPB	09 Jan 2023 12:09:19	U	SMPL