



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/10/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:20
In Date: 01/09/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 01/10/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123590

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1070-01	MDBXS6	1	1.16	8.44	9.6	9.47	98.5	
O1070-02	MDBXS6D	2	1.16	8.44	9.6	9.47	98.5	
O1070-03	MDBXS6S	3	1.16	8.44	9.6	9.47	98.5	
O1070-04	MDBXV2	4	1.13	8.58	9.71	9.57	98.4	
O1070-05	MDBXV4	5	1.15	8.82	9.97	9.77	97.7	
O1070-06	MDBXW3	6	1.13	8.65	9.78	9.58	97.7	
O1070-07	MDBXW4	7	1.12	8.87	9.99	9.81	98.0	
O1070-08	MDBXX7	8	1.11	8.66	9.77	9.53	97.2	
O1070-09	MDBYA4	9	1.12	8.49	9.61	9.4	97.5	
O1070-10	MDBYA5	10	1.19	8.51	9.7	9.5	97.6	
O1070-11	MDBYB1	11	1.12	8.56	9.68	9.45	97.3	
O1070-12	MDBYC2	12	1.12	8.53	9.65	9.47	97.9	
O1070-13	MDBYE5	13	1.19	8.73	9.92	9.62	96.6	
O1070-14	MDBYE6	14	1.12	8.85	9.97	9.64	96.3	
O1070-15	MDBYE8	15	1.14	8.57	9.71	9.53	97.9	
O1070-16	MDBYE9	16	1.13	8.75	9.88	9.7	97.9	
O1070-17	MDBYF0	17	1.11	8.68	9.79	9.63	98.2	
O1070-18	MDBYF1	18	1.16	8.53	9.69	9.41	96.7	
O1070-19	MDBYF2	19	1.19	8.74	9.93	9.68	97.1	
O1070-20	MDBYF3	20	1.15	8.60	9.75	9.44	96.4	
O1070-21	MDBYF4	21	1.13	8.64	9.77	9.61	98.1	
O1070-22	MDBYF9	22	1.13	8.70	9.83	9.66	98.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$