

LB123593


7470A

INSTRUMENT ID: CV4

Linear

 μ Abs.:

44801



A= 0.0000e+000

B= 2.1683e-004

C= 1.4427e-004

Rho= 0.9985117

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.0	0.000	0.035	0.035	161	0.000	161					
0.2	0.200	0.189	-0.011	870	0.0 %	870					-6
2.5	2.500	2.335	-0.165	10767	0.0 %	10767					-7
5.0	5.000	5.070	0.070	23380	0.0 %	23380					1
7.5	7.500	7.858	0.358	36238	0.0 %	36238					5
10.0	10.000	9.714	-0.286	44801	0.0 %	44801					-3

LB123593 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Jan 2023 13:53:04

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	161	-	0.0000	7470A	PPB	09 Jan 2023 13:55:50	S	Std
0.2 - 1	51.2	870	-	0.2000	7470A	PPB	09 Jan 2023 13:58:07	S	Std
2.5 - 1	52.5	10767	-	2.5000	7470A	PPB	09 Jan 2023 14:00:24	S	Std
5.0 - 1	55	23380	-	5.0000	7470A	PPB	09 Jan 2023 14:02:41	S	Std
7.5 - 1	57.1	36238	-	7.5000	7470A	PPB	09 Jan 2023 14:04:58	S	Std
10.0 - 1	51.1	44801	-	10.0000	7470A	PPB	09 Jan 2023 14:07:14	S	Std
ICV96 - 1	ICV96	17600	3.8163	-	7470A	PPB	09 Jan 2023 14:10:04	U	SMPL
ICB96 - 1	ICB96	181	0.0394	-	7470A	PPB	09 Jan 2023 14:12:18	U	SMPL
CCV23 - 1	CCV23	22478	4.8740	-	7470A	PPB	09 Jan 2023 14:14:33	U	SMPL
CCB23 - 1	CCB23	118	0.0257	-	7470A	PPB	09 Jan 2023 14:16:48	U	SMPL
CRA - 1	CRA	874	0.1897	-	7470A	PPB	09 Jan 2023 14:19:03	U	SMPL
HighStd - 1	HighStd	44753	9.7039	-	7470A	PPB	09 Jan 2023 14:21:18	U	SMPL
ChkStd - 1	ChkStd	32885	7.1305	-	7470A	PPB	09 Jan 2023 14:26:07	U	SMPL
PB150099BL - 1	PBW	109	0.0238	-	7470A	PPB	09 Jan 2023 14:28:23	U	SMPL
PB150099BS - 1	LCSW	17988	3.9005	-	7470A	PPB	09 Jan 2023 14:30:39	U	SMPL
O1007-01 - 1	TREMONT-DRUM-23-PPE	89	0.0194	-	7470A	PPB	09 Jan 2023 14:32:55	U	SMPL
O1007-02 - 1	TREMONT-DRUM-24-POLY	54	0.0119	-	7470A	PPB	09 Jan 2023 14:35:11	U	SMPL
O1027-02 - 1	COMP-1-1-4-2023	53	0.0116	-	7470A	PPB	09 Jan 2023 14:37:28	U	SMPL
O1029-02 - 1	TP-205	109	0.0238	-	7470A	PPB	09 Jan 2023 14:39:45	U	SMPL
O1029-05 - 1	TP-204	232	0.0504	-	7470A	PPB	09 Jan 2023 14:42:00	U	SMPL
CCV24 - 1	CCV24	22279	4.8309	-	7470A	PPB	09 Jan 2023 14:44:15	U	SMPL
CCB24 - 1	CCB24	44	0.0097	-	7470A	PPB	09 Jan 2023 14:46:31	U	SMPL
O1030-02 - 1	TP-5	401	0.0871	-	7470A	PPB	09 Jan 2023 14:48:47	U	SMPL
O1038-01 - 1	1X10-GLASS-BEADS	436	0.0947	-	7470A	PPB	09 Jan 2023 14:51:02	U	SMPL
O1043-02 - 1	TP-203	316	0.0687	-	7470A	PPB	09 Jan 2023 14:53:18	U	SMPL
O1043-05 - 1	TP-202	260	0.0565	-	7470A	PPB	09 Jan 2023 14:55:34	U	SMPL
O1047-02 - 1	TP-6	23	0.0051	-	7470A	PPB	09 Jan 2023 14:57:50	U	SMPL
O1047-02DUP - 1	TP-6DUP	47	0.0103	-	7470A	PPB	09 Jan 2023 15:00:06	U	SMPL
O1047-02MS - 1	TP-6MS	17034	3.6936	-	7470A	PPB	09 Jan 2023 15:02:22	U	SMPL
O1047-02MSD - 1	TP-6MSD	16911	3.6669	-	7470A	PPB	09 Jan 2023 15:04:40	U	SMPL
PB150100BL - 1	PBW	24	0.0053	-	7470A	PPB	09 Jan 2023 15:06:57	U	SMPL
PB150100BS - 1	LCSW	16547	3.5880	-	7470A	PPB	09 Jan 2023 15:09:13	U	SMPL
CCV25 - 1	CCV25	22161	4.8053	-	7470A	PPB	09 Jan 2023 15:11:29	U	SMPL
CCB25 - 1	CCB25	131	0.0285	-	7470A	PPB	09 Jan 2023 15:13:44	U	SMPL
O1038-02 - 1	1X1-GREEN-POWER	883	0.1916	-	7470A	PPB	09 Jan 2023 15:16:00	U	SMPL
O1038-03 - 1	1X2-GREEN-POWER	488	0.1060	-	7470A	PPB	09 Jan 2023 15:18:15	U	SMPL
O1038-04 - 1	1X3-GREEN-POWER	470	0.1021	-	7470A	PPB	09 Jan 2023 15:20:31	U	SMPL
O1038-05 - 1	1X4-GREEN-POWER	354	0.0769	-	7470A	PPB	09 Jan 2023 15:22:47	U	SMPL
O1038-06 - 1	1X5-GREEN-POWER	108	0.0238	-	7470A	PPB	09 Jan 2023 15:25:03	U	SMPL
O1038-07 - 1	1X6-GREEN-POWER	64	0.0140	-	7470A	PPB	09 Jan 2023 15:27:19	U	SMPL
O1038-08 - 1	1X7-GREEN-POWER	82	0.0179	-	7470A	PPB	09 Jan 2023 15:29:35	U	SMPL
O1038-09 - 1	1X8-GREEN-POWER	88	0.0192	-	7470A	PPB	09 Jan 2023 15:31:53	U	SMPL
O1038-10 - 1	1X9-GREEN-POWER	105	0.0229	-	7470A	PPB	09 Jan 2023 15:34:09	U	SMPL
O1038-10DUP - 1	1X9-GREEN-POWERDUP	84	0.0184	-	7470A	PPB	09 Jan 2023 15:36:25	U	SMPL
CCV26 - 1	CCV26	22248	4.8241	-	7470A	PPB	09 Jan 2023 15:38:41	U	SMPL
CCB26 - 1	CCB26	133	0.0290	-	7470A	PPB	09 Jan 2023 15:40:56	U	SMPL
O1038-10MSD - 1	1X9-GREEN-POWERMSD	16534	3.5852	-	7470A	PPB	09 Jan 2023 15:45:32	U	SMPL
PB150130BL - 1	PBW	35	0.0077	-	7470A	PPB	09 Jan 2023 15:47:47	U	SMPL
PB150130BS - 1	LCSW	16347	3.5448	-	7470A	PPB	09 Jan 2023 15:50:03	U	SMPL
O1013-02 - 1	COMP	235	0.0511	-	7470A	PPB	09 Jan 2023 15:52:18	U	SMPL
O1013-02DUP - 1	COMPDUP	220	0.0478	-	7470A	PPB	09 Jan 2023 15:54:35	U	SMPL
O1013-02MS - 1	COMPMS	11428	2.4781	-	7470A	PPB	09 Jan 2023 15:56:52	U	SMPL
O1013-02MSD - 1	COMPMSD	12315	2.6704	-	7470A	PPB	09 Jan 2023 15:59:08	U	SMPL
O1040-02 - 1	DSN001	1177	0.2554	-	7470A	PPB	09 Jan 2023 16:01:26	U	SMPL
O1040-04 - 1	DSN002	561	0.1218	-	7470A	PPB	09 Jan 2023 16:03:41	U	SMPL
CCV27 - 1	CCV27	22797	4.9432	-	7470A	PPB	09 Jan 2023 16:05:57	U	SMPL
CCB27 - 1	CCB27	100	0.0218	-	7470A	PPB	09 Jan 2023 16:08:13	U	SMPL
O1040-05 - 1	DSN003	513	0.1114	-	7470A	PPB	09 Jan 2023 16:10:28	U	SMPL
O1044-01 - 1	COMP-SAMPLE-TANK-10-34-35	256	0.0557	-	7470A	PPB	09 Jan 2023 16:12:43	U	SMPL
O1045-01 - 1	COMP-SAMPLE-TANK-11-12-31-32-33	431	0.0936	-	7470A	PPB	09 Jan 2023 16:14:59	U	SMPL
O1046-01 - 1	COMP-SAMPLE-API-3	262	0.0570	-	7470A	PPB	09 Jan 2023 16:17:15	U	SMPL
O1052-01 - 1	UST-DRUM	3901	0.8460	-	7470A	PPB	09 Jan 2023 16:19:30	U	SMPL
O1061-12 - 1	TOTE-COMP	1187	0.2575	-	7470A	PPB	09 Jan 2023 16:21:47	U	SMPL
O1066-01 - 1	OILY-WATER-1	305	0.0663	-	7470A	PPB	09 Jan 2023 16:24:03	U	SMPL
PB150056TB - 1		126	0.0275	-	7470A	PPB	09 Jan 2023 16:26:20	U	SMPL
PB150095TB - 1		70	0.0153	-	7470A	PPB	09 Jan 2023 16:28:37	U	SMPL
CCV28 - 1	CCV28	23087	5.0061	-	7470A	PPB	09 Jan 2023 16:30:52	U	SMPL
CCB28 - 1	CCB28	166	0.0361	-	7470A	PPB	09 Jan 2023 16:33:08	U	SMPL
O1047-02LX5 - 1		42	0.0093	-	7470A	PPB	09 Jan 2023 16:35:24	U	SMPL
O1047-02A - 1		18051	3.9141	-	7470A	PPB	09 Jan 2023 16:37:40	U	SMPL
O1038-10LX5 - 1		171	0.0372	-	7470A	PPB	09 Jan 2023 16:39:56	U	SMPL
O1038-10A - 1		17609	3.8183	-	7470A	PPB	09 Jan 2023 16:42:12	U	SMPL
O1013-02LX5 - 1		60	0.0132	-	7470A	PPB	09 Jan 2023 16:44:29	U	SMPL
O1013-02A - 1		17633	3.8235	-	7470A	PPB	09 Jan 2023 16:46:46	U	SMPL
O1038-10MS - 1	1X9-GREEN-POWERMS	17044	3.6958	-	7470A	PPB	09 Jan 2023 16:49:02	U	SMPL
CCV29 - 1	CCV29	22969	4.9805	-	7470A	PPB	09 Jan 2023 16:51:19	U	SMPL
CCB29 - 1	CCB29	164	0.0357	-	7470A	PPB	09 Jan 2023 16:53:36	U	SMPL